

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kozhikode
2	Name of Hospital	PHC Ayancherri
3	Name of Equipment with Make, Model and Serial Number	UPS for Semi Auto Biochemistry Make: APC Model: BX 600 Serial: BC21916007659 No
4	Equipment ID & Barcode	82016 9 1141793
5	Date of purchase / Year of manufacture /Installation Date	12/09/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC No CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	23/11/2023
9	Action taken	Checked and found main board, Battery, Transformer & Relay unit defective
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	Battery - Rs 1100 Main Board - Rs 2500 Other components - Rs 500 [Transformer, Relay etc] Rs 4100 <u>Rs 4100</u>

13	Asset Value	₹4500
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	91%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report Attached & Vendor quotation
16	Reasons for recommending the equipment as BER	Checked & Found Main Board, Battery Transformer, Relay unit other components defective. Equipment installed on 12/09/2019 completed 4 years and 3 months. The repair cost is beyond economical. Met both criteria So recommended for Condemnation process
17	Name & Signature of CYRIX Authority	Akhil Suresh S <i>Akhil Suresh S</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & found that main board, Battery, transformer, & relay unit are damaged. Equipment installed on 12/09/2019

Signature of JC BM (NHM)
National Health Mission
Kozhikode

Date _____ Seal _____



 Director's Office
 Superintendent / Medical Officer (i/c)

Kudiyil



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 108025

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>PHC Ayancheri</u></p> <p>Address : <u>Kozhikode</u></p> <p>Ph : <u>7907973618</u></p>	<p>Call Registration Date : <u>23/11/23</u></p> <p>Caller ID : <u>82016</u></p> <p>Date of Visit : <u>28/11/23</u></p> <p>Asset No. : <u>1141793</u></p> <p>EQPT Name : <u>UPS for Semi Auto Biochemistry Analyzer</u></p> <p>Manufacture <u>APC</u> Model : <u>BX600C</u></p> <p>S. No. <u>B21916007688</u> Dept. <u>Laboratory</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not getting ON

Action Taken : Checked and found Mainboard, Battery, Transformer and Relay unit defective need to replace those spares for further checking and working condition of equipment.

Completed Date : 04/12/23 Time : 10:30am Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	<u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
2.	<u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
3.				

Cyrux Engineer	Date	Start Time	End Time
<u>Akhil Suresh S</u>	<u>04/12/23</u>	<u>10am</u>	<u>10:30am</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Akhil Suresh S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>04/12/23</u></p> <p>Contact Number : <u>9995706213</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>04/12/23</u></p> <p>Contact Number : <u>04/12/23</u></p> <p>Designation :</p> <p>Hospital Seal :</p>
--	---



DICKSON F S
PHARMACEUTIST

GSTIN: 32AAJFB3520H1ZY
SSI No: 3204020549
SSI Date: 31-7-04

Ph: 0496 2525176 (Shop)
2512680 (Service)

Battery House

BATTERY, UPS, HUPS, SOLAR & INVERTER SALES & SERVICE
Narayana Nagar Ground, Vatakara - Pin: 673 101

WEXIDE
INDUSTRIES LIMITED

Whole sale & Retail Outlet

Quotation

Date: 04/12/2020

To,
The Manager
Cyrix Health Care Pvt. Ltd

Sir,

Good via APC - UPS - Company checked and
found Complaint. spare rate is

- 1. Battery - Rs. 1100
- 2. Main board - Rs. 2500
- 3. Other Components
(Transformer, Relay, capacitor
etc.) } Rs. 500

Rs.

4100

(Rupees four thousand and one hundred only)

Vatakara



TO BATTERY HOUSE
Manager

	NAME OF EQUIPMENT	DETAILS
10	UPS APC 600 VA	INVOICE NO: C0192000R142 dated: 12/9/19 Product Code: 22340004 Received from KMSCL
11	3-PART HAEMATOLOGY ANALYSER	Model No: Medonic M32B Serial NO: 114727 Order NO 50/750/2019-2020/57 dated 15/10/20
12	C	

Recommended for condemnation

Mudya
 മെഡിക്കൽ ഔദ്യോഗിക
 കുടുംബാരോഗ്യകേന്ദ്രം ആയുർവേദ
 ട്രസ്റ്റ് പാസ്റ്റ് 67:542





Recommended for condemnation

Krudys

മെഡിക്കൽ ഓഫീസർ
കുടുംബാരോഗ്യകേന്ദ്രം ആയഞ്ചേരി
കടമേരി പോസ്റ്റ് 67 547

