


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	WAYANAD
2	Name of Hospital	DH Mananthavady
3	Name of Equipment with Make, Model and Serial Number	Equipment-X-ray machine Make-G E Model-HF 400 Sn-9011567
4	Equipment ID/ Barcode	1210566//TICKET 91932
5	Date of purchase/ Year of manufacture/Installation Date	14-10-1998
6	Warranty details (Yes/No)	No
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	21/01/2024/Toll Free
9	Action taken	Checked, found programmed Jedi rotation V3 board and Jedi KV Control board defective. Enquired OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	programmed Jedi rotation V3 board and Jedi KV Control spare rate is 395000/-

13	Asset Value	368040/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	107.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Quotation from Diagnostic equipment services P Ltd (Authorized Dealer-Wipro GE Health care) Cyril service report
16	Reasons for recommending the equipment as BER	Checked the equipment and found programmed Jedi rotation V3board and Jedi KV Control spare defective. Equipment completed 25 years Repairing cost is 107.3% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	NIYAS HADI AP 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Jedi rotation V3 board and Jedi KV Control spare defective. Already 25 years completed. Recommended condemnation


CONSULTANT
Bio-Medical Engineer
Signature of JC BM (NHM)
(W. of Kerala)
Wayanad-673122
25/03/24

old and defective machine, not working since 2 months.

Date:




Signature of
Superintendent/Medical Officer
25/3/24
Registered: VP
SUPERINTENDENT
DISTRICT HOSPITAL
WAYANAD, MANANTHAVAD.



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 181857

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>District Hospital</u></p> <p>Address <u>Manantharady</u> <u>Mayanad, Kerala</u></p> <p>Ph : <u>9947918269</u></p>	<p>Call Registration Date : <u>21/1/2024</u></p> <p>Caller ID : <u>91932</u></p> <p>Date of Visit : <u>14/3/2024</u></p> <p>Asset No. : <u>1210566</u></p> <p>EQPT Name : <u>X ray Machine</u></p> <p>Manufacture <u>GE</u> Model : <u>Column 90 1/2</u></p> <p>S. No. <u>9011567</u> Dept. <u>X ray dept.</u></p>
---	--

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : JEDI communication error, Error 60, Error 120.


Action Taken : checked the unit and found programmed jedi rotation V3 board and jedi kv control board defective. Crosschecked both boards and confirmed the complaint.

Completed **Date :** 14/3/24 **Time :** 1:00pm **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>NIYAS</u>	<u>14-3-24</u>	<u>11:00pm</u>	<u>1:00 PM</u>



Customer Remark Completed Pending RBER

SUPERINTENDENT
DISTRICT HOSPITAL
MANANTHARADY

Service Engineer Name : N. Niyas

Signature : [Signature]

Date : 14-03-2024

Contact Number : 7593847171

Customer Name : [Signature]

Signature : [Signature]

Date : 14/3/2024

Contact Number : 9947918269

Designation : [Signature]

Hospital Seal : [Signature]



DIAGNOSTIC EQUIPMENTS SERVICES PVT. LTD

Aravindan Mansion, CC No : 39/1479, North Janatha Road, Palarivattom, Ernakulam, Kerala - 682 025
Mob : 0484 351 2324 | 9947246593 | 9947246591 | 9947246595 | Email : decochin.ge@gmail.com
GST#32AABCD0243L1Z3

Date: 15-2-2024

To,

DISTRICT HOSPITAL
MANANTHAVADI

Sir,

Sub: Offer For the spare part of HF 400 X Ray Machine

This has reference to the discussion you had with our Engineer regarding the requirement of spare parts for your HF 400 X Ray machine. We are pleased to give below our best offer for the same.

Description	Model	Tax Rate	Excl. Price	Incl. Price
Programmed Jedi Rotation V3 Board	HF 400	12%	352679/ -	395000/-
Jedi KV Control Spare				

(Rs. THREE LAKH NINETY FIVE THOUSAND ONLY)

Terms and Condition

- a) Price : Price indicated is on Exchange Basis
- b) Delivery : 4-6 weeks on receipt of your order complete in all respects.
- c) Order : Diagnostic Equipments Services Pvt Ltd
- d) Validity : This Quote shall be valid for a period of 30 days only
- e) Payment : 100% Advance
- f) Cancellation : 5% of the order value would be deducted from the advance if order before dispatch and 20% of the order value if cancelled after dispatch and adjusted towards sales tax/freight liability.

For any clarifications please feel free to contact the undersigned. We now look forward for your order and assuring you of our best services of always, we remain,

Yours truly,

For Diagnostic Equipments Services pvt ltd

Arun Kumar V
9947246595



Authorised Dealer
Wipro GE Healthcare



DIAGNOSTIC EQUIPMENTS SERVICES PVT. LTD

Aravindan Mansion, CC No : 39/1479, North Janatha Road, Palarivattom, Ernakulam, Kerala - 682 025
Mob : 0484 351 2324 | 9947246593 | 9947246591 | 9947246595 | Email : decochin.ge@gmail.com
GST#32AABCD0243L1Z3

Date: 15-2-2024

To,

DISTRICT HOSPITAL
MANANTHAVADI

Sir,

Sub: Offer For the spare part of HF 400 X Ray Machine

This has reference to the discussion you had with our Engineer regarding the requirement of spare parts for your HF 400 X Ray machine. We are pleased to give below our best offer for the same.

Description	Model	Tax Rate	Excl. Price	Incl. Price
Programmed Jedi Rotation V3 Board	HF 400	12%	352679/ -	395000/-
Jedi KV Control Spare				

(Rs. THREE LAKH NINETY FIVE THOUSAND ONLY)

Terms and Condition

- Price** : Price indicated is on Exchange Basis
- Delivery** : 4-6 weeks on receipt of your order complete in all respects.
- Order** : Diagnostic Equipments Services Pvt Ltd
- Validity** : This Quote shall be valid for a period of 30 days only
- Payment** : 100% Advance
- Cancellation** : 5% of the order value would be deducted from the advance if order before dispatch and 20% of the order value if cancelled after dispatch and adjusted towards sales tax/freight liability.

For any clarifications please feel free to contact the undersigned. We now look forward for your order and assuring you of our best services of always, we remain,

Yours truly,

For Diagnostic Equipments Services pvt ltd

Arun Kumar V
9947246595



Received from
25/2/24
SUPERINTENDENT
DISTRICT HOSPITAL
WAYANAD, MANANTHAVADI



Authorised Dealer
Wipro GE Healthcare



SUPPLIER'S CODE :	H14408	
VOLTS / CYCLES :	415V 3- / 50Hz	
PART NO :	6289052	
SER. NO. :	SM-13-05-0103	
MANUFACTURED :	JANUARY - 2013	
DESCRIPTION :	GENERATOR CABINET	

GE Medical Systems (India) Pvt. Ltd.
 #122, EPIP, Part 1, Whitefield, Bangalore-560088
 Made in India

VOLTS / CYCLES :	415V / 50Hz	PHASE :	3-
Z.CAT NO. :	Z1000260	MANUFACTURED :	04-2013
SER. NO. :	T3117880		
DESCRIPTION :	HF-ADVANTAGE 32KV SYSTEM		



DISTRICT HOSPITAL MANANTHAVADY
 PH: 04935 240 264
 WAYANAD - 670 645

25/3/20
 SUPERINTENDENT
 DISTRICT HOSPITAL
 WAYANAD, MANANTHAVADY



Fwd: Spare Quotation - D H Mananthavadi, HF Advantage

2 messages

Aneesh P B <mklse@cyrix.in>
To: niyascyrix@gmail.com <niyascyrix@gmail.com>

Thu, 29 Feb 2024 at 9:54 am

Kind Regards,
Aneesh P B
Deputy Manager - Service
Ph: 9744827770

Cyrix Healthcare Pvt Ltd | 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

Ph: 9744555073 | M 9072522093 | W www.cyrixhealthcare.com

KERALA || KARNATAKA || UTHAR PRADESH || TAMIL NADU || KUWAIT || KSA || UAE

From: DES COCHIN <descochin.ge@gmail.com>
Sent: Thursday, February 15, 2024 11:30:27 AM
To: Aneesh P B <mklse@cyrix.in>
Subject: Re: Spare Quotation - D H Mananthavadi, HF Advantage

Sir
Please find the Quotation

Regards
Treesa



Diagnostic Equipments Services (P) Ltd
Aravindan Mansion, CCno:39/1479
North Janatha Road
Palarivattom-682025

On Thu, Feb 15, 2024 at 9:58 AM Aneesh P B <mklse@cyrix.in> wrote:

Dear Sir,

Please provide a quotation for the Rotor Control Board and KV control board spares for the GE HF Advantage equipment installed at DH Mananthavadi.

Thank you,
Aneesh P B
Deputy Manager - Service
Cyrix Healthcare Pvt Ltd
30/641 B, Pettah Jn, Poonithura, Cochin -38, Kerala, India

Ph: 9744555073 | M 9072522093 | W www.cyrixhealthcare.com

KERALA || KARNATAKA || UTHAR PRADESH || TAMIL NADU || KUWAIT || KSA || UAE

Niyas Hadi AP <niyascyrix@gmail.com>
To: Sarang KM <sarangcyrix@gmail.com>

Thu, 14 Mar 2024 at 9:07 am

Dear Sir
For your information

Thanks & regards

NIYAS HADI
District incharge
BEMP project
Wayanad
kerala
mob -7593847171

email -niyascyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

[Quoted text hidden]