


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	WAYANAD
2	Name of Hospital	PH LAB Sulthan Bathery
3	Name of Equipment with Make, Model and Serial Number	Equipment-Centrifuge Make-Tarsons Model-spin win MC100 Sn-1501R341
4	Equipment ID/ Barcode	1212072//TICKET 94799
5	Date of purchase/ Year of manufacture/Installation Date	09/06/2016
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	07/02/2024/Toll free
9	Action taken	Checked and found motor and sample holder defective
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	RS 8300/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not applicable
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Letter from hospital authority Cyrix service report
16	Reasons for recommending the equipment as BER	Checked and found motor and sample holder defective. Equipment installed on 09.06.2016 and aged up to 7 years and 9 months. The equipment gets by donation and there are no supplier details available So we recommended for Condemnation as per tender clause 5.3.14.2.
17	Name & Signature of CYRIX Authority	NIYAS HADI AP 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and sample holder of equipment is defective. No supplier details available - Recommended Condemnation


25/3/24
CONSULTANT
Signature of J.C. BM (NHM)
District Health
Arogyakeralam
Wayanad-673122



Date: 25-03-2024


CONSULTANT
Signature of
Superintendent/Medical Officer (I/c)
(Arogyakeralam)
Wayanad-673122

Medical Officer
District Public Health Laboratory
Wayanad, Sulthan Bathery



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 181827

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : PUBLIC HEALTH	Call Registration Date : 7/2/2024
Address : LABORATORY	Caller ID : 94799
WADIONAD KERALA	Date of Visit : 8/2/2024
Ph : 9846309569	Asset No. : 1212032
	EQPT Name : CENTRIFUGE
	Manufacture JORSON Model : SPIN WIN 100
	S. No. 1501R241 Dept. LDC

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Motor sample holder defective and motor not rotating

Action Taken : Inspected the equipment and found that the motor failure and sample holder defective. Need to repair or replace the motor and sample holder

Completed Date : 8/2/2024 Time : 1:00PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Dijukam. EGOS	8/2/2024	12:00PM	1:00PM

Customer Remark *Medical Officer* Completed Pending
District Public Health Laboratory, Wayanad, Sulfh
Sabina

Service Engineer Name : <i>Dijukam</i>	Customer Name : Febna. P
Signature : <i>[Signature]</i>	Signature : <i>[Signature]</i>
Date : 8/2/2024	Date : 8/2/24
Contact Number : 8086752545	Contact Number : 949681058
	Designation : SGT
	Hospital Seal :

"ഭരണഭാഷ മാതൃഭാഷ"

**ജില്ലാ പബ്ലിക് ഹെൽത്ത് ലബോറട്ടറി, വയനാട്,
സുൽത്താൻ ബത്തേരി, പിൻ- 673 592**

ഫോൺ: 04936 223113, e-mail: dphlwayanad@gmail.com

നമ്പർ: പി.എച്ച്.എൽ-123/2024

തീയതി: 22/03/2024

പ്രേഷകൻ,

മെഡിക്കൽ ഓഫീസർ
ജില്ലാ പബ്ലിക് ഹെൽത്ത് ലാബ്
സുൽത്താൻ ബത്തേരി, വയനാട്

സ്വീകർത്താവ്

Cyrux Health Care Pvt.Ltd

സർ,

വിഷയം:- non repairable certificate - അപേക്ഷ സമർപ്പിക്കുന്നത് സംബന്ധിച്ച്
സൂചന:-

ജില്ലാ പബ്ലിക് ഹെൽത്ത് ലാബിലെ RTPCR ലാബിൽ ഉപയോഗിച്ച് കൊണ്ടിരിക്കുന്ന SPINNER (TARSON SPINWIN Mc-00,CA-1010,SL No.150IR341) കേടായതിനാൽ complaint രജിസ്റ്റർ ചെയ്തിരുന്നു .എന്നാൽ അത് non repairable ആണെന്നും non repairable certificate ലഭിക്കുന്നതിനായി ഉപകരണത്തിന്റെ ബില്ല് ,വാറണ്ടി കാർഡ്, ഇൻസ്റ്റാളേഷൻ റിപ്പോർട്ട് തുടങ്ങിയ രേഖകൾ ആവശ്യമാണെന്ന് അറിയിച്ചിരുന്നു .എന്നാൽ പ്രസ്തുത ഉപകരണം മണിപ്പാൽ ഇൻസ്റ്റിട്യൂട്ടിൽ നിന്നും DMO യ്ക്ക് നൽകുകയും DMO മുഖാന്തിരം ജില്ലാ പബ്ലിക് ഹെൽത്ത് ലാബിന് കൈമാറുകയും ചെയ്താണ്. ഈ ഉപകരണത്തോടൊപ്പം ബില്ല്, വാറണ്ടി കാർഡ് , ഇൻസ്റ്റാളേഷൻ റിപ്പോർട്ട് തുടങ്ങിയ രേഖകൾ കൈമാറി കിട്ടിയിട്ടില്ല എന്നറിയിക്കുന്നു. Non repairable certificate ലഭ്യമാക്കിത്തരുവാൻ നടപടി സ്വീകരിക്കണമെന്നും അപേക്ഷിക്കുന്നു . .



വിശ്വസ്യതയോടെ
(Handwritten Signature)
മെഡിക്കൽ ഓഫീസർ,
ജില്ലാ പബ്ലിക് ഹെൽത്ത് ലാബ്
സുൽത്താൻ ബത്തേരി, വയനാട്

പകർപ്പ്:- ഫയൽ

