



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : **189381**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <i>phe chidtariparamba</i></p> <p>Address ..... <i>Kensar</i></p> <p style="padding-left: 100px;"><i>Kerala</i></p> <p>Ph : ..... <i>7510237592</i></p>	<p>Call Registration Date : ..... <i>13-3-24</i></p> <p>Caller ID : ..... <i>101040</i></p> <p>Date of Visit : ..... <i>14-3-24</i></p> <p>Asset No. : ..... <i>1345245</i></p> <p>EQPT Name : ..... <i>UPS</i></p> <p>Manufacture ..... <i>EMERSON</i> Model : ..... <i>na</i></p> <p>S. No. ..... <i>na</i> Dept. ..... <i>Laboratory</i></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... *nod switching on. Main board & 3 batteries are defective*

**Action Taken :** ..... *checked the equipment. Found Equipment is not switching on. Found main board and 3 batteries are defective. Need to replace these spares for further checking*

Completed  Date : ..... *14-3-24* Time : ..... *11:40 am* Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Shankrishnan ppk</i>	<i>14-3-24</i>	<i>10:30 AM</i>	<i>11:40 am</i>

**Customer Remark** Completed  Pending

*For M.O. Dr. Ajith V Jagan*

<p><b>Service Engineer Name :</b> <i>Shankrishnan ppk</i></p> <p><b>Signature :</b> <i>[Signature]</i></p> <p><b>Date :</b> <i>14-3-24</i></p> <p><b>Contact Number :</b> <i>8281188514</i></p>	<p><b>Customer Name :</b> <i>Shaferase A.R</i></p> <p><b>Signature :</b> <i>[Signature]</i></p> <p><b>Date :</b> <i>14/3/24</i></p> <p><b>Contact Number :</b> <i>7510237592</i></p> <p><b>Hospital Seal :</b> <i>KT CH-I</i></p>
---	---






**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**


**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KANNUR
2	Name of Hospital	PHC CHITTARIPARAMB
3	Name of Equipment with Make, Model and Serial Number	UPS Make: EMERSON Model: NA SN: NA
4	Equipment ID/ Barcode	1345245 #101040
5	Date of purchase/ Year of manufacture/Installation Date	27-11-2013
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 13-03-2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, Found main board and 3 batteries are defective. Need to Replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Main Board(1)-1500/- Battery(3)-3300/- Total=4800/-

13	Asset Value	4500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	106.666%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine found machine is not switching ON, Found main board 3 batteries are defective. Need to Replace these spares for further checking Machine is installed on 27-11-2013 aged up to 10 years 3 months. Repairing cost is 106.666% both criteria RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:  
 checked & identified that the main board & 3 batteries are defective and aged up to 10 years above. its repairing cost economically not feasible & recommended for BER as per BER protocol.  
  
 26/3/14 Signature of JC BM (NHM)


CONSULTANT BIOMEDICAL  
 NATIONAL HEALTH MISSION  
 KANNUR - 670002

Recommending for Condemnation

MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 CHITTARIPARAMBA  
 P. O. KANNOTH-670 650

Date: 30/03/14



  
 Signature of Superintendent/Medical Officer (i/c)



# STOCK BOOK (MEDICINES)

Name of drug Automatic Biochemist Name of drug Analyses

Date	Receipt	Issue	Balance	Date	Receipt	Issue	Balance
	kg./gm./ml.	kg./gm./ml.	kg./gm./ml.		kg./gm./ml.	kg./gm./ml.	kg./gm./ml.
27.11.13			NIL				NIL
24.7.13	Received from KMBCI.						
25.10.13	Accessories:-						
	1. Sine wave UPS				1 NO.		1 NO.
	2. Halogen Lamps				1 No		
	3. Thermal Print Paper				02 No.		
	4. Dust Covers				1 NO.	Issued to	
	5. Variable volume Pipette 10-100ml				1 NO.	Lab	
	6. One fixed volume Pipette 500ml				1 No.		
	Reagents:-						
	1. Urea Mod				500ml x 1 bat		
	2. Creatinine				500ml x 1 bat		
	3. Billirubin TAD				500ml x 1 bat	Issued to	
	4. Cylucose				500ml x 1 bat	Lab	
	5. Cholesterol				500ml x 1 bat		
	6. Q-C. Abnormal				05ml x 1 bat		
	7. Q-C Normal				05ml x 1 bat		
	St. 101 AT 1/8.10.13 RBK				Dy: Rebornik DVT LD. Ch		
	Grand total Rs: 84,321/-				Warranty: 26.11.2016 (3 years)		
27.11.13	Issued to Lab technician.						NIL
	Stock entry page no: 104/vd III						





# IT WORLD

## COMPUTER SALES & SERVICE

1<sup>st</sup> Floor,, Malabar Plaza, Near SBT ATM  
MATTANUR.

Mob: 9744041752, Office: 8089136301

No. **106**

Date 19/3/20

### QUOTATION

To Cyril healthcare Pvt. Ltd Poonithura Erngkulam  
682038

Mob No..... Email : .....

Sl. No.	DESCRIPTION	Qty	Unit	Price
	1) main board UPS	1	1	1500/-
	2) Battery	1	3	3300/-
Total				4800/-

For IT WORLD  
Signature



Model: GXT MT+ 1K  
Input: 208VAC/240VAC, 50/60Hz, 4 BA, 1Φ  
Output: 208VAC/240VAC, 50/60Hz, 1Φ  
Capacity: 1000VA/800W  
16312513091001411  
MADE IN CHINA



**EMERSON**  
Network Power



FUNCTION DISPLAY



**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890615225)  
1345245



**Liebert GXT MT+**  
AC Power



*sh*  
2/4/2024  
MEDICAL OFFICER  
PRIMARY HEALTH CENTRE  
CHITTARIPARAMBA  
P. O. KANNOTH-670 650