


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THRISSUR
2	Name of Hospital	PHC CHOWANNUR
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR Make: N/A Model: NY-10ZW SN: 2106243411086
4	Equipment ID/ Barcode	0843515 #83074
5	Date of purchase/ Year of manufacture/Installation Date	27-07-2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 29-11-2023
9	Action taken	Checked the machine and found that compressor motor, sieve bed and main board defective , need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. Unfortunately, the supplier and installation stock information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	Sinto joy 

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

OEM not found.

  
Signature of JC BM (NHM)  
Junior Consultant (Biomedical)  
NHM, Thrissur

Date:



  
Signature of  
Superintendent Medical Officer (I/c)  
NHM, Thrissur



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 189837

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

<p>Health Facility <u>PHC CHOWANNUR</u></p> <p>Address <u>Chowannur</u> <u>Thrissur</u></p> <p>Ph : <u>9497064304</u></p>	<p>Call Registration Date : <u>29/11/2023</u></p> <p>Caller ID : <u>83074</u></p> <p>Date of Visit : <u>30/11/2023</u></p> <p>Asset No. : <u>0843515</u></p> <p>EQPT Name : <u>Oxygen Concentrator</u></p> <p>Manufacture <u>N/A</u> Model : <u>NY-102W</u></p> <p>S. No <u>2106243411086</u> Dept. <u>Prin. &amp; Palliative</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Machine not switching on

**Action Taken :** When checked the machine and found that sieve bed, Compressor motor and main board defective. Need to replace these spares for further checking

Completed  Date : 30/11/2023 Time : 1:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrlix Engineer	Date	Start Time	End Time
<u>SINTO JOY</u>	<u>30/11/23</u>	<u>10:00AM</u>	<u>1:00pm</u>

**Customer Remark** Completed  Pending


**Service Engineer Name :** SINTO JOY  
**Signature :**   
**Date :** 30/11/23  
**Contact Number :** 7907406353



**Customer Name :**   
**Signature :**   
**Date :**   
**Contact Number :**   
**Designation :**   
**Hospital Seal :**   
*കോടതി കമ്മീഷൻ ഇൻ ചാർജ്  
പ്രൈവറ്റ് ഹോസ്പിറ്റൽ ട്രസ്റ്റ്, ചെമ്മൂട്  
ചെമ്മൂട്, പി.ഒ. സ്റ്റാമ്പ് - 680601*



*Srinivas*



*Srinivas*  
Sr Consultant (Biomedical)  
NHM, Thrissur

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7888  
BAR CODE - (8004-890615229)  
0843515

*Handwritten signature*

Consultant (Biomedical)

**CYRIX HEALTH CARE PVT.LTD**  
**BEMP SERVICE PROVIDER**  
**PREVENTIVE MAINTENANCE SERVICE**  
PMS DONE ON: 31/3/12  
NEXT PMS DUE: 30/11/12  
PERFORMED BY: JH 12

Oxygen Concentrator

Mod. 92W  
Power 0V/50HZ  
Inlet 550W  
Flow rate: 1-10 L/min  
Oxygen purity: 93% ± 3%



Other details can be found in the instructions  
Date of production: fuselage bar code

Model No: 92W-102W  
Date Of Manufacture: 24th. 06. 2021



Serial Number: 2106243411086

*M. D. D.*  
Consultant (Biomedical)  
Sur