


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	PALAKKAD
2	Name of Hospital	PHC PUDUR
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE MAKE : SAMSO MODEL : NA SN : NA
4	Equipment ID/ Barcode	#103246 0943890
5	Date of purchase/ Year of manufacture/Installation Date	04/09/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	25/03/2024 (toll free)
9	Action taken	Checked the machine found load cell and Display damaged. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA




13	Asset Value	RS.1600/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found not working. Found load cell and display defective. Equipment installed on 04/09/2018. aged upto 5 year 7 months. Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Both RBER Criteria met. Hence recommended for BER.

  
3/04/2018  
Signature of ~~IC~~ Biomedical Engg)  
Consultant (NHM) Palakkad  
NHM (Arogyakeralam) Palakkad



Date: 16/04/2018

  
MEDICAL OFFICER  
FAMILY HEALTH CENTRE  
PUDUR, ATTAPPADY  
PIN - 678581

Signature of  
Superintendent/Medical Officer (i/c)





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 179956

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>24x7 P.H.C. PUDUR.</u>	Call Registration Date : <u>25/3/2024</u>
Address <u>PUDUR</u>	Caller ID : <u>103246</u>
<u>PALAKKAD</u>	Date of Visit : <u>26/3/2024</u>
Ph : <u>8075243181</u>	Asset No. : <u>0943890</u>
	EQPT Name : <u>Weighing Machine</u>
	Manufacture <u>SANSO</u> Model : <u>NIL</u>
	S. No. <u>NIL</u> Dept. <u>Op.</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Machine not working

Action Taken : checked and found that load cells (four numbers), display defective. Need to replace these spares for further checking

Completed  Date : 26/3/2024 Time : 11:30 Am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>SREETHA C</u>	<u>26/3/2024</u>	<u>10:00AM</u>	<u>11:30AM</u>

Customer Remark



Completed

MEDICAL OFFICER  
FAMILY HEALTH CENTRE  
PUDUR, ATTAPPADY  
PIN - 678581

Service Engineer Name : SREETHA C  
Signature : [Signature]  
Date : 26/3/2024  
Contact Number : 9072522058

Customer Name : Jayakumar  
Signature : [Signature]  
Date : 26/3/24  
Contact Number : 8075243181  
Designation : SN0.  
Hospital Seal : -

NRI-1  
2020/11/1

# Digital Weighing Machine.

	Rd.	issued	Balance	sq
4/9/18	Rd from store	1	1	
	issued for use	1	1	
20/12/18	Rd from the store	2	3	
14/1/2020	(Found on physical Vag)		3	



*[Handwritten signature]*

MEDICAL OFFICER  
FAMILY HEALTH CENTRE  
PUDUR, ATTAPADY  
PIN - 678561



