


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KANNUR
2	Name of Hospital	CHC IRIKKUR
3	Name of Equipment with Make, Model and Serial Number	CENTRIFUGE Make: DLAB Model: DM0412 SN: ID169AB0000206
4	Equipment ID/ Barcode	1330613 #94560
5	Date of purchase/ Year of manufacture/Installation Date	31-07-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 06-02-2024
9	Action taken	Checked the machine found machine is not switching ON, found that Main Board is defective, enquired OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Main Board-12500/- TOTAL =12500+GST@18% =14750/-





13	Asset Value	23200/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	63.578%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & OEM quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching ON, found Main Board of the machine is Defective. Enquired OEM for quotation. Equipment installed on 31-07-2018 aged up to 5 years 7 months. Repairing cost is 63.578% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:  
 checked & identified that main board issue and aged up to 5~~7~~ years above. As per quotations its service cost economically not feasible. So recommended for BER as per BER protocol.  
  
 02/03/2024 Signature of JC BM (NHM)

Layana. C.V.   
 9947992683  
 Date: 7/2/2024



Signature of  
 Superintendent/Medical Officer (C)  
  
 MEDICAL OFFICER  
 COMMUNITY HEALTH CENTRE  
 KANNUR-670593





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 134134

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC IRIKKUR</u></p> <p>Address <u>IRIKKUR</u> <u>KANNUR</u></p> <p>Ph : <u>9947992683</u></p>	<p>Call Registration Date : <u>6/2/2024</u></p> <p>Caller ID : <u>94560</u></p> <p>Date of Visit : <u>7/2/2024</u></p> <p>Asset No. : <u>1330613</u></p> <p>EQPT Name : <u>Centrifuge</u></p> <p>Manufacture <u>DLab</u> Model : <u>DM0412</u></p> <p>S. No. <u>1D169ab0000206</u> Dept. <u>Laboratory</u></p>
---	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not switching ON

**Action Taken :** checked the machine found machine is not switching ON, found that Main Board is defective, enquired OEM for quotation.

Completed  Date : 7/2/2024 Time : 12:30 pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin</u>	<u>7/2/2024</u>	<u>11:45 am</u>	<u>12:30 pm</u>

**Customer Remark** Completed  Pending

**Service Engineer Name :** Aswin  
Signature :   
Date : 7/2/2024  
Contact Number : 8306012759



**Customer Name :** Layana.C.V.  
Signature :   
Date : 7/2/2024  
Contact Number : 9947992683  
Designation : LT  
Hospital Seal :

**MEDICAL OFFICER**  
**COMMUNITY HEALTH CENTRE**  
**IRIKKUR-670893**

LabX/Q-LFCR-SP/QT 025/2024

Date : 19-02-2024

To,

Cyrix Healthcare PVT Ltd  
30/641 B, Petta Junction  
Poonithura , Kochi  
Kerala-682038

Dear Sir/Madam,

**Subject: Quotation for Spare Parts –CLINICAL CENTRIFUGE (DM0412)**

We are pleased to offer you our quote for the same as under.


Description	Qty	Amount (Rs)	GST@18%	Total Rate (Rs.)
MAIN BOARD	1	12500	2250	14750
<b>Grand Total</b>				<b>14750.00</b>

**TERMS&CONDITIONS**

- **Payment: Payment is to be made in favor of M/S LabX.**
- **Issue Work Order a long with the 100% payment in Advance.**
- **Limited warranty of 120 days from the date of invoice.**

Thank you,

LabX



Authorized Signatory

---

**PRICE ENQUIRY 1330613 #87426 CENTRIFUGE @ CHC IRIKKUR**

7 messages

**Sarang KM** <sarangcyrix@gmail.com>

Fri, Jan 5, 2024 at 2:25 PM

To: servicesupport@labx.in

Cc: anurag p &lt;anuraganu126@gmail.com&gt;, BLESSEN JOSE &lt;zm2.klbemp@cyrix.in&gt;

Dear sir,

Please share the price and availability of MAIN BOARD of CENTRIFUGE @ CHC IRIKKUR please send the quotation as soon as possible

CLINICAL CENTRIFUGE  
MODEL: DM0412  
SN: LD169AB0000206

IMAGES ATTACHED FOR YOUR REFERENCE

--

**Thanks & Regards****Sarang.k.m**

District Incharge | Kannur

Bemp-Kerala

Mobile : **+91 7593847150**Email : [sarangcyrix@gmail.com](mailto:sarangcyrix@gmail.com)**CYRIX HEALTHCARE (P) LTD**

30/641B | Petta | Punithura | Kochi | Kerala- 682038

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Save Environment, Please do not print this e-mail unless it's essential

---

**6 attachments****WhatsApp Image 2024-01-05 at 2.21.47 PM (2).jpeg**  
161K**WhatsApp Image 2024-01-05 at 2.21.47 PM (1).jpeg**  
91K






WhatsApp Image 2024-01-05 at 2.21.47 PM.jpeg  
68K



WhatsApp Image 2024-01-05 at 2.21.46 PM.jpeg  
164K



WhatsApp Image 2024-01-02 at 2.35.34 PM.jpeg  
478K



WhatsApp Image 2024-01-02 at 2.35.33 PM.jpeg  
1607K

---

**Sarang KM** <sarangcyrix@gmail.com>  
To: servicesupport@labx.in  
Cc: anurag p <anuraganu126@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Mon, Jan 8, 2024 at 11:36 AM

Dear sir,

**REMINDER**

Please share the quotation as soon as possible

[Quoted text hidden]

---

**Sarang KM** <sarangcyrix@gmail.com>  
To: servicesupport@labx.in  
Cc: anurag p <anuraganu126@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Wed, Feb 7, 2024 at 4:33 PM

Dear sir,

**REMINDER**

Please share the quotation as soon as possible

[Quoted text hidden]

---

**Sarang KM** <sarangcyrix@gmail.com>  
To: servicesupport@labx.in  
Cc: anurag p <anuraganu126@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Fri, Feb 16, 2024 at 4:01 PM

Dear sir,

REMINDER

Please share the quotation as soon as possible

[Quoted text hidden]

---

Sarang KM <sarangcyrix@gmail.com>

Sat, Feb 17, 2024 at 1:31 PM

To: servicesupport@labx.in

Cc: anurag p <anuraganu126@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Dear sir,

REMINDER

Please share the quotation as soon as possible

[Quoted text hidden]

---

ISD LMS <servicesupport@labx.in>

Mon, Feb 19, 2024 at 10:51 AM

To: Sarang KM <sarangcyrix@gmail.com>

Cc: anurag p <anuraganu126@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Dear Sir,

***Greetings of the day!!!***

We are very grateful to offer the Quotation for the main board of CLINICAL CENTRIFUGE (DM0412) for the customer CHC IRIKKUR. Please find the attached quotation and kindly do the needful at the earliest.

Waiting to hear from you soon.

*Thanks & Regards,*

***Arya M S***

*Resident Engineer - ISD*

*Mob. No.: 94460 04866*

**LabX Medical Systems**

TC 95/1621/1, Opp SFS Tiara Apartments

Kumarapuram, Medical College P.O

Trivandrum - 695011 Kerala

Tel No : 0471-2447639

email: [servicesupport@labx.in](mailto:servicesupport@labx.in)

Web : [www.labx.biz](http://www.labx.biz)

[Quoted text hidden]

---

 **QT 025-Cyrix Healthcare PVT Ltd - CHC IRIKKUR.pdf**  
687K

---

Sarang KM <sarangcyrix@gmail.com>

Mon, Feb 26, 2024 at 11:45 AM

To: avinash t <avinashtcyrix@gmail.com>

FYI

[Quoted text hidden]

---

 **QT 025-Cyrix Healthcare PVT Ltd - CHC IRIKKUR.pdf**  
687K



Layana. cv. @  
LT





PERFORMED BY

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890615225) 1330613

CARE PROVIDER  
ID  
CALC  
CALC  
VENTIVE  
G DONE ON  
PMS DUE: 08/12  
FORMED BY

**DM0412**  
110-240 V- 50/60HZ, 70W  
Cat No: 9034002122  
Ver No: 0906027777  
Ser No:  
I.D169AB0000206



Layana. cv. ~~CP~~  
LT.





