


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendatlons for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	THIYI MANNARKKAD
3	Name of Equipment with Make, Model and Serial Number	BICARBONATE MIXER Make : TUSCAVO(CURA) Model: ABM-FS SN: 45/09/17
4	Equipment ID/ Barcode	0921140 #107178
5	Date of purchase/ Year of manufacture/Installation Date	30-12-2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registrationof complaint through email/ Toll free)	19-04-2024 toll free
9	Action taken	Checked the machine and found SMPS,driver board , valves, main board, transformer damaged . Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS. 60073/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the DEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Regarding the inspection of the equipment, bicarbonate mixer. Upon examination, it has been discovered that critical spare parts, including the Driver board, Main board, Solenoid valve, Flow sensor, Display, and Transformer, are severely damaged. Equipment installed on 30/12/2017 and aged upto 6+ year. In our effort to source replacement parts, we inquired with the original manufacturer, Cura Healthcare. However, it has come to our attention that Cura Healthcare has ceased its operations and has permanently closed down. Consequently, there are no available spares from the manufacturer. Considering the circumstances and in adherence to the tender clause 5.3.14.2, which stipulates that manufacturer not existing, the equipment should be evaluated for condemnation, we recommend that these equipment is marked for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine aged up to 6 years. RBER criteria not met.

Consultant (Biomedical Engg)
NHM (Arogyakeralam), Palakkad


10/5/24
Signature of JC BM (NHM)

Date: 15/5/24





Signature of
Superintending Medical Officer (Ue)
അലൂക്ക് ആസ്ഥാന ആശുപത്രി
മുസ്സാർലാട്



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 207463

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERD Approved Service Agency

Service Report

30/64 1 B, Pelta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrilx.in

Health Facility : <u>THQIL MANNARKAD</u>	Call Registration Date : <u>19/4/2024</u>
Address : <u>MANNARKAD</u>	Caller ID : <u>107178</u>
<u>PALAKKAD</u>	Date of Visit : <u>20/4/2024</u>
Ph : <u>9207617256</u>	Asset No. : <u>0921140</u>
	EQPT Name : <u>Bicarbonate mixer</u>
	Manufacture <u>TUSCANA ABM</u> Model : <u>ABM B</u>
	S. No. <u>45/09/17</u> Dept. <u>Dialysis unit</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not working

Action Taken : checked the machine and found smps, driver board, valves, main board, transformer damaged. need to replace these spare for further checking

Completed Date : 20/4/2024 Time : 2:30pm Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>SREETHA</u>	<u>20/4/2024</u>	<u>11:30pm</u>	<u>2:30pm</u>

Customer Remark : Completed Pending

Service Engineer Name : SREETHA C
Signature : [Signature]
Date : 20/4/2024
Contact Number : 9072522058


Customer Name : [Signature]
Signature : [Signature]
Date : 20/4/24
Contact Number : 9847299500
Designation : NO 1496837567
Hospital Seal : [Seal]



BICARBONATE MIXTURE

Date		Received	Issued	Balance	Main Stock Balance	Sign
30/8/18	Received from store on 30/8/18	1		1	1	
	Power code - 1		2 ✓			
	User manual - 1					
	Issued to Dialysis unit.		1	NIL	1	<u>Amu</u>
	15-					




 സൂപ്രണ്ട്
 താലൂക്ക് ആസ്ഥാന ആശുപത്രി
 മണ്ണാർകട



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
www.telcare.org.in

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0921140




CYRIX
 HEALTHCARE PVT LTD
 S.N: 45 09/17
 CN: TAMP 145
 Alkxer



സുപ്രണ്ട്
 അലൂക്ക് ആസ്ഥാന ആരോഗ്യ
 ഉത്തരവാകാട്