


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC PERINJANAM
3	Name of Equipment with Make, Model and Serial Number	Fetal doppler Make : BPL Model :9613 Sr.No- AZTA7H4266
4	Equipment ID & Barcode	#101212 – 0831830
5	Date of purchase / Year of manufacture /Installation Date	15-07-2011
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14-03-2024
9	Action taken	Checked the machine and found main board , battery and PROBE are complaint. Need to replace the main board, battery and display for further checking and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Nil

13	Asset Value	4000
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and oem EOL letter attached
16	Reasons for recommending the equipment as BER	Checked the machine and found main board, battery and probe complaint. Equipment installed on 15-11-2011 and aged up to 12 YEARS 6 month. oem informed that this machine is discontinued model and spare not available, oem letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:



Repairs not possible



Date:



Signature of JG-BM (NHM)

Superintendent / Medical Officer (i/c)	Signature of seal	Date
		<p>6/4/24</p>

SUPERINTENDENT
Community Health Centre
Perinjanam - 680 686

ASSISTANT SURGEON
Community Health Centre
Perinjanam-680686



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 195057

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <i>C.H.C. Perinjalam</i></p> <p>Address <i>Perinjalam</i> <i>Thiruvananthapuram</i></p> <p>Ph : <i>9495005134</i></p>	<p>Call Registration Date : <i>14-03-2024</i></p> <p>Caller ID : <i>101212</i></p> <p>Date of Visit : <i>15-03-2024</i></p> <p>Asset No. : <i>0831830</i></p> <p>EQPT Name : <i>Fetal dopler</i></p> <p>Manufacture <i>BPL</i> Model : <i>9613</i></p> <p>S. No. <i>727A-744266</i> Dept. <i>Casualty</i></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Equipment not switching on.*

Action Taken : *checked the machine and found main board, battery and probe are complaint need to replace the main board, battery and probe for further checking and spare required to OEM.*

Completed Date : *15-03-2024* Time : *10:00 Am* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<i>Bjo</i>	<i>15-03-2024</i>	<i>9:30 Am</i>	<i>10:00 Am</i>

Customer Remark Completed Pending

<p>Service Engineer Name : <i>Bjo Joy</i></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <i>15-03-2024</i></p> <p>Contact Number : <i>892870010</i></p>	<p>Customer Name : <i>Photo Community Health Centre</i></p> <p>Signature : <i>[Signature]</i></p> <p>Date : <i>[Blank]</i></p> <p>Contact Number : <i>9495005134</i></p> <p>Designation : <i>SNO</i></p> <p>Hospital Seal : <i>[Seal]</i></p>
--	---



SUPERINTENDENT
Community Health Centre
Perinjalam - 680 886

ASSISTANT SURGEON
Community Health Centre
Perinjalam-680686



RINS
MARKETING & SERVICES

City Tower, TM 364 N / 1st Floor
Opp. Chinmaya Balabhavan, Kannur-670 002
Tel: 0497-2704624, Fax: 2709789, Mob: 9946100601, 9447064624
E-mail: rinsmarketing@gmail.com
www.rinsmarketing.com **28/08/2024**

TO,
CYRIX HEALTH CARE PVT LTD,
ERANAKULAM

SUB: End Of Service /Spare support

Dear Sir,

This has specific reference to the repair inspection done by our engineer on your BPL FETAL DOPPLER FD9613 bearing the unit serial number AZTA7H4266

The said equipment is more than 7 years old and the usage exploited from the unit is pretty good. As the machine is getting old the economical usage of the said equipment may not be viable and spare parts may not be available.

Thanking you and assuring you of our best services at all time se

For,
Rins Marketing & Services

Authorized Signatory



**REQUIRED QUOTATION FOR FETAL DOPPLER PARTS -0831829 #101211 -0831830
#101212**

3 messages

Sarang KM <arangcyrix@gmail.com>
To: rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>

Tue, Mar 26, 2024 at 11:06 AM

Dear sir,

Kindly share the quotation for fetal doppler spare parts , details given below

HOSPITAL NAME: CHC PERINJANAM
BARCODE: 0831829 CALL ID: 101211
BARCODE: 0831830 CALL ID 101212
make: BPL
Model: 9613

Required spares : Probe, Battery , Main board

Do the needful as soon as possible.

--

Thanks & Regards

Sarang.k.m
Divisional Incharge|North Zone
Bemp-Kerala
Mobile : **+91 7593847150**
Email : sarangcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com



Save Environment, Please do not print this e-mail unless it's essential

rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>
To: Sarang KM <arangcyrix@gmail.com>

Wed, Mar 27, 2024 at 11:40 AM

Dear Sir,

Sorry for the inconvenience FETAL DOPPLER 9613 parts(probe, battery, mainboard) are not available .

Regards,
Anagha



TM 364 N, 1st Floor, City Tower
Opp. Chinmaya Balabhavan, Kannur-2
PH : 0497-2704624
MOB : 9846008441
MOB :9447064624

[Quoted text hidden]

rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>
To: Sarang KM <sarangcyrix@gmail.com>

Thu, Mar 28, 2024 at 12:35 PM


Please find the attachment.

**Regards,
Anagha**



TM 364 N, 1st Floor, City Tower
Opp. Chinmaya Balabhavan, Kannur-2
PH : 0497-2704624
MOB : 9846008441
MOB :9447064624

[Quoted text hidden]

 **SERVICE LETTER.pdf**
127K

66

Doppler
CNRHM)

67
Ulceme Dilators

	Rel	Cond	Rel		Rel	Cond	Rel
24/3/10							
10/3/10 Rel	1		1	10/3/10	Rel (CNRHM)	5	
15/3/11 Rel (room store)	1		2				
4P Biomedical Engineering					5 scan		
24/3/2022 - C/O to the equipment stock Vol. II Page No: -110				24/3/2022 - C/O to the equipment stock Vol. II Page No: -111			
24/3/2022 - C/O to the equipment stock Vol. II Page No: -110							



