


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	THQH KODUNGALLUR
3	Name of Equipment with Make, Model and Serial Number	UPS I BALL Nirantar UPS-621 EL4314C15273
4	Equipment ID/ Barcode	0821137
5	Date of purchase/ Year of manufacture/Installation Date	06/02/2018
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	22/12/2023
9	Action taken	Checked the equipment and found that the control PCB,battery, transformer are defective.Enquired for spare
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	3086/-
13	Asset Value	4,500/-




14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report and quote attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found that the control PCB, battery, transformer are defective, Equipment installed on 06/02/18 and aged up to 6 years, Spare cost is 68% of asset value and it is beyond economical repairable, We recommend this equipment for condemnation.
17	Signature of CYRIX Authority	Ramdas K 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Board cost is not economical


Signature of JC BM (NHM)
Dr. Consultant

Date


Signature of
Superintendent/Medical Officer (i/c)

Superintendent
Taluk Head Quarters Hospital
Kodungallur





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 167494

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THQH Kodungallur</u></p> <p>Address <u>Kodungallur</u> <u>Tharissur</u></p> <p>Ph : <u>86.06.46.9732</u></p>	<p>Call Registration Date : <u>22/12/2023</u></p> <p>Caller ID : <u>87052</u></p> <p>Date of Visit : <u>23/12/23</u></p> <p>Asset No. : <u>0821137</u></p> <p>EQPT Name : <u>UPS</u></p> <p>Manufacture <u>i Ball</u> Model : <u>Minanta</u></p> <p>S. No. <u>FL4314C15773</u> Dept. <u>Lab</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment not switching on.

Action Taken : Checked the equipments and found transformer control PCB, battery are defective. Enquired spare locally.

Completed Date : 23/12/23 Time : 12:00 Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ramdas</u>	<u>23/12/23</u>	<u>11:00 am</u>	<u>12:00 pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Ramdas</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>23/12/23</u></p> <p>Contact Number : <u>9745520163</u></p>	<p>Customer Name : <u>Limi.KK.</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>23/12/23</u></p> <p>Contact Number : <u>86644732</u></p> <p>Designation : <u>LT</u></p> <p>Hospital Seal :</p>
--	--



[Signature]
Superintendent
Taluk Head Quarters Hospital
Kodungallur

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 27/01/2024

Contact : 8921950217

Email : yp@backups.in

GSTIN: 32CGUPM9450L1ZN

600VA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTH CARE PVT .LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	12 -230 inverter PCB	1	1100.00	1100.00	18	1,298.00
2	7Ah Better y	1	890	890	28	1,139.00
3	12v transformer	1	550	550	18	649.00
PAYABLE AMOUNT						3086.00
Commercial terms and conditions						
1. Payment 80% advance along with PO balance before dispatch, 20% After installation.						
2. Delivery period; 2weeks.						
3. Tax :GST@18% for spares, 28%@Battery						
4. Installation; Our scope.						
Please arrange to release the PO at our above address						
Bank name :HDFC		Account no:50200082743176		IFSC Code:HDFC0004083		
We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.						
Thank you, Sincerely, BHAGATH MOHAN						



Fwd: 600VA UPS SERVICE PROPOSAL

1 message

Thu, Jun 6, 2024 at 11:30 AM


Bijo T Joy <bijocyrix@gmail.com>
To: Sarang KM <arangcyrix@gmail.com>

----- Forwarded message -----

From: **Bhagath Mohan** <aionbackups12@gmail.com>
Date: Thu, 6 Jun, 2024, 11:04 am
Subject: 600VA UPS SERVICE PROPOSAL
To: bijocyrix@gmail.com <bijocyrix@gmail.com>

Thanks & Regards

Bhagath Mohan
PH:-8921950217

 **600VA UPS SERVICE PROPOSAL_27-01-24.pdf**
1007K

UPS

232

UPS	Received	Issue	Balance
62/12 Received from store (for ESR Analyser)	1		1

[Signature]
 P. M.
 Lab Technician gr I



Superintendent
 Taluk Head Quarters Hospital
 Kodungallur

KEY BOARD (for computer)

9/3/200 9/3/	Received from store (1p)	1	-	1	25454
-----------------	--------------------------	---	---	---	-------



Princy
~~Supintendent~~ Princy P.M
Taluk Head Quarters Hospital Lab Technician 301
Kodungallur

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7569
BAR CODE - (8004-890615225)
0821137



[Signature]
Superintendent
Taluk Head Quarters Hospital
Kodungallur

[Signature]
Prize P.M
Lab Technician gr I

OUTPUT 230Vac, 50Hz, 2.22A, 1Ø

INPUT 230Vac, 50Hz, 5AMAX, 1Ø

MODEL: Ball Niranlar UPS-621

EL4314C15273



Panja
Briya . PM
Superintendent Technician JRI
Taluk Head Quarters Hospital
Kodungallur