


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Wayanad
2	Name of Hospital	UPHC Munderi
3	Name of Equipment with Make, Model and Serial Number	Equipment- Stabilizer Make-V-guard Model- N/A Sn- N/A
4	Equipment ID/ Barcode	I240302//Ticket 98883
5	Date of purchase/ Year of manufacture/Installation Date	11/06/2020
6	Warranty details (Yes/No)	No warranty.
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	29/02/2024 /Toll free
9	Action taken	Checked machine, found stabilizer PCB and stabilizer transformer defective. Enquired spare from local vendor
10	Present status of the equipment - (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	V-guard stabilizer pcb-1000/- V-guard stabilizer transformer -1200/- Total=INR 2200/-

13	Asset Value	RS 1799/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	122%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Quotation from local vendor Cyrix service report
16	Reasons for recommending the equipment as BER	Checked machine, found stabilizer PCB and stabilizer transformer defective. Equipment completed 3 years and 8 months. Repairing cost of equipment more than 60 % with respect to asset value both criteria of RBER Met. So we recommended for condemnation as per tender clause 5.3.14.1
17	Name & Signature of CYRIX Authority	NIYAS HADI AP 


Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

PCB and stabilizer transformer defective.
Recommended Condemnation


 Signature of JC(BM) (NHM)
 National Health Mission
 (Arogyakeralam)
 Wayanad-673122


 Date.....

Date: 25/3/2024

For M.O

 Signature of
 Superintendent/Medical Officer (i/c)

Medical Officer (NUHM)
UPHC Munderi



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 201801

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>UPHC</u></p> <p>Address <u>Munday</u> <u>Wayanad, Kerala.</u></p> <p>Ph : <u>7356070020</u></p>	<p>Call Registration Date : <u>29/02/2024</u></p> <p>Caller ID : <u>98883</u></p> <p>Date of Visit : <u>01/03/2024</u></p> <p>Asset No. : <u>1240302</u></p> <p>EQPT Name : <u>Stabilizer</u></p> <p>Manufacture <u>V.GUARD</u> Model : <u>N/A</u></p> <p>S. No. <u>N/A</u> Dept. <u>LABORATORY</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Unwanted sound from the unit.
Not working properly.

Action Taken : Checked and found board and transformer complaint. Need to replace the same.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ajnu.p</u>	<u>01/03/24</u>	<u>8:30AM</u>	<u>9:40AM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Ajnu.p</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>01/03/24</u></p> <p>Contact Number : <u>7593847137</u></p>	<p>Customer Name : <u>Athiba PG</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>01/03/24</u></p> <p>Contact Number : <u>9747621775</u></p> <p>Designation : <u>Lab technician</u></p> <p>Hospital Seal : </p>
--	--



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 181751

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>UPHC</u></p> <p>Address <u>Mundery</u> <u>Wayanad, Kerala.</u></p> <p>Ph : <u>7356070020</u></p>	<p>Call Registration Date : <u>29/02/24</u></p> <p>Caller ID : <u>98883</u></p> <p>Date of Visit : <u>04/03/24</u></p> <p>Asset No. : <u>1240302</u></p> <p>EQPT Name : <u>Stabilizer</u></p> <p>Manufacture <u>V Guard</u> Model : <u>VA50</u></p> <p>S. No. <u>N/A</u> Dept. <u>LABORATORY</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not getting on.

Action Taken : Checked the machine, found stabilizer board and transformer defective. Need to replace the same.

Completed Date : 04/03/24 Time : 10:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Arjun.P</u>	<u>04/03/24</u>	<u>9:00 AM</u>	<u>10:00 AM</u>

Customer Remark _____ Completed Pending

Service Engineer Name : Arjun.P
Signature : [Signature]
Date : 04/03/24
Contact Number : 7593847137



Customer Name : Aswathi
Signature : [Signature]
Date : 9/3/24
Contact Number : 7902468039
Designation : Lab Tech.
Hospital Seal :

[Signature]
Medical Officer (NHM)
UPHC Mundery
Lab Technician
UPHC Mundery

HYBRID SOLAR

Near KFWW Board, Kalpetta Wayanad

Mob: 8089386389, 9745772211

HS/147/23-24
01/04/2024

TO
THE CYRIX HEALTH CARE
KOCHI

Attn: Sir,
Ref :

SUB: -

Ref: - As Per Your Enquiry Notice No:

We are pleased to quote our UPS & Battery Sales Price as below

Sl.No	Description	Rate	Qty	Amount
1	V-GURD STABILIZER PCB	1000.00	1	1000.00
2	V-GURD STABILIZER TRANSFORMER	1200.00	1	1200.00
	INCLUDING SERVICE CHARGE			
TOTAL				2200.00

Yours truly
For Hybrid Solar
Kalpetta
Wayanad
E-mail: hybridsolar8911@gmail.com

Dr JJS





Stabilizer service

1 message

Anoop PK <hybridsolar8911@gmail.com>
To: niyascyrix@gmail.com

Mon, 1 Apr 2024 at 10:48 am

STABILIZER

DATE	PARTICULARS	RECEIVED	ISSUED	BALANCE	Remarks
------	-------------	----------	--------	---------	---------

6-11-20	Received from the flow	1	1	1	
---------	------------------------	---	---	---	--

Dr. Jiji

Medical Officer (NCHM)
UPHC Munderi





URBAN PRIMARY HEALTH
 Date.....
 MANDERI, KALPETT

2/2/25
 2/2/25

Medical Officer (NUTM)
 UPHC Murala