


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Wayanad
2	Name of Hospital	PHC Vengapally
3	Name of Equipment with Make, Model and Serial Number	Equipment-Weighing machine Make-Samsco Model-na Sn-na
4	Equipment ID/ Barcode	1240847//Ticket 98532
5	Date of purchase/ Year of manufacture/Installation Date	06-07-2018
6	Warranty details (Yes/No)	NO
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	28/02/2024 /Toll free
9	Action taken	Checked machine, found load sensor and display board defective. Enquired spare from local vendor
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Load sensor for samsco weighing scale -1416/- Display board -448/- Total=INR 1864/-

13	Asset Value	RS 1600-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	116.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Quotation from local vendor Cyrix service report
16	Reasons for recommending the equipment as BER	Checked machine, found load sensor board and display board defective Equipment completed 5 years Repairing cost of equipment is more than 60 % with respect to asset value Both criteria of RBER Met. So we recommended for condemnation as per tender clause 5.3.14.1 .
17	Name & Signature of CYRIX Authority	NIYAS HADI AP 


Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Load sensor board and display board are defective. Recommend Condemnation


 Signature of JC BM (NHM)
 Biomedical Engineer
 National Health Mission
 (Trogva Pralam)
 Wayanad 673122

Date: 25/3/24




 Signature of
 Superintendent/Medical Officer (i/c)
 National Health Mission
 Wayanad 673122



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/098

CYRIX

HEALTHCARE PVT LTD

No. : 181753

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERS Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>PHC</u>	Call Registration Date : <u>28/02/2024</u>
Address : <u>Vengappally</u>	Caller ID : <u>98532</u>
<u>Wayanad, Kerala</u>	Date of Visit : <u>29/02/2024</u>
Ph : <u>8921738203</u>	Asset No. : <u>1240847</u>
	EQPT Name : <u>Weighing machine</u>
	Manufacture : <u>SANCO</u> Model : <u>Adult</u>
	S. No. : <u>N/A</u> Dept. : <u>Subcentre</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not getting on.

Action Taken : Checked and found load sensor and board complaint. Need to replace the same.


Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrlix Engineer	Date	Start Time	End Time
<u>Ajan P</u>	<u>29/02/24</u>	<u>8:30AM</u>	<u>9:30AM</u>

Customer Remark : Completed Pending



Service Engineer Name : <u>Ajan P</u>	Customer Name : <u>Sjils V</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>29/02/24</u>	Date : <u>29/2/24</u>
Contact Number : <u>8593847137</u>	Contact Number : <u>8921738203</u>
	Designation : <u>N/A</u>
	Hospital Seal :

QUOTATION



Ospio Healthcare Kannadikal, calicut, Kerala, GSTIN: 32AZLPT7342N1ZP Mobile: 9447338356	Quotation No. 136	Quotation Date 26/03/2024	Expiry Date 25/05/2024
	BILL TO CYRIX HEALTHCARE PVT LTD Address: Poonithura, Ernakulam , GSTIN: 32AAFCC2499H2ZM State: Kerala		

S.NO.	ITEMS	QTY.	RATE	TAX	AMOUNT
1	Load sensor for samsco digital scale	1 PCS	1,200	216 (18%)	1,416
2	Display for weigh scale	1 PCS	400	48 (12%)	448
TOTAL		2		₹ 264	₹ 1,864

HSN/SAC	Taxable Value	CGST		SGST		Total Tax Amount
		Rate	Amount	Rate	Amount	
-	400	6%	24	6%	24	₹ 48
-	1,200	9%	108	9%	108	₹ 216

Total Amount (in words)
 One Thousand Eight Hundred Sixty Four Rupees

Terms and Conditions
 1. Goods once sold will not be taken back or exchanged
 2. All disputes are subject to [ENTER_YOUR_CITY_NAME] jurisdiction only



Invoice created using



Handwritten signature and text: **Job**
 കൂടുതൽ വിവരങ്ങൾക്ക് ക്ഷേത്രം കോമ്പ്, മിഡൽ.
 പി.ഒ. തെക്കൂന്നി, വയനാട്.
 ഫോൺ: 673 121



Request for spare quotation

3 messages

Niyas Hadi AP <niyascyrix@gmail.com>
To: ospiohealthcare@gmail.com

Tue, 26 Mar 2024 at 11:55 am

Dear Sir
Please send the quotation for following spare of Samsco digital weight machine
1-load sensor and board
2-dispay of weighing machine

Thanks & regards

NIYAS HADI
District incharge
BEMP project
Wayanad
kerala
mob -7593847171

email -niyascyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

ospio healthcare <ospiohealthcare@gmail.com>
To: Niyas Hadi AP <niyascyrix@gmail.com>

Tue, 26 Mar 2024 at 12:10 pm

I will send it soon.
[Quoted text hidden]

ospio healthcare <ospiohealthcare@gmail.com>
To: Niyas Hadi AP <niyascyrix@gmail.com>

Mon, 1 Apr 2024 at 3:25 pm

Dear Sir/Madam,

Please refer to the attached DOC.

Thanks & Regards,

OSPIO HEALTHCARE

11/502, KRISHNA KRIBA BUILDING,
PARAMBIL KADAVU, KANNADIKKAL,
VENGERI (P.O), CALICUT -673010
MOB: 919447338356

[Quoted text hidden]



Handwritten signature and text in Malayalam, including the phone number 673 121.


Name of Items: Weighing Scale

Date	From whom Report	Report	Issue	Balance	CoA	Signature	Remarks
22/6/23	found on physical exam			2			Infant - 1
2/8/23	Transferred to Unservicell Registers Page no - 6 Issue to clinic			1			Infant weighing Scale Not working actinax -

Name of Items: Hub center

Date	From whom Report to whom Issue	Report	Issue	Balance	CoA	Signature	Remarks
22/6/23	found on physical exam			1			Not working
2/8/23	Transferred to Unservicell Registers Page no - 6			1	0		
5-7-23	- no file	1		1			Hub center NS800 Manual CP.
7-7-23	Issue to FHC Clinic		1	0			




 ജി. അനീസ്
 മെഡിക്കൽ ഓഫീസർ
 കുടുംബ ആരോഗ്യ കേന്ദ്രം, നെൽക്കുളം
 പി.ഒ. നെൽക്കുളം, നയം
 ഫോൺ: 675 121