


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	CHC ALANALLUR
3	Name of Equipment with Make, Model and Serial Number	CENTRIFUGE MAKE : REMI MODEL : C-854/8 SN : NA
4	Equipment ID/ Barcode	0931191 # 107079
5	Date of purchase/ Year of manufacture/Installation Date	01/12/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	18/04/2024 (toll free)
9	Action taken	Checked the machine found that electric motor, speed regulator, tube s.s 15ml damaged. Enquire OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Electric motor = 3313/- Knob for speed regulator = 58/- Tube s.s 15 ml = 195/-  TOTAL = 3313/-


13	Asset Value	RS.5000/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.2%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine found that electric motor, speed regulator, tube s.s. 15ml damaged. Enquire OEM for quotation. Equipment installed on 01-12-2018 aged upto 5 years 3 months. Repairing cost is 66.2% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P 

\*Not mandatory      #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine aged up to 5 years. Spare cost is the 66.2% of asset value. Hence recommended for BER.

  
Signature of JC S. J. Joshi  
Consultant (Biomedical Equip)  
NHM (Anagardhan), Palitana

Date: 23/5/24



  
Signature of  
Superintendent/Medical Officer (I/c)  
PALITANA DISTRICT





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 207486

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrilx.in

### Service Report

<p>Health Facility ..... <u>CHC ALANALLUR</u>.....</p> <p>Address ..... <u>ALANALLUR</u>.....</p> <p>..... <u>PALAKKAD</u>.....</p> <p>Ph : ..... <u>9562922468</u>.....</p>	<p>Call Registration Date : ..... <u>18/4/2024</u>.....</p> <p>Caller ID : ..... <u>107079</u>.....</p> <p>Date of Visit : ..... <u>19/4/2024</u>.....</p> <p>Asset No. : ..... <u>0931191</u>.....</p> <p>EQPT Name : ..... <u>CENTRIFUGE</u>.....</p> <p>Manufacture ..... <u>REMI</u>..... Model : <u>C 854 18</u></p> <p>S. No. ..... <u>NA</u>..... Dept. <u>Laboratory</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... Machine not working.....

**Action Taken :** ..... checked the machine and found that electric motor, speed regulator, tube 5.5.15ml damaged. Need to replace these spares for further checking......

Completed  Date : 19/4/2024 Time : 11:30AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>SREETHAN C</u>	<u>19/4/2024</u>	<u>9:30AM</u>	<u>11:30AM</u>


**Customer Remark** .....  Completed  Pending

<p><b>Service Engineer Name :</b> <u>SREETHAN C</u></p> <p><b>Signature :</b> <u>[Signature]</u></p> <p><b>Date :</b> <u>19/4/2024</u></p> <p><b>Contact Number :</b> <u>9072522058</u></p>	<p><b>Customer Name :</b> <u>[Signature]</u></p> <p><b>Signature :</b> <u>[Signature]</u></p> <p><b>Date :</b> <u>19/4/2024</u></p> <p><b>Contact Number :</b> <u>9562922468</u></p> <p><b>Designation :</b> <u>[Signature]</u></p> <p><b>Hospital Seal :</b> <u>[Seal]</u></p>
---	---







  
SUPERINTENDENT  
COMMUNITY HEALTH CENTRE  
ALANALLUR  
PALAKKAD D.STRIC



REMI SALES & ENGINEERING LIMITED



70/1750,5th Cross Road , Link Park ,Pottakuzhi Road, Pachalam P.O, Kochi-682012  
 Phone : 0484-2401361,Email Id.:instservicekochi@remigroup.com/service@remilabworld.com

Corporate Off : Remi House, 3rd floor, 11 Cama Ind. Estate, Goregaon (East) Mumbai 400 063. INDIA  
 • Board: +91 22 4058 9888 SALES SUPPORT NO.8879448682 • Email Id.: sales@remilabworld.com

QUOTATION CUM PROFORMA INVOICE

QN :15

Complaint No.	Customer Name & Address.:					Quotation No	PRINTED		
Dated	4/5/2024	CYRIX HEALTHCARE (P) LTD					Date	4/5/2024	
ENGINEER NAME	Mpin	30/6418, Petta, Punnthura Ernakulam- 682038					REF	Phone	
REF.	Phone	7593847154					DATE	4/5/2024	
Email	akhileshcyrix@gmail.com	AKHILESH P							
Model no.	C-254/s	SR NO. /MFG DT.							
Sr No.	Description	Part code	HSN/SAC	Rate Rs.	Qty	Amount Rs. (A)	GST % (B)	GST Amt (C)	Total Amt (A+C)
1	ELECTRIC MOTOR ( RT )	ML25000154-	850140	2808	1	2808	18%	505	3313
2	KNOB FOR SPEED REGULATOR ( NRT )	L22000116	392390	49	1	49	18%	9	58
3	TUBES.S. 15 ML(NRT )	L22000496	721990	165	1	165	18%	30	195
								TOTAL ( D )	3313
CALIBRATION CHARGES		0	0	0	0	0	0	0	0
SERVICE CHARGES		0	0	0	0	0	18%	0	0
								TOTAL ( E )	3313
Rupees in Words:	Rs. Three thousand three hundred and thirteen only .								3313

# Terms & Conditions:

- Spares are suitable only to use with REMI instruments & not for any other make.
- We will accept return of spares only if they are unused and not damaged. At our discretion, the returned spare will be replaced. The difference in prices, if any, will have to be borne by you.
- To ensure supply of correct spares, we suggest you to send us defective spare in advance with your request for which you want to purchase new spares.
- 100% Advance along with your order confirmation.
- Payment can be made by Cheque/ NEFT in our Bank account. For Cash Payment you must collect Receipt from recipient of cash falling which REMI will not be held responsible for any discrepancy.
- Courier/ transportation charges are payable at destination & not included in our price.
- Validity: The quotation is valid for 30 days from the date of this quotation / Proforma Invoice.
- Subject to Ernakulam Jurisdiction only.

REMARKS : PAYMENT 100% ADVANCE

For REMI SALES & ENGINEERING LTD



# BANK DETAILS FOR REMI SALES & ENGINEERING LTD.

- A/C NO: 35699134671
- IFSC CODE: SBIN0013223
- GST NO: 32AAACR0421M125
- Bank Name: STATE BANK OF INDIA
- MICR CODE: 682002053
- PAN NO: AAACR0421M
- BRANCH: ELAMAKKARA
- SWIFT CODE: SBININBB395
- CIN NO. U31100MH198PL022314

---

**REQUIRED QUOTATION FOR CENTRIFUGE MACHINE 0931191 #107079 @ CHC ALANALLUR**

---

**Akhilesh P** <akhileshcyrix@gmail.com>  
To: Sarang KM <sarangcyrix@gmail.com>

Thu, Jun 6, 2024 at 12:58 PM

Thanks &amp; Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

----- Forwarded message -----

From: **instservicekochi** <[instservicekochi@remigroup.com](mailto:instservicekochi@remigroup.com)>  
Date: Mon, 6 May, 2024, 11:54 am  
Subject: FW: FW: REQUIRED QUOTATION FOR CENTRIFUGE MACHINE 0931191 #107079 @ CHC ALANALLUR  
To: <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>  
Cc: Raman <[raman@remilabworld.com](mailto:raman@remilabworld.com)>, REMI Group - Kochi - Vishnu S K <[kochisales.hcd@remigroup.com](mailto:kochisales.hcd@remigroup.com)>, <[kochi@remigroup.com](mailto:kochi@remigroup.com)>

Dear Sir,

Please find the attached PERFORMA INVOICE for your reference. **PLEASE RELEASE THE PAYMENT IN ADVANCE.**

Regards

REMI KOCHI

INST DIV.

**From:** Akhilesh P <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>  
**Sent:** Monday, May 6, 2024 10:06 AM  
**To:** [instservicekochi@remigroup.com](mailto:instservicekochi@remigroup.com)  
**Cc:** REMI Group - Kochi - Vishnu S K <[kochisales.hcd@remigroup.com](mailto:kochisales.hcd@remigroup.com)>  
**Subject:** Re: FW: REQUIRED QUOTATION FOR CENTRIFUGE MACHINE 0931191 #107079 @ CHC ALANALLUR

Dear sir

Kindly update the status.

Thanks &amp; Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

On Sat, 4 May, 2024, 11:00 am Akhilesh P, <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)> wrote:

Dear sir

Gently Reminder Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

On Thu, 2 May, 2024, 4:53 pm Akhilesh P, <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)> wrote:

Dear mam

PFA

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

On Thu, 2 May, 2024, 4:01 pm instservicekochi, <[instservicekochi@remigroup.com](mailto:instservicekochi@remigroup.com)> wrote:

Dear sir,

Please share this spare image .

1. POTTENSIOMETER
2. TRANSFORMER
3. SAMPLE CUP HOLDER

Regards

Rabitha

**From:** Akhilesh P <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>

**Sent:** Thursday, May 2, 2024 3:57 PM

**To:** [Instservicekochi@remigroup.com](mailto:instservicekochi@remigroup.com)

**Subject:** Fwd: REQUIRED QUOTATION FOR CENTRIFUGE MACHINE 0931191 #107079 @ CHC ALANALLUR



Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

----- Forwarded message -----

From: **Akhilesh P** <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>  
Date: Thu, 2 May, 2024, 10:58 am  
Subject: Re: REQUIRED QUOTATION FOR CENTRIFUGE MACHINE 0931191 #107079 @ CHC ALANALLUR  
To: <[kochi@remigroup.com](mailto:kochi@remigroup.com)>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

On Wed, 1 May, 2024, 1:37 pm Akhilesh P, <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)> wrote:

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

[Quoted text hidden]