


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	DH WADAKKANCHERY
3	Name of Equipment with Make, Model and Serial Number	Diathermy Delcatt 250-U
4	Equipment ID/ Barcode	0812315
5	Date of purchase/ Year of manufacture/Installation Date	28/11/2007
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	07/12/2023
9	Action taken	Checked the equipment and found Main board defective. Enquired spare
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	NA
13	Asset Value	10,505/-


14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report ,OEM Mail copy attached
16	Reasons for recommending the equipment as BER	Checked equipment and found that Main board defective. Enquired for spare from OEM.Equipment installed on 28/11/2007 and aged up to 16 years.OEM informed that this model is obsolete,end of life letter issued,As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Signature of CYRIX Authority	Ramdas K 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Obselete model.



 Jr. Consultant (Biomedical)
 NHM Thrissur
 Signature of JC BM (NHM)



Date *12/11/2024*

SUB INTENDENT
 DISTRICT HOSPITAL, WADAKKANANCHERY
 KUMARANELLUR P.O. THRISSUR-680 590

 Signature of
 Superintendent/Medical Officer (i/c)



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 166791

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>DH Wadakkancherry</u>	Call Registration Date : <u>7/12/23</u>
Address <u>Trissur</u>	Caller ID : <u>84363</u>
Ph : <u>9946623074</u>	Date of Visit : <u>8/12/2023</u>
	Asset No. : <u>0812315</u>
	EQPT Name : <u>CAUTERY</u>
	Manufacture <u>DELTA MEDICAL</u> Model : <u>DELCONT</u>
	S. No. <u> </u> No <u> </u> Dept. <u>OT</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Cutting, coagulation not working

Action Taken : checked the equipment found that cautery machine main board defective. Required spare

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrrix Engineer	Date	Start Time	End Time
<u>Tejas</u>	<u>8/12/23</u>		

Customer Remark Completed Pending

Service Engineer Name : Tejas
Signature : [Signature]
Date : 8/12/23
Contact Number : 8848098402



Customer Name : SURESH K P
Signature : [Signature]
Date : 8/12/23
Contact Number : 9447299500
Designation : PHYSICIAN
Hospital Seal : 04886235274



0812315 Main Board Price enquiry

3 messages

Ramdas K <ramdascyrix@gmail.com>
To: Deepak Gala <galadeepak7@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>

Thu, 21 Dec, 2023 at 11:51 pm

Dear Deepak,

Kindly send me the quotation of the main board for delcatt 250 Diathermy.

Hospital Name - DH wadakkanchery
Department - OT
Make&Model - Delcat 250 U
S/n - Not available(Faded)

With Regards,

Ramdas K

DISTRICTS ENGINEER THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia M:-9745520163 | E:- ramdacyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Ramdas K <ramdascyrix@gmail.com>
To: Deepak Gala <galadeepak7@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>

Mon, 1 Jan, 2024 at 10:42 pm

Dear Sir,

Waiting for your reply.

With Regards,

Ramdas K

Biomedical ENGINEER THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038

Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

| E:- ramdascyrix@gmail.com

| W:- www.cyrixhealthcare.com | "Genius in Healthcare"

[Quoted text hidden]

Deepak Gala <galadeepak7@gmail.com>
To: Ramdas K <ramdascyrix@gmail.com>

Fri, 5 Jan, 2024 at 12:44 pm

To,
Ramdas K
Cyrix Healthcare Private Limited
Kerala

Dear Sir,



SUPERINTENDENT
DISTRICT HOSPITAL, WADAKKANCHERY
KUMARAN KURUP
THRISSUR-680 590
21/5/24
9-1-24

We have received your mail and noted that Surgical Diathermy Delcatt - 250B model is very old and we have stop manufacturing that model, so we are not having its part i.e., Mains Board.

Hence we are not able to quote or supply mains board for the same.

Thanking you,

[Quoted text hidden]

Regards,
Mr. Deepak Gala
Delta Medical Corporation
108, Damji Shamji Industrial Estate,
Mahakali Caves Road, Chakala,
Andheri (e), Mumbai-400 093
Tel no. 022-26873762/9820228951
Email: galadeepak7@gmail.com
Website: www.deltamedicorp.com

Please consider our environmental responsibility before printing this e-mail or any other print - Save Paper, Save Woods, Save Trees.....




*N.P. 21st
9-1-24*

SUPERINTENDENT
DISTRICT HOSPITAL, WADAKKANCHERY
KUMARANELLUR P.O. THRISSUR-680 590

69

Dra Themy Surgical Unit (Canton)

Date	Particulars	Receipt	Issue	Balance	Sign
8/9/22	Tr in from Main OT Dra Themy Surgical unit (Canton) Stock Book Volume II (180) (Delta Medical apparatus) Chir polas Dra Themy, Rada Rd date 28/11/2007		1	1	
8/9/22	Issued to minor OT		1	Nil	



~~K.P.~~
 2/1/24
 9-1-24
 SUPERINTENDENT
 DISTRICT HOSPITAL WADAKKANCHERY
 KUMARANELLUR P.O. THRISSUR-680 590



R.P.
9/1/24
SUPERINTENDENT
DISTRICT HOSPITAL, KUMARANELLUR,
KUMARANELLUR PO THRISSUR-680 590



K. A. Q. M.
9-1-24

SUPERINTENDENT
DISTRICT HOSPITAL, WADAKKANCHERY
KUMARANELLUR P.O. THRISSUR-680 590