


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	RPH LAB KOZHIKODE
3	Name of Equipment with Make, Model and Serial Number	Equipment : Electrolyte Analyzer Make: Meizhou Cornley Hi Tech Co.Ltd Model: AFT 300 SN:DJ05A11110019
4	Equipment ID & Barcode	109952 & 1110142
5	Date of purchase / Year of manufacture /Installation Date	15-02-2016 (Actual installation date is 03-07-2012,stock book attached for reference)
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	07-05-2024
9	Action taken	Checked and found found Main board ,Multiplexer (6 way) Na Electrode, K Electrode and Reference Electrode defective. Enquired spares with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	1)Main Board -1 nos = 70500/- 2) Multiplexer(6 way) -1 nos= 24750/- 3) Na Electrode -1 nos = 9250/- 4) K Electrode 1 nos = 9250/- 5) Ref. Electrode 1 nos =9250/- TOTAL =123000/- GST =22140/- GRAND TOTAL =145140/-

13	Asset Value	95238/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	152.39 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor quotation attached
16	Reasons for recommending the equipment as BER	Checked and found found Main board , Multiplexer (6 way) Na Electrode, K Electrode and Reference Electrode defective. In software installation date is 15-02-2016, As per the stock book machine is installed on 03-07-2012 Equipment out lived 11 years 10 month. Repairing cost of the equipment is 152.39% both criteria for RBER met, As per tender clause 5.3.14.1 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	JINEESH PP 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that main board (multiplexer) (6 way) NA electrode, K electrode and reference electrode defective. Machine installed on 15/2/2012. So recommended for condemnation.

Consultant Bio Medical Engineer
National Health Mission
Signature of JC BM (NHM)
Kozhikode



Seal

Signature of 

Superintendent Medical Officer (IC)

കോഴിക്കോട്, പിൻ: 673 032

Date



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 173243

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>RPH LAB</u></p> <p>Address <u>Kozhikode</u></p> <p>Ph : <u>9446385087</u></p>	<p>Call Registration Date : <u>7/05/2024</u></p> <p>Caller ID : <u>109952</u></p> <p>Date of Visit : <u>08/05/2024</u></p> <p>Asset No. : <u>1110142</u></p> <p>EQPT Name : <u>Electrolyte Analyser</u></p> <p>Manufacture <u>Meizhou Cornley Hi-Tech Co.Ltd</u> Model : <u>AFT-300</u></p> <p>S. No. <u>DJ05A11110019</u> Dept. <u>Laboratory</u></p>
-----------------------------------------------------------------------------------------------------	--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : → NO CALA reasprate.

Action Taken : checked and found that main board multiplexer assembly, kt electrode, Nat electrode and referene electrode are defatine. need to replace these spares for further checking

Completed Date : 8/5/2024 Time : 12-30pm Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Jineesh PP</u>	<u>8/5/2024</u>	<u>9 AM</u>	<u>12-30 pm</u>

Customer Remark Completed Pending

Keenal
നിയമിത മെന്റേണൽ ഓഫീസർ
സർവ്വീസ് ഓഫീസ് ഹെൽത്ത് പ്ലാനിംഗ്
കോർപ്പറേഷൻ, പിൻ: 673 032

Service Engineer Name : Jineesh PP
Signature : [Signature] Sr. BME
Date : 8/05/2024
Contact Number : 8943527662

Customer Name : [Signature]
Signature : [Signature]
Date : 08/5/2024
Contact Number : 9446385087
Designation : Scientific officer
Hospital Seal :



Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



ISO 9001:2015
CERTIFIED COMPANY

Proforma Invoice No.: RDPL/S/AD/20240516/24-24			Date : 16-05-2024		
To, Cyrix Healthcare Pvt. Ltd. 30/641B, Petta, Punithura Poonithura, Ernakulam-682038, India			Ref. : By Mail Date : 16-05-2024 Ref.RPH LAB,Kozhikode		
Sr. No	Cat No.	Description	Price Each Ex- Work	Qty	Total (In Rs.)
		(Instrument Model: Electrolyte Analyzer) (Model :AFT-300) (S/NO.DJ05A11110019)			
A		Spare & Consumables Procurement Details:			
(I)		Main Board	70,500.00	1	70,500.00
(II)		Multiplexer (6 Way)	24,750.00	1	24,750.00
(III)		Na Electrode	9,250.00	1	9,250.00
(IV)		K Electrode	9,250.00	1	9,250.00
(V)		Ref. Electrode	9,250.00	1	9,250.00
Total					1,23,000.00
(+) GST @ 18%					22,140.00
Total: One Lakh Fourty Five Thousand One Hundred Fourty Only.					1,45,140.00

Terms & Conditions : { Validity: 90 Days } , {Delivery : 3-4 Weeks}, {Payment Terms : 100% Advance }

Bank Account Details :-

Name of the Firm	: Aspen Diagnostic Pvt. Ltd.	Aspen Diagnostics Pvt. Ltd.
Bank Name	: ICICI Bank Limited	
Account No.	: 016005003239	
Account Type	: Current Account	
MICR Code	: 110229029	
IFSC Code	: ICIC0000160	



Required Spare Quotation -1110142-109952- Electrolyte Analyzer-RPH LAB,Kozhikode

5 messages

Jineesh P P <jineeshcyrix@gmail.com>

Sun, May 12, 2024 at 10:55 PM

To: binto@rdgc.in

Cc: service@rdgc.in, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear Sir

Please share price and spare availability of main board, multi plexer assembly ,K+ electrode ,Na+ electrode and reference electrode.

Electrolyte Analyzer

Make : Meizhou cornley Hi Tech co.Ltd.

Model: AFT -300

SN: DJ05A11110019

Thanks & Regards,

Jineesh PP

District Incharge - Kozhikode

Mob-7593847163

jineeshcyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Jineesh P P <jineeshcyrix@gmail.com>

Tue, May 14, 2024 at 1:26 PM

To: binto@rdgc.in

Cc: service@rdgc.in, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear Sir,

Reminder

[Quoted text hidden]
[Quoted text hidden]

Jineesh P P <jineeshcyrix@gmail.com>

Thu, May 16, 2024 at 12:18 PM

To: binto@rdgc.in

Cc: service@rdgc.in, aspenkochi@gmail.com, Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

[Quoted text hidden]

service support <servicesupport@rdgc.in>

Thu, May 16, 2024 at 3:52 PM

To: jineeshcyrix@gmail.com

Cc: BINTO JOSEPH <binto@rdgc.in>, rustam@rdgc.in, aspenkochi@gmail.com, Rajaram.rdg <rajaram.k@rdgc.in>

Dear sir,

Please find the attached P.I. & share the payment details .

Thanks & Regards

Hari Ram

[Quoted text hidden]

Jineesh P P <jineeshcyrix@gmail.com>

Thu, May 16, 2024 at 8:55 PM

To: Sarang KM <sarangcyrix@gmail.com>



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)


TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
1110142



Supplied and Maintained by
KERALA MEDICAL SERVICES CORPORATION LTD.
(Department of Health & Family Welfare, Govt. of Kerala)
Thycaud P.O. Thiruvananthapuram - 696014

Under
Tender No.
Eqpt Code



ELECTROLYTE ANALYZER

POWER SUPPLY: 100V-240V~

FREQUENCY: 50/60Hz

RATED POWER: 100VA

Type **AFT-300**




SN **DJ05A11110019**

2011.11.28

CAUTIONS:
100V-240V ~ Inside. Turn off power before opening the back cover.

Meizhou Cornley Hi-Tech Co., Ltd.
Add: 6/F Block E, Dahan Science & Technology Estate, XII,
Nanshan, Shenzhen, Guangdong, China 518055
Tel: 86-755-66330908
Email: sales@cornley.com

Fax: 86-755-66330962
Website: WWW.CORNLEY.COM



കേരളം പബ്ലിക് ഹെൽത്ത് ഡെപാർട്ട്മെന്റ്
ലാബോറട്ടറി
കേരളം പബ്ലിക് ഹെൽത്ത് ഡെപാർട്ട്മെന്റ്
ലാബോറട്ടറി
കേരളം പബ്ലിക് ഹെൽത്ത് ഡെപാർട്ട്മെന്റ്
ലാബോറട്ടറി