

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	CHC CHALISSERY
3	Name of Equipment with Make, Model and Serial Number	DIGITAL BP APPARATUS Make : OMRON Model: NA SN: NA
4	Equipment ID/ Barcode	0931064 #111618
5	Date of purchase/ Year of manufacture/Installation Date	25-02-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	17-05-2024 (toll free)
9	Action taken	Checked the machine and found main board, air pump damaged . Need to replace air pump and main board for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS. 1323/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipments as BER	Checked the machine found air pump and main board damaged. Equipment installed on 25/02/2018 aged upto 6 year 2 months. Quotation not submitted since spare parts are not available on market. So we are recommended the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILESH P


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine aged up to 6 years. Air pump and main board damaged. Quotation not submitted.


 Signature of Consultant (Biomedical Engg) NHM (Arogya) Palakkad (NHM)

Date:



Signature of Superintendent / Medical Officer (I/c) Community Health Centre, Chalissery, Palakkad 679536



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 178884

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

Health Facility <u>CHC CHALISSERY</u>	Call Registration Date : <u>17/05/2024</u>
Address <u>PALAKKAD</u>	Caller ID : <u>111618</u>
Ph : <u>9388191096</u>	Date of Visit : <u>18/05/2024</u>
	Asset No. : <u>0931064</u>
	EQPT Name : <u>BP Apparatus digital</u>
	Manufacture <u>OMRON</u> Model : <u>VA</u>
	S. No. <u>VA</u> Dept. <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked the machine and found main board, air pump damaged. need to replace air pump and main board for further checking

Completed **Date :** 18/05/24 **Time :** 11:00AM **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Tejas</u>	<u>18/05/24</u>	<u>10:00AM</u>	<u>11:00AM</u>



Customer Remark Completed Pending

Service Engineer Name : Tejas
Signature : [Signature]
Date : 18/05/24
Contact Number : 9848098499

Customer Name :
Signature :
Date :
Contact Number :
Designation :
Hospital Seal :



Tejas
Dr. Jayaraj
17/05/24
0948261236
16/05/24
Community Health Centre
Chalissery, Palakkad - 686501



Sox

SUPERINTENDENT
Community Health Centre
Chalissery, Palakkad - 678 636