


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	DH WADAKKANCHERRY
3	Name of Equipment with Make, Model and Serial Number	Ocular Ultrasound Make - Biomedix Model - Echo Rule Pro S/N - 310143
4	Equipment ID/ Barcode	#94863 - 0812450
5	Date of purchase/ Year of manufacture/Installation Date	29-09-2014
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	08-02-2024
9	Action taken	Checked the equipment and found probe is defective. spare enquired to OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	NAME OF SPARE - PROBE 1,31,040.00

13	Asset Value	1,51,525/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	86.46%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached And Quotation attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found that probe is defective. The machine installed on 20-09-2014 and aged up to 9 years 5 month. The repair cost is 86.46% and both criteria met for RBER. As per tender clause 5.3.14.1 we are recommending equipment for condemnation,
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph




Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Probe cost is vey high


Signature of JC BM (NHM) 

Date: 21/03/2024




21/03
SUPERINTENDENT.
DISTRICT HOSPITAL, WADAKKANCHERY
Signature of 
Superintendent/Medical Officer (i/c) 



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 179031

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility DH Inlada Mhonessey</p> <p>Address Thriessur</p> <p>Ph : 944 7805900</p>	<p>Call Registration Date : 08/02/2024</p> <p>Caller ID : 94863</p> <p>Date of Visit : 9/02/2024</p> <p>Asset No. : 0812480</p> <p>EQPT Name : AScan, ocular ultra sound</p> <p>Manufacture Biomedica Model : Scholale Pro</p> <p>S. No. 5101A3 Dept. Eye OP</p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the equipment found that Prob defective
Engined for spare from OEM

Completed **Date :** 9/02/24 **Time :** 12:00 Pm **Spare Required**

Spare Replaced **Requested**


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Tejas	9/02/24	11:00Am	12:00Pm

Customer Remark **Completed** **Pending**

<p>Service Engineer Name : Tejas</p> <p>Signature : </p> <p>Date : 9/02/2024</p> <p>Contact Number : 884809422</p>	<p>Customer Name : </p> <p>Signature : </p> <p>Date : 9/2/24</p> <p>Contact Number : </p> <p>Designation : LAY SECRETARY & TREASURER</p> <p>Hospital Seal : DISTRICT HOSPITAL MADAKKANCHERY KUMARANELLUR P.O, THRISSUR-680 590</p>
--	--



	<p style="text-align: center;">BioMedix Optotechnik & Devices Pvt. Ltd. No-258 Amar jyoti Layout, Domlur. Bangalore-560071</p>	<p style="text-align: center;">BMOD/F/023</p>
---	--	---

QUOTATION

Date: 22.02.2024

Ref: - BMOD/HO/SER/24-25/119
BMOD –CIN : U33112KA1993PTC014968
GST No : 29AAACB7372B1ZU

Customer Details:

Name : Cyrix Health Care Pvt Ltd
Address : Cochin

Equipment – Biomedix A-Scan Probe

SL No.	Particulars	Qty	Amt. Per Unit	Amount Rs
1	Towards A-Scan Probe of Echorule Pro	1	Rs.1,17,000.00	Rs.1,17,000.00
			Total	Rs.1,17,000.00
			GST@12%	Rs.14,040.00
			Final Price	Rs.1,31,040.00

TERMS :-

Taxes : GST @12%

Payment: 100% in Advance

Validity : This quotation valid for 30 days from the date of quotation

The above price offer is based on current custom duty tariff. Any change in custom duties will invalidate above price offer.


If any other parts need to replaced, the same can be done on the basis of the revised estimate.

For BioMedix Optotechnik & Devices Pvt. Ltd.,



(Authorized Signatory)

(Signature & Seal of the customer)

	<p>BioMedix Optotechnik & Devices Pvt. Ltd. No-258 Amar jyoti Layout, Domlur. Bangalore-560071</p>	<p>BMOD/F/023</p>
---	--	-------------------

QUOTATION

Date: 22.02.2024

Ref: - BMOD/HO/SER/24-25/119
 BMOD -CIN : U33112KA1993PTC014968
 GST No : 29AAACB7372B1ZU

Customer Details:

Name : Cyrix Health Care Pvt Ltd
 Address : Cochin

Equipment – Biomedix A-Scan Probe

SL No.	Particulars	Qty	Amt. Per Unit	Amount Rs
1	Towards A-Scan Probe of Echorule Pro	1	Rs.1,17,000.00	Rs.1,17,000.00
			Total	Rs.1,17,000.00
			GST@12%	Rs.14,040.00
			Final Price	Rs.1,31,040.00

TERMS :-

Taxes : GST @12%

Payment: 100% in Advance

Validity : This quotation valid for 30 days from the date of quotation

The above price offer is based on current custom duty tariff. Any change in custom duties will invalidate above price offer.

If any other parts need to be replaced, the same can be done on the basis of the revised estimate.

For BioMedix Optotechnik & Devices Pvt. Ltd.,



(Authorized Signatory)

(Signature & Seal of the customer)



(Signature)
 LAY SECRETARY & TREASURER
 DISTRICT HOSPITAL WADAKKANCHERY
 KUMARANELLUR P.O, THIRISSUR-680 500

Fwd: Quotation-Echorule Pro A scan probe for Cyrix Health Care Pvt Ltd, Cochin

1 message

Ramdas K <ramdascyrix@gmail.com>
To: Sarang KM <arangcyrix@gmail.com>

Thu, Jun 6, 2024 at 3:32 PM

----- Forwarded message -----

From: **Ajit Srivastava** <Servicesupport@biomedixdevices.com>
Date: Thu, Feb 22, 2024, 4:41 PM
Subject: Quotation-Echorule Pro A scan probe for Cyrix Health Care Pvt Ltd, Cochin
To: <ramdascyrix@gmail.com>, <tejasvcyrix@gmail.com>
Cc: P Vijaykumar <Vijayakumar@biomedixdevices.com>, Sanjay Pai <sanjaypai@biomedixdevices.com>, BioMedix Service <Service@biomedixdevices.com>

Dear Customer,

Greetings from Biomedix Optotechnik & Devices Pvt Ltd.

Thank you for giving us an opportunity to serve you.

As per your discussion with Mr.Vijay here with this mail we are forwarding the spare estimation.

Kindly find the attached quote for the same & provide your approval/PO along with Advance payment, so that will proceed further.

Also here attached our bank account details for payment purposes.

For further assistance, please do email. I would be delighted to help you.


--

Thanks & Regards

Ajit Srivastava

service coordinator

BioMedix Optotechnik & Devices Pvt Ltd,
258,Amarjyoti Layout Domluru
Bangalore -560071
Ph : 080-49332900, Ext No 114

2 attachments **Quote-119-Cyrix Health Care Pvt Ltd.pdf**
482K **BOB Bank Details (1).pdf**
273K

8024

22

A Scan-Axiomeli

-/NO: 20-9-2011

Sr. 31014 3 model Biomed/ech.

Rs. 1,83,570.00.



SECRETARY & TREASURER
DISTRICT HOSPITAL WADAKKANCHERY
KUMARANELLUR P.O., THIRISSUR-680 590

BioMedix



CYRIX HEALTH CARE PVT.LTD
 BEMP SERVICE PROVIDER
 PREVENTIVE MAINTENANCE SERVICE
 PMS DONE ON _____
 NEXT PMS DUE _____
 PERFORMED BY _____

CYRIX HEALTH CARE PVT.LTD
 SERVICE MAINTENANCE
 UNIT - SERVICE DIVISION, KAMBEL
 TOLL FREE : 1800 425 000 4
 A HALL OF ALL HEALING THINGS - 12, 30, 40 & 50, 7th
 FLOOR, 100 Feet Road, Kambel, Kambel
 41 1111 - 405547

DISTRICT HOSPITAL, MADAKKANCHERY
 KUMARANELLUR, P.O., THRISSUR-680 590

ECHORULE PRO
 CYRIX
 PMS



ECHORULE PRO

Serial No : 310143



15V/4A

DANGER - Risk of explosion if used in the presence of flammable atmosphere.

FUSE

Rating : 230V/3A

Size : 5x20mm



BioMedix
Optotechnik & Devices Pvt. Ltd.
#258, Amarjyothi Layout, Domlur,
Bangalore - 560 071 .INDIA

Made in India

LAY SECRETARY & TREASURER.
DISTRICT HOSPITAL, NADAKKANCHERY
PIN CODE: 680-590

Handwritten notes and a red-bordered label on the top of the device. The label includes the text "TOLL FREE - 1800 425 7669".

ECHORULE PRO

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
HEALTHCARE PVT LTD

TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)
0212450



Handwritten signature and official stamp of the District Health Officer, Kandy, Sri Lanka.

