


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

1	Name of District	KANNUR
2	Name of Hospital	TALUK HOSPITAL PERAYOOR
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS ELKO, EL510 CANC 20190416EL-51001002
4	Equipment ID/ Barcode	13 20079
5	Date of purchase/ Year of manufacture/Installation Date	28-5-2019
6	Warranty details	No
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	19-10-2023
9	Action taken	Checked the equipment & found not working. Found switch mainboard, motor are defective
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Details of service required (Recommendations for repair)	Not recommended
12	Cost of spares (specify parts and cost)	—


13	Asset Value	2200
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	90%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Service report attached
16	Reasons for recommending the equipment as BER	checked the equipment Found not working, checked and Found Switch main board, motor are defective. Equipment is installed in 25/8/19 and aged more than 4 years comparing the lifespan used the repairing & replacement cost are beyond economical. So recommend for condemnation as per MSEL tender clause 5.3.14.2
17	Signature of CYRIX Authority	shyamkrishnan pph  25/10/23

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that main board, motor defective. Aged up to 4 years. Hence cost economically not feasible. So recommended for BER.


28/10/2023.

Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

Mohananik. 

Date 25/10/2023




Signature of
Superintendent Medical Officer (Bc)
TALUK HOSPITAL
PERAVOOR
KANNUR
670673



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 189531

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.ki@cyrix.in

<p>Health Facility : THQH PERAVOOR.</p> <p>Address : Peravoor KANNUR</p> <p>Ph : 8595 463 755</p>	<p>Call Registration Date : 19/10/2023</p> <p>Caller ID : 75763</p> <p>Date of Visit : 20/10/2023</p> <p>Asset No. : 1320079</p> <p>EQPT Name : BP APPARATUS</p> <p>Manufacture : ELKO Model : EL510.</p> <p>S. No. : CYN20A0416EL Dept. : Subcentre 51001082.</p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Mainboard, Motor, Switch are defective.

Action Taken : checked the equipment and found Mainboard, Motor, Switch are found defective. Needed to replace the spares for further checking of the equipment.

Completed **Date** : 20/10/2023 **Time** : 11.30AM. **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Sayanth	20/10/2023	10.30 Am	11.30 Am.


Customer Remark : Completed Pending

Service Engineer Name : Sayanth
Signature : *Sayanth*
Date : 20/10/2023
Contact Number : 7593847147

Customer Name : M...
Signature : *M...*
Date : 20/10/2023
Contact Number : 9446682048
Designation : Health Inspector
Hospital Seal : TALUK HOSPITAL PERAVOOR KANNUR 670671




ELKO[®]



SYS
mmHg

DIA
mmHg

PUL.
/min







START
STOP

MEM

EL-510

Product Name: Digital Blood Pressure Monitor

Model NO.: EL-510

Rated Input: 6.0V  IP21   

Made in P.R.C.

Imported and marketed by:
Elex Surgical & Allied Products LLP
29, Bhagat Singh Market, New Delhi-110001, INDIA

 SN: 1201201610EL-51001281

STOCK BOOK OF ~~2000~~ Other's.....Name of Article No: Subcentre Kit

Date	No. and date of Voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of receiver	Remarks
		Brought forward					
09-8-2019		Received from ^{Superintendent} TH Peravoor.					
	1	BP apparatus Digital		1			
	2	Dustbin - Black	1		1		
	3	Dustbin - Red	1		1		
	4	Dustbin - Yellow	1		1		
	5	Examination table (LED)	1		1		
	6	Fibroscope (Sakuma)	1		1		
	7	Gloves Sterile -	25				
	8	Glucose strips.	300				
	9	Glucometer (ear-cuff type)	1		1		
	10	Hemometer test strips	900				
	11	Hemometer (True Hb)	1		1		
	12	Subcutter	1		1		
	13	Infantometer (200PS)	1		1		
	14	Kit for testing residual chlorine Darding water	1		1		
	15	Lancets (one cell)	100				
	16	measuring tape	1				
	17	nebulizer	1				
	18	Oxygen cylinder	1	Issue to Labour room			
	19	Oxygen cylinder spare	1	Issue to Labour room			
	20	Oxygen cylinder bracket	1	Issue to Labour room			
	21	Oxygen flow meter	1	Issue to Labour room			
	22	Self testing kit for iodine	5				
	23	Stadiometer	1				
	24	Stethoscope (super deluxe)	1				
	25	Thermometer (Digital)	1				
		Carried Over					



Am
SUPERINTENDENT
TALUK HOSPITAL
PERAVOOR
KANNUR
670 673

Smitha Mol PK
SPHN



AR MEDICARE
ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL: armedicare07@gmail.com
MOB: 7994654108, 9446057350

OUR REF : AR/QTN/113-2023
DATE : 25/10/2023

TO
CYRIX HEALTHCARE
ERNAKULAM

Respected Sir,

SUB: QUOTATION FOR ELKO DIGITAL BP APPARATUS PARTS

Kindly note with reference to the discussions had with yourself regarding the above mentioned subject, we are pleased to submit our very special competitive offer for your reference and further perusal.

In case of any further clarifications kindly feel free to contact us.

Thanking you and assuring you of our best attention and engineering services at all times.

Waiting for your earliest reply.

Yours truly,

For AR MEDICARE,
KANNUR.



TRUST US WITH YOUR NEEDS

OUR REF : AR/QTN/113-2023

DATE : 25/10/2023

SLNO:	DESCRIPTION	TOTAL(INCL.ALL TAXES)
01	ELKO DIGITAL BP APPARATUS MAINBOARD	1200/-
02	MOTOR	600/-
03	SWITCH	200/-
	TOTAL	2,000/-

TERMS & CONDITIONS OF SALE

Payment Terms : Complete Payment Within 30 Days

Delivery : Within 4 Days from the date of purchase order .



AR MEDICARE

TRUST US WITH YOUR NEEDS

QUOTATION

1 message

ar medicare <armedicare07@gmail.com>
To: Sarang KM <sarangcyrix@gmail.com>

Sat, Oct 28, 2023 at 9:07 AM

Sir,
PFA**AR MEDICARE****Thanks & Regards****AR MEDICARE**

Room No:865,2nd floor

AL-RAHA Arcade

Thazhe Chovva, Kannur

Mob: 9446057350, 7994654108, 8138091108**Email: armedicare07@gmail.com****GSTIN :32BDYPV5030P1ZP****ARQ113. CYRIX DIGITAL BP THQH PERAVOOR.pdf**

482K