


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC MAYYIL
3	Name of Equipment with Make, Model and Serial Number	STERILIZER Make: REICO Model: NA SN: NA
4	Equipment ID/ Barcode	1331125#109525
5	Date of purchase/ Year of manufacture/Installation Date	29-03-2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 04-05-2024
9	Action taken	Checked the Equipment found Equipment Blasted Element, Gasket, Autoclave knob, Inner Chamber, Outer Chamber, Autoclave Door Damaged, Enquired OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Elements For Autoclave- 800/- Autoclave Knob-180/- Autoclave gasket- 875/- Inner chamber- 30000/- Outer chamber- 13795/- Auto clave code wire- 850/- Autoclave door- 13500/- Total- 60000/-


13	Asset Value	49,491/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	121.234%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment Blasted Element, Gasket, Autoclave knob, Inner Chamber, Outer Chamber, Autoclave Door Damaged Enquired OEM for quotation Equipment installed on 29-03-2017 aged up to 7 years 2 months. Repairing cost is 121.234% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

*Not mandatory #Based on the period of life and value as per the BER guidelines

^ Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


checked & identified that machine blasted, gasket, door, knob are physically damaged. inner chamber, out chamber also damaged. Aged up to 7 years above. The repairing cost 80% above the asset value. So recommended for BER as per BER Protocol.

Signature of JE BM (NHM)

15/05/2024

**REGIONAL BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002**

Date: 23.05.24




Signature of
Superintendent/Medical Officer (i/c)
മെഡിക്കൽ ഓഫീസർ
സാമൂഹികാരോഗ്യ കേന്ദ്രം
കണ്ണൂർ



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :
202292

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038; Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC</u></p> <p>Address <u>CHC MAYYIL</u> <u>KANNUR</u></p> <p>Ph : <u>9656199072</u></p>	<p>Call Registration Date : <u>04.05.2024</u></p> <p>Caller ID : <u>109525</u></p> <p>Date of Visit : <u>06.05.2024</u></p> <p>Asset No. : <u>1331125</u></p> <p>EQPT Name : <u>Sterilizer</u></p> <p>Manufacture <u>Reico</u> Model : <u>Reico</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Sterilizer not working.

Action Taken : checked the Equipment Found Equipment Blasted Element, Basket, Auto cleve knob, Inner chamber, Outer chamber, Auto cleve Door. Damaged Enquired OEM for quotation.

Completed Date : 06/05/24 Time : 11AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ABDULLA</u>	<u>06/05/24</u>	<u>10AM</u>	<u>11AM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>ABDULLA</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>06/05/2024</u></p> <p>Contact Number : <u>6282953548</u></p>	 <p>Customer Name : <u>Rashmi-S.U</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>6/05/2024</u></p> <p>Contact Number : <u>[Number]</u></p> <p>Designation : <u>NSG Officer ungraded</u></p> <p>Hospital Seal :</p>
---	---



Sreeraj Surgicals

GST No: 32ABYF53674H1ZL
DI No : 13/023/208/2011
13/024/218/2011
Dt.: 29-01-2011

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002
Ph: 0497 2706307. Mob: 9349846307, 9387529848
Email: sreerajsurgical@gmail.com, rajeshsreerajsurgical@gmail.com
www.sreerajsurgical.com Follow us on  www.facebook/ sreerajsurgicals

Date:

ESTIMATE BILL

DATE: 15-05-24

TO,
CYRIX HEALTH CARE
THRIPOONITHARA
ERNAKULAM

SIR,

THANK YOU FOR YOUR ENQUIRY TO THE RATE OF THE FOLLOWING ITEMS.

SL.NO.	ITEMS	RATE
1	ELEMENTS FOR AUTO CLAVE	800.00
2	AUTO CLAVE KNOB	180.00
3	AUTO CLAVE GASKET	875.00
4	INNER CHAMBER	30000.00
5	OUTER CHAMBER	13795.00
7	AUTO CLAVE CODE WIRE	850.00
8	AUTO CLAVE DOOR	13500.00
TOTAL		60000.00

GST INCLUDED

THANKING YOU



Authorised Dealer:



Suppliers of Surgical instruments, Hospital Furnitures, X-ray Materials, Lab Equipments & Disposables



Autoclave quotation required

2 messages

avinash t <avinashtcyrix@gmail.com>
To: sreerajsurgicals@gmail.com
Cc: Sarang KM <sarangcyrix@gmail.com>

Wed, 15 May, 2024 at 2:53 pm

Dear sir,
Please provide quotation of below mentioned parts

1. Element for autoclave
2. Autoclave knob
3. Autoclave gasket
4. inner chamber
5. outer chamber
6. code wire
7. autoclave door

rajesh vk <rajeshsreerajsurgical@gmail.com>
To: avinash t <avinashtcyrix@gmail.com>

Mon, 27 May, 2024 at 12:54 pm

[Quoted text hidden]

Autoclave quotation required

Inbox



avinash t 15 May

to sreerajsurgicals, Sarang



Dear sir,

Please provide quotation of below mentioned parts

1. Element for autoclave
2. Autoclave knob
3. Autoclave gasket
4. inner chamber
5. outer chamber
6. code wire
7. autoclave door



rajesh vk 12:54 pm

to me



Show quoted text

Sreeraj Surgicals

GST No: 32ABYF53674H1ZL
DI No : 13/023/208/2011
18/024/218/2011
Dt.: 29-01-2011

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002
Ph: 0497 2706307. Mob: 9349846307, 9387529848
Email: sreerajsurgicals@gmail.com, rajeshsreerajsurgical@gmail.com
www.sreerajsurgical.com Follow us on www.facebook/ sreerajsurgicals



Date:

ESTIMATE BILL



CYRIX.jpg





Handwritten signature in green ink.

Handwritten text in purple ink: **ബോധകരം എമ്മിൾ**
പി.എസ്.എസ്. റവീന്ദ്രൻ

Official circular stamp of the Kerala State Police (കേരളം പൊലീസ്) with the number 670 652.

Name of Items : Autoclave B/P Page No. _____ of Vol. _____

Date	No. and date of invoice	From whom received / To whom issued	Received	Issue	Balance	Batch No.	Expri. Date	Manu- facture	Cost
20/6/16	Received date:	BF from old sub-stock registers			2				9603/-
27.09.10	(NRHM)	voll no. II Pg no. 240 Rs - 9603/-							
28.3.16	(LP)	Sreejay Surgicals. phone - 93498246307.							9250/-
28/3	28/3/17	Ret from stock	1		3				Rs 23000/-
2017	Bill No-	2725 dated 24-3-17. Sreejay. Surgicals ICNR.							
28317	Invoice N.	phone - 9387529848. put in use							
27/3/19	50/1469 2018-19 B	Vertical steam jacketed.						Universal Agencies.	
11/22-1-19.		Small model - ESTA - 106						Traissus - 650001	
		make - ESTEEM INDUSTRIES.	1		4			01 872440319	Rs 87,700/-
								01 872427031.	
								mobile 8943341569.	
								RINS CL - 'thy cad' TV. P.O	
								70130932/19.	
11/12/19	Ret date 27/3/19	In. out to dentals op							
		op istock: register	1		3				Rs
12/12/20		condemned and disposed							
		as per order no (MS) No. 62/2013							
		H S Food dt 22/2/2013 and B 575/20							
11	17.11.20	no CHC May	2		1 210			(1) Barcode	671324 48
								(2) Barcode	671324 55



Handwritten signatures and official stamps in blue ink, including a date stamp '20/12/2019' and other illegible text.

Dentals