



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 183543

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>L.S. NOORANAD</u>	Call Registration Date : <u>01.05.2024</u>
Address <u>NOORANAD</u>	Caller ID : <u>108951</u>
<u>ALAPPUZHADU</u>	Date of Visit : <u>09.05.2024</u>
Ph : <u>8289899628</u>	Asset No. : <u>0411756</u>
	EQPT Name : <u>WAX BATH</u>
	Manufacture <u>BMS</u> Model : <u>WAXBATH</u>
	S. No. <u>0000</u> Dept. <u>PHYSIOTHERAPY</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT WORKING

Action Taken : CHECKED THE MACHINE FOUND THAT HEATING ELEMENT AND THERMOSTAT ARE FAULTY MACHINE IS RUSTED CONDITION AND WAX LEAKING.

Completed Date : 16.05.2024 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.	-	-	-
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>SAUBAN BIN YASEEN</u>	<u>16.05.24</u>	<u>10:00am</u>	<u>12:00pm</u>

Customer Remark Completed Pending
Machine cannot be repaired request for condonance

Service Engineer Name : SAUBAN
Signature : [Signature]
Date : 16.05.2024
Contact Number : 9746859470

Customer Name [Signature]
Signature : [Signature]
Date : 16/5/24
Contact Number : [Signature]
Designation : [Signature]
Hospital Seal :




SUPERINTENDENT
Leprosy Sanatorium
Nooranad

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	LS NOORANAD
3	Name of Equipment with Make, Model and Serial Number	WAX BATH Make :- BMS Model :- NA SN :- NA
4	Equipment ID & Barcode	108951 & 0411756
5	Date of purchase / Year of manufacture /Installation Date	20/03/2017
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	01/05/2024
9	Action taken	Checked the machine found Heating element and Thermostat are faulty. Machine is rusted condition and wax is leaking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair. (Machine is rusted and body is leaking)
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	RS:- 7425/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT
16	Reasons for recommending the equipment as BER	Checked the machine found heating element & Thermostat are faulty. Machine is rusted condition and wax is leaking. Wear and Tear. The equipment installed on 20/03/2017 aged up to 7 years. The machine is full of rusted, body is leaking and irreparable Condition. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	SAUBAN BIN YASEEN 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER


 23/05/24

JYOTHISH THOMAS
Signature of JC BM (NHM)
NHM AT

Date




Seal



Signature of Superintendent / Medical Officer (M.O.)

SUPERINTENDENT
Leprosy Sanatorium
Noorankal

WIA* BOTH APPARATUS

Date	nature of transaction	Amount	Receipt	Issue	Balance	Remarks
30/3/14	Received from store M.S.R page 31 Vol IV	Rs-7425/-	1		1 0	
	(07)(014)					




 SUPERINTENDENT
 Leprosy Sanatorium
 Nooranad

