


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	DH KANNUR
3	Name of Equipment with Make, Model and Serial Number	HAEMATOLOGY ANALYZER Make: MINDRAY Model: BC 3000 PLUS SN: RJ22110574
4	Equipment ID/ Barcode	1312665 #94585
5	Date of purchase/ Year of manufacture/Installation Date	04.05.2015 (Actual installation date is 22.05.2012, stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 06-02-24
9	Action taken	Checked the machine found machine is not switching ON, found that WBC bath with ruby, RBC Bath with Ruby, 3 way valve, 2 way valve, Analog Board, Volumetric Board, Sample syringe and TFT display are defective. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Wbc bath with ruby-60,000/- Rbc bath with ruby-50,000 /- 3 way valve-10,000/- 2 way valve-10,000/- Analog Board-60,000/- Volumetric Board-14,480/- Sample Syringe-25,500/- TFT Display-30,000/- TOTAL= 2,59,980+18%GST = 3,06,776.40/-


13	Asset Value	225714/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	135.913%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & OEM quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine found machine is not switching ON, found that WBC bath with ruby, RBC Bath with Ruby, 3 way valve, 2 way valve, Analog Board, Volumetric Board, Sample syringe and TFT display are defective. Enquired OEM for quotation. In software installation date is 04.05.2015 As per the stock book Machine is installed on 22.05.2012 aged up to 12 years 10 months. Repairing cost is 135.913% both criteria RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that machine not switching on . 3 wave valve, 2 way valve, Board like volumetric, Analog are in faulty. TFT & syringe are defective. Aged upto 12 years above. As per quotation from OEM the recommending for BER. and also met the BER protocol.

 Signature of JC BM (NHM)
02/04/2024

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

 R. K. Rave
JUNIOR SCIENTIFIC OFFICER
D H KANNUR

Date: 24-04-24

 DISTRICT HOSPITAL
KANNUR - 670 017

 Superintendent
DISTRICT HOSPITAL
KANNUR (c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 201274

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : DH Kanner</p> <p>Address : Kanner Kerala</p> <p>Ph : 9633383806</p>	<p>Call Registration Date : 06/02/24</p> <p>Caller ID : # 94585</p> <p>Date of Visit : # 07/02/24</p> <p>Asset No. : # 1312665</p> <p>EQPT Name : # Haematology</p> <p>Manufacture Minelray Model : BC 3000plus</p> <p>S. No. R522110574 Dept. Lab</p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : # machine not switching on

Action Taken : # Checked the machine found machine is not switching on, found that WBC bath with ruby, RBC with ruby, 3 way valve, 2 way valve, analog board, volumetric board, sample syringe and IFT display are defective
engined oem for quotation

Completed Date : 07/02/24 Time : 12:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			


Cyrrix Engineer	Date	Start Time	End Time
Shebin V S	07/02/24	11:00 am	12:30 pm

Customer Remark Completed Pending

<p>Service Engineer Name : Shebin V S</p> <p>Signature : </p> <p>Date : 07/02/24</p> <p>Contact Number : 7593847149</p>	<p>Customer Name : JSD KANNUR</p> <p>Signature : </p> <p>Date : 07/02/24</p> <p>Contact Number : 9633383806</p> <p>Designation : SUPERINTENDENT</p> <p>Hospital Seal : </p>
---	---



Quotation

Golden Harvest Industrie(Spares) Admin Off: 1st Floor, New No 83, Old No 33, New Boag Road, Opp. Khannammapet Corp. School, T Nagar, Chennai - 600017. Ph: 9282111674 Regd. Off: 8/c, Jawaharlal Nehru Rd, Ekkattuthangal Chennai - 97 DL No 4940/MIII/20-B, 4821/MIII/21-B GST Provisional ID: 33ACVPG7238F2ZE ARN NO: AA330217006488C GSTIN/UIN: 33ACVPG7238F2ZE State Name: , Code :		Invoice No. GHI/23-24/Q01041		Dated 15-Feb-24			
		Delivery Note:		Mode/Terms of Payment 100% Advance			
		Supplier's Ref. GHI/23-24/Q01041		Other Reference GST			
		Buyer's Order No. By Phone		Dated 15-Feb-24			
Buyer (Bill to) Cyrix Healthcare Private Limited Ernakulam, FIRST FLOOR,PETTA JUNCTION 30/641B POONITHURA,ERNAKULAM-682038,INDIA,sales.kl2@ cyrix.in,9744555073 GST No : 32AAFCC2499H2ZM State Name : Kerala, Code : 32		Despatch Document No. DIRECT		Delivery Note Date			
		Despatched through Professional		Destination			
		Terms of Delivery At Your Place					
SI No.	Description of Goods	HSN/ SAC	Qty	Rate	per	Disc. %	Amount
1	Mindray BC-3000+ Rbc bath with ruby	90279090	1	₹ 50,000.00	Nos		₹ 50,000.00
2	Mindray BC-3000+ Wbc bath with ruby	90279090	1	₹ 60,000.00	Nos		₹ 60,000.00
3	Mindray BC-3000+ 3 way valve	90279090	1	₹ 10,000.00	Nos		₹ 10,000.00
4	Mindray BC-3000+ Analog Board	90279090	1	₹ 60,000.00	Nos		₹ 60,000.00
5	Mindray BC-3000+ Volumetric Board	90279090	1	₹ 14,480.00	Nos		₹ 14,480.00
6	Mindray BC-3000+ 2 way valve	90279090	1	₹ 10,000.00	Nos		₹ 10,000.00
7	Mindray BC-3000+ 10ml Syringe	90279090	1	₹ 25,500.00	Nos		₹ 25,500.00
8	Mindray BC-3000+ Display	90279090	1	₹ 30,000.00	Nos		₹ 30,000.00
							₹ 2,59,980.00
	IGST 18%						₹ 46,796.40
	Total		8 Nos				₹ 3,06,776.40
Amount Chargeable(in words):							E. & O.E
Indian Rupees Three Lakh Six Thousand Seven Hundred SeventySix Paise Forty Only							
HSN/SAC		Taxable Value	Integrated Tax		Total Tax Amount		
			Rate	Amount			
90279090		2,59,980	18%	46,796	46,796		
Total		2,59,980		46,796	46,796		
Tax Amount (in words) : Indian Rupees FourtySix Thousand Seven Hundred NinetySix Paise Forty Only							
Company Bank Details Bank Name : HDFC Bank [12162000001380] A/C No : 12162000001380 Branch & IFS Code : Padma Complex & HDFC0001216							
Company's VAT TIN : 33840902845 Company's CST No : 717439 Company's Service Tax No. : ACVPG7238FSD002 Company's PAN : ACVPG7238F		for Golden Harvest Industrie(Spares)					
Declaration We declare that this invoice shows the actual price of the goods described and that all particulars are true and correct							
						 Authorised Signatory	



Haematology Analyser quotation

2 messages

avinash t <avinashtcyrix@gmail.com>
To: Jithin U <jithin.uu@gmail.com>

Fri, 7 Jun, 2024 at 9:26 am

Dear sir,
please provide DH kannur Haematology Analyser quotation.

Regards
Avinash T
District charge, kannur
Cyrix Health care

Jithin U <jithin.uu@gmail.com>
To: avinash t <avinashtcyrix@gmail.com>

Fri, 7 Jun, 2024 at 9:31 am

Sent from my iPhone

> On 7 Jun 2024, at 9:26 AM, avinash t <avinashtcyrix@gmail.com> wrote:

>

>

[Quoted text hidden]

STOCK REGISTER OF MEDICINES/EQUIPMENTS

Name of Item:

HEALTH CARE ANALYSER

Date	From whom received or to whom issued	Receipt	Issue	Balance	Batch No	Expiry Date	Price	Remarks
7.2.08	Received from Store BC 3000 PLUS BAN 3027 Price Rs. 10000	1		1				
27.3.08	Received from store Model AL 820C972 Lablife 1130 P/Manure Price Rs. 31500	1		2				Condensed
22.5.12	HEALTH CARE ANALYSER MODEL NO BC 3000 PLUS SERIAL NO RJ 2210574 FANP. Code: 55-103 Price Rs. 223354/- With 2-year warranty	1		3				Store stock book Pg. no 35/11/14
27.6.15	Condensed as per order no. HC/10821/2015 dated 27.6.15 (AL 820C972 Lablife)		1					
5.5.17	Received from store (KMSCL supply) with Accessories SI NO: A60061701000 23EH- warranty 3yrs from 5/5/17 VSP with Back up CMO 21178005 Installed on 5/5/17 Service Engineer Name: Equipment Code: 55 Model: PEG000	1						
5.4.21	Received from store 5 Part BC 5130 PRINTER R08397	1		4				
16.6.21	Received from store AMODEL BC 6000 Computer UPS online 3 KVA	1		5				Returned to store model no PE 6000 Store Stock Pg. 366 Transferred to 9/11/21 Peravur



Rekha
JUNIOR SCIENTIFIC OFF
D H KANNUR

Stores Superintendent
DISTRICT HOSPITAL KANNUR
4.60.20/



V. Relcha
JUNIOR SCIENTIFIC OFFICER
D. H. KANNUR