



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 204209

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : GH Palu	Call Registration Date : 19/5/2024
Address : Kottayam	Caller ID : 111964
Ph : 9495018721	Date of Visit : 20/5/2024
	Asset No. : 0514057
	EQPT Name : BP Apparatus
	Manufacture Meditech Industries Model : MCP
	S. No. 603072 Dept. Casualty

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Below zero level mercury and mercury leakage

Action Taken : Checked the BP Apparatus and refill the mercury but mercury leaked the tank. This BP Apparatus repair not possible. This reasons for recommending the BP Apparatus as RBER.

Completed Date : 20/5/2024 Time : 11am..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Josekutty George	20/5/2024	10am	11am

Customer Remark : Completed Pending



Service Engineer Name : Josekutty George
Signature : *[Signature]*
Date : 20/5/2024
Contact Number : 8848230407

Customer Name : KRISHNAPRIYA
Signature : *[Signature]*
Date : 20/5/24
Contact Number : 9495655671
Designation : Nursing officer
Hospital Seal :

REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	GH PALA
3	Name of Equipment with Make, Model and Serial Number	Equipment :BP APPARATUS Make: MEDITECH Model :MCP SN: 603072
4	Equipment ID & Barcode	111964 & 0514057
5	Date of purchase / Year of manufacture /Installation Date	25/3/2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	19/05/2024 (Toll free)
9	Action taken	Checked and found out the mercury tank leaked and spilled out the mercury. Screw of hinges are broken.
10	Present status of the equipment (Fully damaged / parually damaged)	Fully damaged ✓
11	Recommendations for repair (required service details)	Not recommending for repair. ✓
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	Rs. 2000/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached ✓
16	Reasons for recommending the equipment as BER	Checked and found out the mercury tank leaked and spilled out the mercury. Screw of hinges are broken. Machine installed in 25/03/2017 and covered up to 7 + years. Mercury not available in the market . So <u>recommending to condemnation</u> ✓
17	Name & Signature of CYRIX Authority	Josekutty George

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset - Rs 2000
 Cost of spare - Not mentioned
 Date of installation - 25/3/2017
 Life period - 7 years

As per the RBER guidelines recommended as RBER

Signature of JC BM (NHM)

23/05/2024
 Date

Signature of
 Superintendent / Medical Officer (i/c)

RMO
 General Hospital
 Pala



B.P. apparatus

	No	Date of receipt	Cost	Conda	Balance	Dr	Cr	Particulars
Transfer in from Imagical								
2/1/07	1	16.8.97	398		1			309
2/1/05	2	2.3.05	Rs. 440/-		2			309
18/6/07	3	18/6/07	-		4			myth
11/4/08	4	11/4/08			5			myth
16/12/08	5	16/12/08	Rs. 437/-		6			myth
16/12/08	3		Rs. 1785/-		10			myth
Main slab page no. 124 vol. IV								
8/6/13		To out to unseizable article -> Condemned as per order No.						
		Registers page No. C2 (C1) 1096/13/GHP dt. 27.9.3.			4	6		
13/8/13	4	13/8/13	1714.29	1-PP	10			
block page no. 10/ vol II								
5/10/13		To out to unseizable						
		Registers page No. 27			1	9		
- condemned as per order no. C2 6111/13 GHP dt. 20/12/13								
10/2/14		BP apparatus cuff 10 -> Rs 230/each						
		BP apparatus bath valve 10 -> Rs 85/- each						
12/2/14		Amount for case in 10						
		Casualty and OP 10						
Condemned as per order								
		No. C2/3352/14/supt. GHP/15/1/2015			3	6		
20/1/15	1				7			
Rs. 2150/- m.s page 10 vol. 5								
Condemned as per order								
		No. C2/1657/15 GHP dated 30/1/16			1	6		
20/1/15	1				7			
Received from store								

contn. p. 94



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 0514057

SPHYGMOMANOMETER

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 0514057

