


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA


Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	PHC VANIYAMKULAM
3	Name of Equipment with Make, Model and Serial Number	STERILIZER Make : SIDCO Model: N/A SN:N/A
4	Equipment ID/ Barcode	0942302 #100509
5	Date of purchase/ Year of manufacture/Installation Date	03-03-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	11-03-2024 toll free
9	Action taken	Checked the machine and found that Pressure gauge, safety valve, relief valve, silicon gasket, coil with connector, door lock knob. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Pressure gauge = 1800.00 Safety valve = 1300.00 Relief valve = 1100.00 Silicon gasket = 2500.00 Heater coil with connector = 1900.00 Door lock knobs = 700 TOTAL = 9300.00/-

13	Asset Value	RS. 15000/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	62%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report & Quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Pressure gauge, safety valve, relief valve, silicon gasket, coil with connector, door lock knob. Damaged Equipment installed on 03-03-2018 aged up to 6 years. Repairing cost is 62% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 Repairing cost is around 62% and machine aged up to 4 years hence both criteria for RBER met. Hence recommended for BER.


 3/04/24
 Signature of Junior Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad

Date: _____




 Signature of Superintendent (NHM) Palakkad
 സൂപ്പർവൈസിംഗ് ഓഫീസർ (NHM) പാലക്കാട്



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 197251

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC VANNIAMKULAM</u></p> <p>Address <u>PALAKKAD</u></p> <p>Ph : <u>9656464685</u></p>	<p>Call Registration Date : <u>11/03/2024</u></p> <p>Caller ID : <u>100509</u></p> <p>Date of Visit : <u>12/03/2024</u></p> <p>Asset No. : <u>0942302</u></p> <p>EQPT Name : <u>Autoclave</u></p> <p>Manufacture <u>SALCO</u> Model : <u>Nil</u></p> <p>S. No. <u>Nil</u> Dept. <u>OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working. (Required more time for heating, steam leak, pressure gauge not working, broken door knob.

Action Taken : checked the machine and found that pressure gauge, safety valve, relief valve, silicon gasket, coil with connector door lock knob need to replaced.

Required further checking after the spare replacement

Completed Date : 12/03/2024 Time : 12:30PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. Pressure gauge	1 NOS		
2. safety valve, Relief valve	1 NOS		
3. Silicon gasket, door lock knob	1 NOS		
Header coil with connector	1 NOS		

Cyrix Engineer	Date	Start Time	End Time
<u>Rahul H V</u>	<u>12/03/2024</u>	<u>11AM</u>	<u>12:30PM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Rahul H V</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>12/03/2024</u></p> <p>Contact Number : <u>9539838931</u></p>	<p>Customer Name : <u>AP BHANU AKAASH HALLAM</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>12/13/24</u></p> <p>Contact Number : <u>9656464685</u></p> <p>Designation : <u>NURSING OFFICER</u></p> <p>Hospital Seal : <u>GRI</u></p>
---	--

JAIGURU SALES & SERVICES

The Team For Sterilization
XIII/385, N.Paravur, Ernakulam, Kerala-683519

Ph: 8089960227
Mob: 09447043821

To

Cyrix Health care Pvt. Ltd,Ernakulam

Ref: Qtn. /162/2023-24 Dated 28-03-2024

Sir,

Sub: Quotation for Autoclave Repair at Vaniamkulam, Palakkad

As per the instructions of your service Engineer, we are giving the rate of the following spares for the service of the small vertical autoclave.

1. Pressure Gauge @ 1800/-
2. Safety Valve @ 1300/-
3. Relief Valve @ 1100/-
4. Silicon Gasket (door) @ 2500/-
5. Heater Coil With Connector @ 1900/-
6. Door Lock Knobs @ 700/-
7. Service Charges @ Rs. 2000/-

Total Rs. 11,300/-

Terms & Conditions

Tax	: 18% GST Extra
Delivery	: Ready
Guaranty	: NA
Payment	: 100% advance with the order

Expecting your valued co operation,

For Jaiguru Sales & Service

Proprietor
Remesh Kumar



REQUIRED QUOTATION FOR AUTOCLAVE #0942302 # 100509 @ PHC VANIYAMKULAM

2 messages

Akhilesh P <akhileshcyrix@gmail.com>
To: Jaiguru Sales & Service Kerala <Jssindiaster@gmail.com>
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>

Thu, 6 Jun, 2024 at 2:52 pm

Dear sir

Please share the price and stock availability of

1. Pressure Guage
2. Safty valve
3. Relief valve
4. Silicon gasket
5. Heating coil
6. Door lock kobe

@ phc vaniyamkulam. Palakkad

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Jaiguru Sales & Service Kerala <jssindiaster@gmail.com>
To: Akhilesh P <akhileshcyrix@gmail.com>

Sat, 8 Jun, 2024 at 3:51 pm

Sir,
We have attached the rates ann terms and conditions.Please go through
it and do the needful.

[Quoted text hidden]

For Jaiguru Sales & Service
Remesh Kumar
Proprietor
Jaiguru Sales & Service
N.Paravur, Ernakulam, Kerala-India
Ph: 8089960227 Mob: 09447043821





Handwritten text in a South Asian script, possibly Sinhala, located at the bottom right of the page. The text is partially obscured by a signature and a stamp.

AUTOCLAVE 12012

Cooker

Buckel Type Stainless Steel Cooker

3/3/18	Purchased from Sidco Kerala (Sd/- of government of Kerala) Marketing Division, P.B. No. 4, Kozhikode Coe 2020 supplied by Sidco Marketing - Centre, Palakkad Invoice - SMC PKD/17-18/II/210 Date: 2/3/18 Model of Size 1242 Rate 12430 tax: 2237.00 Total 14667/-			
12/18	Issued to S/N Ambale	1	1	1



12012
 14667/-
 12/18