



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 192930

A 83

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Thiruvananthapuram</u></p> <p>Ph : <u>7012896198</u></p>	<p>Call Registration Date : <u>01-04-2024</u></p> <p>Caller ID : <u>104090</u></p> <p>Date of Visit : <u>02-04-2024</u></p> <p>Asset No. : <u>0110982</u></p> <p>EQPT Name : <u>B.P Apparatus</u></p> <p>Manufacture <u>HLL Lifecare</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Female Surgical Ward</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : BP Cuff not Inflating

Action Taken : Checked the machine and found air pump motor damage. Also found control board damage. Need to replace these spares to further check working condition of machine.

Completed Date : 02/04/24 Time : 5:00 PM Spare Required


Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.	NA	1	1	1
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>ADITHYA. J.S</u>	<u>02/04/24</u>	<u>4:00pm</u>	<u>5:00pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>ADITHYA. J.S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>02-04-24</u></p> <p>Contact Number : <u>9785054344</u></p>	<p>Customer Name : <u>Superintending General Hospital Thiruvananthapuram</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>2/4/24</u></p> <p>Contact Number :</p> <p>Designation :</p> <p>Hospital Seal :</p>
--	--



PARVATHY
2/4/24


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

283

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GH THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : BP APPARATUS Make: LIFECARE Model :BSX 515 SN : NA
4	Equipment ID & Barcode	104090 & 0110982
5	Date of purchase / Year of manufacture /Installation Date	31.03.2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	01.04.2024 (Toll free)
9	Action taken	Checked the machine and found the motor and control board defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 1201.76
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached.or.Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found that motor and control board are defective. Machine installed on 31.03.2021 and aged up to 3 years. Quotation not attached as spare not available in the market. Hence recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	ADITHYA J S 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph



Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

air pump motor and board damaged. So can be
 RBER:

MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram


 22.4.2024

Signature of JC BM (NHM)



	 Superintendent General Hospital Thiruvananthapuram
Date	Signature of Superintendent / Medical Officer (i/c)
Seal	



74 Digital BP Apparatus.

Date	Particulars.	R/d	Issuit	Balance	Remarks
22/11/19	R/d from store (True check)	1		1	
22/2/21	R/d from store [Elko] (Exp Date 28/12/2030)	1	✓	2	By H/W
31/3/21	(L.P) Received from store Tris from local purchase Register page no. 26 Exp date (RS 1201-46) 30/6/2030 model BSx 575 Lot no S15C1B119 Batch no BH 19 - model BS Box code 0115893 i) 0115816 ii) 0110082 iii) 0110592 (extra)	1		3	By JND
20/10/2021	Physically found BPL 120/80 B9			4	By JND

Superintendent
General Hospital
Thiruvananthapuram
Kerala



Superintendent
General hospital
Thiruvananthapuram