



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 193867

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility ... CMC Ambalapuzha	Call Registration Date : 29/4/2024
Address ... Dist. Alappuzha	Caller ID : 108537
Ph : 9544776756	Date of Visit : 30/4/2024
	Asset No. : 0431306
	EQPT Name : BP Apparatus
	Manufacture Anita Industries Model : Elkonnector
	S. No. 295315 Dept. OP

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Mercury spilled out, BP Valve, BP Bulb, Bladder & Mercury Tank are defective

Action Taken : Checked and found that Mercury spilled out, BP Valve, BP Bulb, Bladder & Mercury Tank are defective. Need to Replace these spares for further checking and working condition of the equipment.

Completed Date : 6/5/2024 Time : 1:30 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Savath Chandran	6/5/2024	12:00 PM	1:30 PM

Customer Remark: Loss of mercury. Completed Pending



Service Engineer Name : Savath Chandran
Signature : Savath
Date : 6/5/2024
Contact Number : 8075052698

Customer Name : ATHULY Administrative Medical Office
Signature : ATHULY Urban Health Training Centre
Date : 6/5/24
Contact Number : 9544776756
Designation : Nursing Officer (OT)
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	CHC AMBALAPPUZHA
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS MAKE :- ANITHA INDUSTRIES MODEL :- ELKOMETER SN :- 395315
4	Equipment ID & Barcode	108537 & 0431306
5	Date of purchase / Year of manufacture /Installation Date	17/04/2015
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29/04/2024
9	Action taken	Checked the machine and found that mercury tank is leaking and insufficient mercury, BP Valve, Bulb and Bladder are defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS:- 1323/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that BP Valve, BP Bulb, Bladder & mercury tank is defective. Mercury is leaking and insufficient mercury in the tank. The equipment was installed on 17/04/2015 and aged up to 9 years. Quotation is not submitted since Mercury is not available in the market. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	SHARATH CHANDRAN

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

[Signature] 07/05/24
JYOTHISH THOMAS
 Consultant (Biomedical Engineering)
 NHM ALAPPUZHA
 Signature of JC BM (NHM)

Date

Seal



[Signature]


Signature of
 Superintendent / Medical Officer (i/c)
 Administrative Medical Officer
 Urban Health Training Centre
 Ambalapuzha

STOCK REGISTER

Name 46 BP Apparatus

Date	Received from or issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	P.				
2-11-09	Found on physical verification							
	Δ_2 346/13 CHTC							
	12/6/10							
12/6/10	Received from store				1			
	Used to actually					1		
9/3/10	Received from store (mercury free)				1			
11/9/10	Received from store				1			
	Issued to OPD					1		
19/3/10	Received from store				6			
	Used to OP					6		
15-22	Received from store							
	(Digital BP Apparatus)				3			
	Issued to OPD + NCD					3		




 Administrative Medical Officer
 Urban Health Training Centre
 Ambalapuzha



Handwritten signature
Administrative Medical Officer
Urban Health Training Centre
Anantnag

