



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 204047

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>Thy</u>	Call Registration Date : <u>27/4/2024</u>
Address : <u>Vaikom</u>	Caller ID : <u>107919</u>
<u>Kattappan</u>	Date of Visit : <u>27/4/2024</u>
Ph : <u>949644410</u>	Asset No. : <u>0520013</u>
	EQPT Name : <u>digital BP</u>
	Manufacture <u>Accusure</u> Model : <u>TMB-1490-A</u>
	S. No. <u>BA154116061007A</u> Dept. <u>Casualty</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

Problem Identified : equipment is not working

Action Taken : checked and found out the problem with main PCB and nbp pump. Need to replace these spares for further working condition.


Completed  Date : ..... Time : ..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AMAL BABU</u>	<u>27/4/2024</u>	<u>2:40pm</u>	<u>3:00pm</u>

Customer Remark  Completed  Pending  
Equipment not working Ajji



Service Engineer Name : AMAL BABU  
Signature : [Signature]  
Date : 27/4/2024  
Contact Number : 9567575983

Customer Name :  
Signature : [Signature]  
Date : 27/4/2024  
Contact Number :  
Designation : Superintendent  
Hospital Seal : Taluk Head Quarters Hospital  
Vaikom, Pin - 686 141




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KOTTAYAM
2	Name of Hospital	THQH VAIKOM
3	Name of Equipment with Make, Model and Serial Number	Equipment :DIGITAL BP Make : Accusure Model :TMB-1490 SN: BA154116061000790D
4	Equipment ID & Barcode	107919 & 0520013
5	Date of purchase / Year of manufacture /Installation Date	1/12/2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24/4/2024 (Toll free)
9	Action taken	Checked and found out the problem with main PCB And Nibp pump
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	Rs .2060 /-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main PCB and nibp pump. Machine installed in 01/12/2017 and covered up to 6 + years. Quotation not submitted since spare are not available in the market So recommending the machine for condemnation as per the tender close 5.3.14.2
17	Name & Signature of CYRIX Authority	Amal babu 

\*Not mandatory #Based on the period of life and value as per the BER guidelines


\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 2060  
 Cost of spare - Spare cost not mentioned  
 Installation Date - 1/12/2017  
 Life period - 6 years  
 As per the guidelines recommended as BER

  
 Signature of DC BM (NHM)

Date

  
 Seal

  
 Superintendent  
 Taluk Head Quarters Hospital  
 Vaikom, Pin - 686 141  
 Superintendent / Medical Officer (I/c)




Digital BHP Appointments

Date	Description	Qty	Value	Remarks
1/10/18	Rd from Hce (3 items) 2 Polyden (1) 15 350 P 5	3		
28/11/18	Received from Govt (Flood) Condemned us per order no C2 3785 MCHM dtd 13/6/18	1		
10/7/20	Good from Govt (Digital Resonance) U 2 Battery (4) Tr-out to Unserviceable Register P-21	3		
	Condemned us per order no C2 1288/2019 T-H-GH VKM Dtd 24-9-2019 (1/12/17-3)	3		
10/3/2021	Tr-out to unserviceable books (24/4/18) Condemned us per order no C2 5735/22 Supt THGH VKM dtd 24/11/18	1		
1/1/2023	Rd from Hce (Coronation)	1		
1/1/23	Rd from Hce (Coronation)	1		
24/2/23	Rd from Hce (Coronation)	1		
17/6/23	Tr-out to unserviceable register papers (32)	3		
23/10/23	RFS	2		
	Barcode - 0452076 Dr. Trust (1) 1860 + GST	5	9	Supply SMO
	Barcode - 0452077 Dr. Odio (1) 1860 + GST			
20/3/2024	Rd from Govt (MUNICIPAL FUND) SUPPLIER NAME MICROTIG INTERNATIONAL bul 0009	5		
	MRP 1/1/2023 Rd from 31/12/2025 Price 1100 each (5499.98)	1		
27/3/2024	Tr-out to desk of	1		



Superintendent  
Taluk Head Quarters Hospital  
Valkom, Pin - 686 141





A handwritten signature in green ink, appearing to be a stylized name.

Superintendent  
Taluk Head Quarters Hospital  
Vaikom. Pin - 686 141

