



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



Service Provider
Order No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 204046

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility : Thy
Address : Vaikom
Kottayam
Ph : 9496114410

Call Registration Date : 26/4/2024
Caller ID : 108285
Date of Visit : 29/4/2024
Asset No. : 0520027
EQPT Name : Digital BP
Manufacture : Accuware Model : TMB-1490
S. No B/A/154/1505/000780 Dept. Physiatry

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : equipment is not working

Action Taken : checked and found out the problem with main PCB and mbp pump. Need to replace these spares for further working condition


Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AMAL BABU</u>	<u>29/4/2024</u>	<u>3:00pm</u>	<u>3:30pm</u>

Customer Remark : Completed Pending
equipment not working



Service Engineer Name : AMAL BABU
Signature : [Signature]
Date : 29/4/2024
Contact Number : 9567575983


Customer Name : [Signature]
Signature : [Signature]
Date : 29/4/2024
Contact Number : [Signature]
Designation : Superintendent
Hospital : Taluk Head Quarters Hospital
Hospital Seal : Vaikom Dist. 686 165

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	THQH VAIKOM
3	Name of Equipment with Make, Model and Serial Number	Equipment :DIGITAL BP Make : Accusure Model :TMB-1490 SN: BA154116061000780D
4	Equipment ID & Barcode	108285 & 0520027
5	Date of purchase / Year of manufacture /Installation Date	1/12/2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	26/4/2024 (Toll-free)
9	Action taken	Checked and found out the problem with main PCB And Nibp pump
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	Rs .2060 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main PCB and nibp pump. Machine installed in 1/12/2017 and covered up to 6 + years. Quotation not submitted since spare are not available in the market So recommending the machine for condemnation as per the tender close 5.3.14.2
17	Name & Signature of CYRIX Authority	Amal babu 



*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 2060
 Cost of spare - Nil
 Life period - 6 year
 As per the RBER guidelines recommended as
 RBER


 Signature of JC BM (NHM)

Date	 Seal	 Superintendent Taluk Head Quarters Hospital Superintendent / Medical Officer (Uc)
------	---	--

Digital B/P Apparatus

1/10/12	Rd from the store 2 Bolepdes (1) ✓	3			3
2P	Rs 6182.40 (3- month) 15 380 P 5				
24/11/18	Received from site (Flood)	1			4
	Condemned us per order no C2 3785 12/11/18 THOH VKM dtd 13/6/23				
10/7/20	Rd from site (Digital BP monitor)	3			7
	U 2 Bolep (14)				
	Tr-out to Unserviceable Register P-21				
	Condemned as per order no C2 1288/2019				
	T-H-THH VKM Dtd 24-9-2019 (1/12/17-3)		3		4
10/3/2021	Tr-out to unserviceable book				
	(24/11/18) Condemned us per order no				
	C2 3785/28 Supt THOH VKM dtd 24/11/18		1		3
11/1/2023	Rd from Hc (Coronation)	1			4
11/1/23					
24/2/23	Rd from Hc (Coronation)	1			5
					2
17/6/23	Tr-out to unserviceable register page				
	(32)		3		2
23/10/23	RFS				
	Barcode - 0452076 Dr. Forest (1) 1860 + GST		2		4
	Barcode - 0452077 Dr. Odib (1) 1860 + GST				
20/3/2024	Rd from store (MUNICIPAL FUND) SUPPLIER NAME MICROTOK INTERNATIONAL bul 0009	5			9
	TRP 1/1/2023 Rd. Dal. Bill no 25 PRYU 1100 cat (5497-78)				
27/3/2024	Tr-out to bank OP				



Superintendent
 Taluk Head Quarters Hospital
 Talukom. Pin - 686 141

12/17



Superintendent
Taluk Head Quarters Hospital
Vaikom, Pin - 686 141

