

**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 204024

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERD Approved Service Agency

Service Report

30/64 1 B, Potta Junction, Poonithurn, Kochi - 002 030, Kerala
Ph : 98472 99500 Website : www.cyrlix.com | E-mail : bomp.kl@cyrlix.in

Health Facility	Thy	Call Registration Date :	7/4/2024
Address	Vaikom Kodiyam	Caller ID :	104977
Ph :	8848909153	Date of Visit :	8/4/2024
		Assot No. :	0521267
		EQPT Name :	UPS
		Manufacture	APC Model : BX600C-1M
		S. No.	NA Dept. lab

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NO output voltage

Action Taken : checked and found out the problem with main PCB and battery. Enquired the spare with OEM.

Completed Date : Time : Spare Required

Description	Qty.	Part Number	PR Number
1. na	na	na	na
2.			
3.			

Cyrlix Engineer	Date	Start Time	End Time
AMAL BABU	8/4/2024	11:30am	12:00pm

Customer Remark: Equipment is not working. **Completed** Pending


Service Engineer Name : AMAL BABU	Customer Name :
Signature :	Signature :
Date : 8/4/2024	Date :
Contact Number : 9567575983	Contact Number :
	Designation : Superintendent
	Designation : Taluk Head Quarters Hospital
	Hospital Name : Vaikom. Pin - 686 131

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	THQH VAIKOM
3	Name of Equipment with Make, Model and Serial Number	Equipment :UPS Make : APC Model :BX600C-IN SN: NIT SN B21750019469 <i>[Signature]</i> COM BMEC
4	Equipment ID & Barcode	104977 & 0521267
5	Date of purchase / Year of manufacture /Installation Date	22/08/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	05-04-2024 (Toll free)
9	Action taken	Checked and found out the problem with main PCB and battery. Enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	1kv line interactive ups pcb -1062/- 7ah battery -1920/- Total =2982/-

13	Asset Value	Rs . 4500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.26%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and quotation are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main PCB and battery. Machine installed in 22/08/2018 and covered up to 5+ years. The repairing cost 66.26%, both criteria met. so recommended the machine for condemnation as per the tender close 5.3.14.1
17	Name & Signature of CYRIX Authority	Amal babu 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value Rs. 4500

Cost of spare Rs 2982 (66.26%)

Life period - 5 year 9 months

The machine supplied with fully Automate Bio-chemistry Analyser - As per the RBER guidelines recommended as RBER

CONSULTANT BIOMEDICAL
Signature of JC BM (NHM)



Date

Seal


Superintendent
Taluk Head Quarters Hospital
Vaikom, Pin - 686 141
Superintendent / Medical Officer (i/c)

GENERAL HOSPITAL


135

STOCK BOOK OF.....

Name of Article FULLY AUTOMATED BIOCHEMISTRY ANALYSER & RELATED EQUIPMENT

Date	No & date of Invoice/ Issue Note	From whom received or to whom issued	Receipt	Issue	Balance	Initials of receiver	Remarks
22/8/2018	Invoice No: CO1819-005162	AGAPPE Diagnostics Hd. Agappe Hills Pallimattom P.O, Eran (Dist) 0424 - 285700	26/7/18				
	Purchase Order No: EQUIPMENT-19-596	TVM, Theyaud P.O, for Rs. - 11,59,345.00					
	① Fully Automated Biochemistry Analyser SNO: NT86001263			1 No			Value 982496.00
	② Desktop Monitor, Mispa Clinic SNO: ETLCH010435LO			1 No			Zero.
	③ Desktop Computer withted Monitor (Mindray higher end equipment) SLNO: PGO192R4			1 No			Zero.
	④ Computer Table			1 No			Zero.
	⑤ UPS Mispa Clinic SNO: B21750019469			1 No.			Zero.
	⑥ PRINTER HP 1108 SLNO: VNEAG18304.			1 No			Zero.
	⑦ Compactible Online UPS 2KVA SNO: 88117300589			1 No.			Zero.
	⑧ Battery for 2/3 KVA UPS SNO: MOQ 04 21353 00998			1 No.			Zero.
	⑨ Battery for 2/3 KVA UPS SNO: MOQ 0421353 01017			1 No			Zero.
	⑩ Battery for 2/3 KVA UPS SNO: MOQ 0421353 01016			1 No			Zero.
	⑪ Battery for 2/3 KVA UPS (SNO: MOQ 04 21353 01018)			1 No:			Zero.




 Superintendent
 Taluk Head Quarters Hospital
 Vaikom, Pin - 686 141



Arun GC <arungcyrix@gmail.com>

1KVA UPS SERVICE PROPOSAL

Bhagath Mohan <aionbackups12@gmail.com>
To: arungcyrix@gmail.com

Mon, Apr 22, 2024 at 3:57 PM

Thanks & Regards

Bhagath Mohan
PH:-8921950217

 **THQH VAIKOM SERVICE PROPOSAL.pdf**
656K

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 06/04/2024

Contact: 8921950217

Email: aionbackups12@gmail.com

GSTIN: 32CGUPM9450L1ZN

1 KVA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTHCARE PVT .LTD, THQH VAIKOM

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	1KVA LINE INTERACTIVE UPS PCB	1	900.00	900.00	18	1062.00
2	7Ah AMARON BATTERY	2	750.00	1500.00	28	1920.00
PAYABLE AMOUNT						2982.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,

Sincerely,

BHAGATH MOHAN



⚠️ ⚠️ WARNING

RISK OF ELECTRIC SHOCK

- Do not remove the cover. Hazardous live parts inside this UPS are energized from battery even when AC supply is disconnected.
- Refer servicing to qualified personnel only.
- Refer user manual and safety guide for other details.

Failure to follow these instructions can result in serious injury or equipment damage.

Output
230 V ~, 2.6 A, 50 Hz
600 VA, 360 W

Self Declaration -
Conforming to
IS 16242 (Part 1) : 2014,
R-83000450

3/3333C-11
222758815469



BATTERY BACKUP
+
SURGE PROTECTION



[Signature]
Superintendent
Taluk Head Quarters Hospital
Vaikom. Pin - 686 141

Back-UPS 600

APC
by Schneider Electric

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (RESUP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 0521267

⏻ ⏪ ⏩ ⏹