



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Order No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 203787

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>CHC</u>	Call Registration Date : <u>20/04/2024</u>
Address : <u>CHC THOTTACKAD</u>	Caller ID : <u>107269</u>
<u>KOTTAYAM</u>	Date of Visit : <u>24/04/2024</u>
Ph : <u>9447855388</u>	Asset No. : <u>0530201</u>
	EQPT Name : <u>SUCTION APPARATUS</u>
	Manufacture : <u>NA</u> Model : <u>NA</u>
	S. No. : <u>NA</u> Dept. : <u>OPD</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Suction Apparatus not working



Action Taken : Checked and found problem with motor and suction jar

Completed  Date : ..... Time : ..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>SANISH SIVAN</u>	<u>24/4/24</u>	<u>2:00 PM</u>	<u>3:30 PM</u>

Customer Remark	Completed <input type="checkbox"/>	Pending <input checked="" type="checkbox"/>
<u>Suction apparatus is not working. Pressure gauge not working.</u>		
Service Engineer Name : <u>SANISH SIVAN</u>	Customer Name	Signature
Signature : <u>[Signature]</u>	Date	Designation
Date : <u>24/4/2024</u>	Designation	Hospital Seal :
Contact Number : <u>9946646571</u>	 	

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	CHC THOTTACKAD
3	Name of Equipment with Make, Model and Serial Number	Equipment : Suction Apparatus Make : NA Model :NA SN: NA
4	Equipment ID & Barcode	107269 & 0530201
5	Date of purchase / Year of manufacture /Installation Date	01-09-1998
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	20-04-2024 (Toll free)
9	Action taken	Checked and found out the problem with suction motor and jar enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Suction motor -4650/- Suction jar – 2500/- Total - 7150/-

13	Asset Value	Rs . 9500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	75 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and quotation are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with suction motor and jar. Machine installed in 01/09/1998 and covered up to 26 + years. The repairing cost 75 % Both criteria met. So recommended the machine for condemnation as per the tender close 5.3.14.1
17	Name & Signature of CYRIX Authority	Sanjish Sivan

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset Value - 9500  
 Cost of spare - 7150  
 Installation Date - 1/9/98  
 Life period - 25 years  
 As per the BER guidelines, recommend as BER

CONSULTANT BIOMEDICAL ENGINEER  
 NATIONAL HEALTH SYSTEMS ENGINEER  
 Signature of JC BM (NHM)



Date

Seal

Signature of **MEDICAL OFFICER**  
 CHC THOTTAKKADI

MUCUS EVACUATOR  
(S.S. P. No: 157)

KNIFE PLASTER  
(S.S. P. No: 62)

B/F from Page 36  
Issued to H/N

(Condemned) No: 250  
265/81-8/1/98

Row 1  
Level 2  
Q.No. 2

B/F from Page 61  
Issued to H/N-2nd

(Condemned as per order no  
B2 1351/03 DMHCK)  
Dt 2.1.04

SUCTION APPARATUS

Help IV: 1637...  
B/F from Page 15...  
Cat No. 5500...  
model - 55...  
Exemption order

Info. Info. 6

21.9.98 To the H/N Smt. PM  
Kathamma SS Page No. 103

Info Nil

5-6-07. IN. NO 86/07-08 dt 5-6-07. DMS. 10TYM

Rs. 5370/1 VIBASH (MEDIVAC) INO INO R

5-6-07 No issued to H/N A/Gie Joseph INO NIL R  
(S.S. Page No 103)



MEDICAL OFFICER  
THOTTAKKADU  
[Signature]



Arun GC <arungcyrix@gmail.com>

---

## CHC THOTTAKKAD SUCTION RBER

1 message

---


**CLOUD 99** <cloud99biotech@gmail.com>  
To: arungcyrix@gmail.com

Sat, May 18, 2024 at 1:02 PM

KINDLY FIND THE ATTACHMENT

SUSAN JOHN  
Cloud99 Biotech  
Medical Engineering Works  
Room No: 22, CSI Golden Jubilee Complex  
Kottayam 686 001  
Kerala, India  
Ph: 9744473043  
email: [cloud99biotech@gmail.com](mailto:cloud99biotech@gmail.com)

---

 **Estimate\_C99B-QTN-8\_03\_05\_24- THOTTAKKAD.pdf**  
197K



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333  
Room No. 22,  
CSI Golden Jubilee Complex  
Baker Hill  
Shastri Road  
Kottayam

# CLOUD99 BIOTECH

GSTIN: 32AMWPTS633Q1ZD  
State: 32 - Kerala

MSME NO : UDYAM-KL-07--0030714

## Estimate

Estimate No.: C99B-QTN-8  
Date: 03/05/2024  
Place of Supply: 32-Kerala

Estimate For:

### ARUN G C

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD  
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 8921992272

GSTIN Number: 32AA FCC2499H2ZM

State: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9015	1	-	₹ 3,940.66	₹ 709.32 (18%)	₹ 4,650.00
2	SUCTION JAR 1000 ML (SUCTION JAR 1000 ML (BABY JAR))	9015	2	Nos	₹ 1,116.07	₹ 267.86 (12%)	₹ 2,500.00
<b>Total</b>						<b>₹ 977.18</b>	<b>₹ 7,150.00</b>



Pay To:  
Bank Name : UNION BANK OF INDIA, KOTTAYAM  
Bank Account No.: 361201010035646  
Bank IFSC code : UBIN0536121  
Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 6,172.82
SGST@6%	₹ 133.93
CGST@6%	₹ 133.93
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
<b>Total</b>	<b>₹ 7,150.00</b>

### Description

THOTTAKAD HOSPITAL  
TICKET ID- 107269  
BAR CODE- 0530201

### Estimate Amount In Words

Seven Thousand One Hundred Fifty Rupees only


### Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



*J.H.*  
15  
MEDICAL OFFICER  
CHC, THOTTAKKADU  
-100 mo/c