



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 191349

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Potta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrilx.in

Health Facility <u>PHC Vazhappally</u>	Call Registration Date : <u>11/4/2024</u>
Address <u>Vazhappally</u>	Caller ID : <u>104046</u>
<u>Kottayam</u>	Date of Visit : <u>4/4/2024</u>
Ph : <u>9495587219</u>	Asset No. : <u>0541338</u>
	EQPT Name : <u>UPS</u>
	Manufacture <u>APC</u> Model : <u>Backup UPS</u>
	S. No. <u>B21651022224</u> Dept. <u>Laboratory</u>

600

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Not turning on. Battery was damaged.

Action Taken : Replaced the battery & checked. Found the main board of the equipment is damaged. Need to replace the main board in order to check the working condition of the equipment. Inquired spare with OEM.

Completed Date : 4/4/2024 Time : 1:49pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>12V battery</u>			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Sujin Suresh</u>	<u>4/4/2024</u>	<u>1:10pm</u>	<u>1:49pm</u>

Customer Remark


Service Engineer Name : <u>Sujin Suresh</u>	Customer Name : <u>Fathima P Haneef</u>
Signature :	Signature :
Date : <u>4/4/2024</u>	Date : <u>30/4/24</u>
Contact Number : <u>7902804749</u>	Contact Number : <u>9495587219</u>
	Designation :
	Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC VAZHAPPALLY
3	Name of Equipment with Make, Model and Serial Number	Equipment :UPS Make : APC Model :BX600C-IN SN : B21651022224
4	Equipment ID & Barcode	104046 & 0541338
5	Date of purchase / Year of manufacture /Installation Date	17/05/17
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	1-04-2024 (Toll free)
9	Action taken	Checked and found out the problem with main PCB and battery. Enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	1 KV line interactive ups pcb- 1062/- 7Ah battery -1920/- Total =2982/-


13	Asset Value	Rs . 4500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	66.26%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and quotation are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main PCB and battery. Machine installed in 17/05/2017 and covered up to 6+ years. The repairing cost 66.26%, both criteria met. so recommended the machine for condemnation as per the tender close 5.3.14.1
17	Name & Signature of CYRIX Authority	Sreeju suresh 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 4500
 Cost of spme - 66.26%
 Installation Date - 17/5/2017
 Life period - around 7 years
 As per the RBER guideline recommended as
 RBER


 18/5/2024
 Signature of JC BM (NHM)
 NATIONAL HEALTH MISSION
 KUTTIYAM



Date 30/4/2024

Seal


 Dr. Sali Sebastian
 Medical Officer in charge
 Signature of
 Superintendent / Medical Officer (I/c)
 Cheeranchira P.O.

UPS. Mispur

Viva

R13.17 Invoice No. CO-DMTC-1617-0195

Pathimattom P.O, Kinnalankulam Batcher

Agrippa Diagnostic Ltd, Agrippa Health
/ Sennathurai 22224

17.5.17) Kessalikalathur
a. Balu Shode pm: 12)

1 No
Bajna

1 No
Selvan

1 No
Ses



~~Dr. Balu Sebastian~~
Medical Officer in Charge
PHC Vazhappally
Cheranchira P.O



Arun GC <arungcyrix@gmail.com>

1KVA LINE INTERACTIVE UPS SERVICE PROPOSAL

Bhagath Mohan <aionbackups12@gmail.com>
To: arungcyrix@gmail.com

Mon, Apr 22, 2024 at 4:02 PM

Thanks & Regards

Bhagath Mohan
PH:-8921950217

 **1KVA UPS SERVICE PROPOSAL - PHC VAZHAPPALLY.pdf**
604K

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 08/04/2024

Contact: 8921950217

Email: aionbackups12@gmail.com

GSTIN: 32CGUPM9450L1ZN

1 KVA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTHCARE PVT .LTD,PHC VAZHAPPALLY

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	1KVA LINE INTERACTIVE UPS PCB	1	900.00	900.00	18	1062.00
2	7Ah AMARON BATTERY	2	750.00	1500.00	28	1920.00
PAYABLE AMOUNT						2982.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% After installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC


Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you,
Sincerely,

BHAGATH MOHAN



Medical Officer in charge
PHC Vazhappally
Cheruvanchira P.O

