



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 218429 495

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Women & Children Hospital</u></p> <p>Address <u>Thycaud,</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>8136810389</u></p>	<p>Call Registration Date : <u>29-5-2024</u></p> <p>Caller ID : <u>113993</u></p> <p>Date of Visit : <u>30-5-2024</u></p> <p>Asset No. : <u>0114531</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>Elcometex</u> Model : <u>Microsil</u></p> <p>S. No. <u>NA</u> Dept. <u>Casualty</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : BP not getting

Action Taken : Checked and found mercury spillage. Identified mercury tank, control valve, glass rod holder, and glass rod defective. Need to replace these spares for further checking.


Completed Date : 30-5-2024 Time : 11:20 a.m Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	NA	NA	NA	NA
2.	NA	NA	NA	NA
3.	NA	NA	NA	NA

Cyrix Engineer	Date	Start Time	End Time
<u>Rehna R</u>	<u>30-5-2024</u>	<u>10:50 a.m</u>	<u>11:20 a.m</u>

Customer Remark Pending



Service Engineer Name : Rehna R
Signature : Rehna
Date : 30-5-2024
Contact Number : 9947239956

Customer Name : VIJAYAKUMARI B.
Signature : [Signature]
Date : 30/5/24
Contact Number : 9447257045
Designation : SUPERINTENDENT
Hospital Seal : Govt. W & C Hospital, Thycaud

Thiruvananthapuram - 14


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

495

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	WOMEN & CHILD HOSPITAL THYCAUD
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP Apparatus Make : Anita Industries Model : Elkometer Serial No : 430557
4	Equipment ID/ Barcode	0114531/113993
5	Date of purchase/ Year of manufacture/Installation Date	20/08/2018
6	Warranty details (Ycs/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	29/05/2024
9	Action taken	Checked and found mercury spillage. Identified mercury tank and glass rod defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	1323/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not Available
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found Mercury spillage, Identified mercury tank and glass rod defective. The unit is installed on 20/8/2018 and covered upto 5+ years. Quotation not attached since the mercury is not available in the market. As per the tender clause 5.3.14.2 recommending the unit for condemnation
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Mercury rod defective. Mercury is spilled off. So as per tender clause, this can be RBER


MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


12-6-2024



Signature of JC BM (NHM)

Date: 10/6/2024




Signature of
Superintendent/Medical Officer (i/c)

SUPERINTENDENT
Govt. W & C Hospital, Thycaud
Thiruvananthapuram - 14



Saanthi
 10/6/24
 SUPERINTENDENT
 Govt. W & C Hospital, Thycad
 Thiruvananthapuram - 14


STOCK REGISTER OF B.P APPARATUS 194

Name of Article.....

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward					
20/8/2018		Received From store		2	24		
		[One for casualty One for initial assessment]					
21/2/19		Received from store		2 (Issued to OP)	26		
18/9/19		Received from store		1	27		
		[Life-medivision Mercury free]					
18/10/2021		Condemned as per order - NO 7094/2019 Dated 26/8/21 (17 Nos)					
9/3/22		To-out to unscrutable } Condemned as per order (5) Register Page No: 1 } 5 Nos. 4498/2022 (Kang) 3. 1807 dt. 23.12.22					
12/10/22		[Bill no 87132 dt. 12/10/22. Kalanya Pharmacy 29 CID No AS 8710] Received from Kal - BP Apparatus - (1) Rate 1073/- Cost = 64.38/- GST - 64.38/- Net Amt - 128.76/-					

[One thousand two hundred and two] Issued to [unclear] [unclear]

HMC Register Vol. 16 (A) 10,000x100/23/539/2007/150/14 Govt. Kerala.



Santhosh
SUPERINTENDENT
Govt. W & C Hospital, Thiruvananthapuram - 14