


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC VELUR
3	Name of Equipment with Make, Model and Serial Number	Nebulizer Make: Life Care Model: CNB69022 SN: LCZZ190401622
4	Equipment ID/ Barcode	#107689 - 0840876
5	Date of purchase/ Year of manufacture/Installation Date	21-12-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 23-04-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Need to replace these spares for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	1353 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Equipment installed on 21-12-18 aged up to 5 years 3 month. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine totally damaged. spare not available


 Signature of JC BM (NHM)

Date: _____




 Signature of
 Superintendent/Medical Officer (i/c)



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 185103

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility P.H.C. Velur	Call Registration Date : 23-04-2024
Address Thiruvissar	Caller ID : 107689
Ph : 9746252660	Date of Visit : 24-04-2024
	Asset No. : 0840876
	EQPT Name : nebulizer
	Manufacture Lite case Model : CNBB9022
	S. No. LC22190401622 Dept. casualty

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : machine not switching on.

Action Taken : checked the machine and found that machine is not switching on. found motor and piston are defective. need to replace these spares. further checking.

Completed Date : 24-04-2024 Time : 10:00am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Bijoy	24-04-2024	9:30am	10:00am

Customer Remark Completed Pending

Service Engineer Name : Bijoy
Signature : *[Signature]*
Date : 24-04-2024
Contact Number : 8921870010



Customer Name : *[Signature]*
Signature : *[Signature]*
Date : *[Signature]*
Contact Number : *[Signature]*
Designation : *[Signature]*
Hospital Seal : *[Signature]*

CYRIX
HEALTHCAREPVT LTD

CN: FHC Vol 4 S.N:
PRO : Nebulizer
ID : 0140710
CAL.ON : 30-4-24
CAL.DTE : 24-4-25



ISO 9001 : 2008 Certified Co. Ph: 574555073

 Life care™

Not Working

✓

CYRIX

HEALTHCARE PVT LTD

CN: FHC006 S.N:
PRO : nebulizer
ID : 0840870
CAL.ON : 5/7/23
CAL.DUE : 11/7/24



[Handwritten signature]

ISO 9001 : 2008 Certified Co. Ph: 9744555073

Compressor Nebulizer

Model No.: CNB69022



Input : AC-220V/50HZ

Power : 180VA

Serial No.: LC22190401770

Mode of operation-continuous

No AP/PG IPXO



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)



TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0840870



**CYRIX HEALTH CARE PVT.LTD
BEMP SERVICE PROVIDER
PREVENTIVE MAINTENANCE SERVICE**

PMS DONE ON: 12/8/23

NEXT PMS DUE: 11/12/23

PERFORMED BY: *[Handwritten signature]*

NEBULISER

Date	Item	Pd from 3/22	Issue	Balance	Signs
20/05/18	Nebuliser i adult and paediatric mask kit life line compress nebuliser	1	OP	ND	d ₂ (damaged and moulded (unusable)) links record
21/12/18	Nebuliser	1			
6/8/19	Nebuliser	1			
2/4/24	Nebulizer (omron). ne-C28 omron health care	1	received from pharmacy store		(3349.92)

