


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC PERINJANAM
3	Name of Equipment with Make, Model and Serial Number	Fetal doppler Make : NIL Model : NIL Sr.No- NIL
4	Equipment ID & Barcode	#101211 – 0831829
5	Date of purchase / Year of manufacture /Installation Date	24-03-2010
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	14-03-2024
9	Action taken	Checked the machine and found main board and PROBE are complaint. Need to replace the main board and probe for further checking .
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Fetal doppler probe -2500/- Main board -3500/- Total -6000/-

13	Asset Value	4000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	150%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found main board and probe complaint. Equipment installed on 15-11-2011 and aged up to 13 YEARS 11 month. Quotation attached and service cost is 150% of asset value both criteria met for RBER. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Foetal doppler is very old.




Signature of JC BM (NHM)


SUPERINTENDENT
Health Centre
Perinjalam - 680 686


Signature of seal

Date



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 190265

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC Perinjalam</u></p> <p>Address <u>Perinjalam,</u> <u>Thriassur</u></p> <p>Ph : <u>04872844484</u></p>	<p>Call Registration Date : <u>14/03/2024</u></p> <p>Caller ID : <u>101211</u></p> <p>Date of Visit : <u>15/03/2024</u></p> <p>Asset No. : <u>0831829</u></p> <p>EQPT Name : <u>Fetal doppler</u></p> <p>Manufacture <u>Na</u> Model : <u>Na</u></p> <p>S. No. <u>Na</u> Dept. <u>Labour room</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the machine and found main board and probe are Complaint. Need to replace the main board and probe are for further checking.

Completed Date : 15-03-24 Time : 1:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin. A-R</u>	<u>15/3/2024</u>	<u>11:00am</u>	<u>1:00 Pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Aswin. A-R</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>15/3/2024</u></p> <p>Contact Number : <u>8589965265</u></p>	<p>Customer Name : <u>Aleena</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>15/03/2024</u></p> <p>Contact Number : <u>04872844484</u></p> <p>Designation : <u>Nursing officer</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	--




SUPERINTENDENT
Community Health Centre
Pernjanam - 680 686

AR MEDICARE

ROOM NO: 865, 2ND FLOOR
AL-RAHA ARCADE, THAZHE CHOVVA
KANNUR-670018
EMAIL:armedicare07@gmail.com
MOB: 9446057350, 7994654108,



QUOTATION

REF: AR/QTN-03/24-25

Date: 02-04-2024

To,
Cyrix Health Care Private Limited
30/641B, Vandipetta Rd, near SBI Bank, Poonithura, Maradu
Ernakulam, - 682038
Kerala 682038

No.	Description	Qty	Unit Price(INR)	Total(INR)
1.	FETAL DOPPLER PROBE	1	2,500.00	2,500.00
2.	FETAL DOPPLER MAIN BOARD	1	3,500.00	3,500.00
			Grand Total	6,000.00 INR
			GST(12%)	720.00 INR
			Net Total	6720.00 INR
INR Six Thousand Seven Hundred And Twenty Only				

Terms and Conditions for sale :

- Validity:** One month from the date of quotation.
- Payment:** Within 30 days from the date of supply.
- Delivery:** Within 7 days after receipt of the signed purchase order.
- If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- Bank Details:
Account holder's name : AR MEDICARE
Bank Name : State Bank Of India, Thazhechovva
Account No. : 39648455024
IFSC code : SBIN0016861



**For AR MEDICARE,
KANNUR**

REQUIRED QUOTATION 0831829 #101211 fetal doppler @ CHC PERINJANAM , THRISSUR

2 messages

Sarang KM <arangcyrix@gmail.com>
To: ar medicare <armedicare07@gmail.com>
Cc: Bijo T Joy <bijocyrix@gmail.com>

Tue, Apr 23, 2024 at 12:34 PM

Dear sir,

Please share the price and stock availability of


1. FETAL DOPPLER PROBE
2. FETAL DOPPLER MAIN BOARD

Kindly do the needful asp

—
Thanks & Regards

Sarang.k.m
Divisional Incharge|North Zone
Bemp-Kerala
Mobile : +91 7593847150
Email : arangcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com



ar medicare <armedicare07@gmail.com>
To: Sarang KM <arangcyrix@gmail.com>

Tue, Apr 23, 2024 at 1:16 PM

Sir,
PFA



AR MEDICARE

Thanks & Regards

AR MEDICARE

Room No:865,2nd floor

AL-RAHA Arcade

Thazhe Chowva, Kannur

Mob: 9446057350, 7994654108, 8138091108

Email: armedicare07@gmail.com

GSTIN :32BDYPV5030P1ZP

[Quoted text hidden]

 **ARQ03. CYRIX DOPPLER ACCESSORIES.pdf**
535K