


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

PROFORMA

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	PHC KEEZHUPARAMBA
3	Name of Equipment with Make, Model and Serial Number	Equipment: UPS For Semi Make: APC Model: Back – UPS 600 Serial Number:E21628008435
4	Equipment ID/ Barcode	#104899 - 1042894
5	Date of purchase/ Year of manufacture/Installation Date	29/03/2017
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	05/04/2024 Toll Free
9	Action taken	Checked the machine and found problem with Main pcb & 12A 7AH battery damaged. Enquire OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	No recommendation
12	Cost of spares (specify parts and cost)	Main pcb – 2000/- 12v 7AH Battery – 1150/- TOTAL=3150/-

13	Asset Value	2341/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	147.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX service report and OEM EOL letter attached
16	Reasons for recommending the equipment as BER	Checked the machine found problem with Ups not on Main Pcb & 12v 7ah battery defective . Enquire OEM for quotation , The equipment installed on 29/03/2017 and aged up to 7 years. Repairing cost is 147.5% both criteria for RBR met. As per the tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	GOKUL KT 


\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

verified with the damage of Main PCB and Battery. Beyond economic repair.

  
22/4/24  
**Or. Consultant Biomedical Engineer**  
Signature of JC BM (NHM)  
NHM Malappuram



Date: 25/4/2024

  
**Superintendent/Medical Officer (i/c)**  
Kizhuparamba Health Centre  
Malappuram Dist - 673 639





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 191631

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <u>PHC</u></p> <p>Address ..... <u>KEEZHUPARAMBA</u> <u>(MALAPPURAM)</u></p> <p>Ph : ..... <u>9846 501911</u></p>	<p>Call Registration Date : ..... <u>5-4-2024</u></p> <p>Caller ID : ..... <u>104899</u></p> <p>Date of Visit : ..... <u>6-4-2024</u></p> <p>Asset No. : ..... <u>1042894</u></p> <p>EQPT Name : ..... <u>UPS</u></p> <p>Manufacture <u>Schneider</u> Model : <u>APC-600</u></p> <p>S. No. <u>E21628008435</u> Dept. <u>Laboratory</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :** UPS not on

**Action Taken :** checked the machine and problem found with main pcb and battery, Enquire these spare from oem for further checking of the equipment.

Completed  Date : 6-4-2024 Time : 1:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Naufal-m</u>	<u>6-4-2024</u>	<u>11:00 am</u>	<u>1:00 pm</u>

Customer Remark  Completed  Pending   
Jr. Consultant Biomedical Engineer  
NHM Malappuram

<p>Service Engineer Name : <u>Naufal-m</u></p> <p>Signature : <u>Naufal</u></p> <p>Date : <u>6-4-2024</u></p> <p>Contact Number : <u>7593847177</u></p>	<p>Customer Name : <u>Shabnam, O.P</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>6/4/24</u></p> <p>Contact Number : <u>Malappuram Dist - 682 038</u></p> <p>Designation : <u>Lab Technician</u></p> <p>Hospital Seal :</p>
---	---





# APS MARKETING COMPANY


27/233A, GROUND FLOOR, APS ARCADE  
NETHAJI ROAD, PUTHIYARA, CALICUT - 673004  
PH: 0495 - 2741313, Mob: 9447184933. Mail: info@apsmarketing.in

QUOTATION

Quotation No : APS-1861  
Date : 09-04-2024  
Valid Upto : 16-04-2024


Customer :  
CYRIX HEALTHCARE LTD  
ERNAKULAM

#	DESCRIPTION OF ITEM	QTY	RATE	TAXABLE	GST %	TOTAL
1	MAIN PCB BX600C-IN	1	2000.00	2000.00	0.00	2000.00
2	12V7AH BATTERY BX600C-IN	1	1150.00	1150.00	0.00	1150.00
<b>G-Total</b>						<b>3150.00</b>

  
22/4/24

Consultant Biomedical Engineer  
NHM Malappuram





Medical Officer  
Primary Health Centre  
Keezhuparamba,  
Malappuram Dist - 673 629

Amount Chargeable ( in words )  
Three Thousand One Hundred Fifty Rupees only

E & O.E

Remarks  
APC BX600C-IN (E21628008435)



FOR APS MARKETING COMPANY

Authorised Signatory





---

## 1042894 UpS need price quotation For main bord & battery

2 messages

---

**Gokul K T** <gokulcyrix@gmail.com>

Wed, 19 Jun, 2024 at 9:49 am

To: apsmarketingservice@gmail.com <apsmarketingservice@gmail.com>

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, Sarang KM <sarangcyrix@gmail.com>

Dear Sir, Madam

Can you send the Price quotation for this Main board & battery For UPS , Make:Apc , model :Back-Ups 600 Machine image attched below

Barcorde:1042894

Ticket id:104899

GOMUL KT  
CYRIX HEALTHCARE PVt Ltd  
DISTRICT INCHARGE  
9809314280  
MALAPPURAM

---

**aps marketing** <apsmarketingservice@gmail.com>

Wed, 19 Jun, 2024 at 1:01 pm

To: Gokul K T <gokulcyrix@gmail.com>

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, Sarang KM <sarangcyrix@gmail.com>

Hi


As per your request quotation attached...

[Quoted text hidden]









*[Signature]*  
**Medical Officer**  
 Primary Health Centre  
 Keezhuparamba,  
 Malappuram Dist - 673 6



*[Signature]*  
 22/4/24

**Assistant Biomedical Engineer**  
 NHM Malappuram