


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	TB CENTRE KANNUR
3	Name of Equipment with Make, Model and Serial Number	UPS Make: ALFA Model: NA SN: NA
4	Equipment ID/ Barcode	#111041- 1312304
5	Date of purchase/ Year of manufacture/Installation Date	12-05-2016
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 14-05-2024
9	Action taken	Checked the Equipment found Equipment is not Switching ON, found Main board and batteries are Defective. Need to replace the main board for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Main PCB- 9800/- Service charge -2000/- Total -11800/-

13	Asset Value	4500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	262.222%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not Switching ON, found Main board and batteries are defective. Need to replace the main board for further checking Equipment installed on 12-05-2016 aged up to 8 years. Repairing cost is 262.222% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that UPS not switching ON. found that main board, battery are defective. need to replace the main board. Aged up to 8 years. The service cost economically not feasible. So recommended for BER as per BER protocol.


Signature of JC BM (NHM)

21/05/2024
Radhika.C.B

Shri: mte
Lab technician
DCC, Kannur


CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002



for OTO
Signature of
Superintendent/Medical Officer (i/c)
District T.B. Officer
District T.B. Centre
P.O. Pallikunnu
Kannur-670004





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/098

CYRIX

HEALTHCARE PVT LTD

No. : 176853

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/84 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>T.B. Centre Kanneer</u></p> <p>Address : <u>Kanneer</u> <u>Kerala</u></p> <p>Ph : <u>9605353961</u></p>	<p>Call Registration Date : # <u>14/05/24</u></p> <p>Caller ID : # <u>111041</u></p> <p>Date of Visit : <u>15/05/24</u></p> <p>Asset No. : # <u>1312304</u></p> <p>EQPT Name : # <u>C.P.S</u></p> <p>Manufacture <u>A.L.F.A</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Lab</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : # machine not switching on

Action Taken : # Visited site and checked the equipment and equipment is not switching on. found main board and batteries are defective need to replace the main board for further checking

Completed Date : 15/05/24 Time : 11:40am Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrrix Engineer	Date	Start Time	End Time
<u>Shobin VS</u>	<u>15/05/24</u>	<u>10:15am</u>	<u>11:40am</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Shobin VS</u></p> <p>Signature : </p> <p>Date : <u>15/05/24</u></p> <p>Contact Number : <u>7593847109</u></p>	<p>Customer Name : <u>Shri...</u></p> <p>Signature : </p> <p>Date : <u>15/05/24</u></p> <p>Contact Number : <u>9605353961</u></p> <p>Designation : <u>Lab technician</u></p> <p>Hospital / Dept : <u>T.B. Centre Poonithura</u></p>
---	---



For DTO
 District T.B. Officer
 Poonithura T.B. Centre
 Poonithura
 15/05/24
 9605353961
 Lab technician

KNR/QTN/SVS/009/24-25
16th May 2024

To

M/s: CYRIX HEATH CARE (P) LTD,
30/641B, PETTA, PONNITHURA,
ERANAKULAM (DT), KERALA.

Dear Sirs,

Sub: - Service Quotation for "ONLINE UPS" Reg.
Ref: - As per your Service request at "TB Centre, Kannur".

With response, we are quoting our lowest special rates for servicing UPS as follow:

Sl.	Particulars	Amount
1.	Cost Of Spares - Main PCB	9800.00
2.	Service Charge	2000.00
Total		11800.00
GST @18%		Included
Grand Total		11800.00

If any other clarification both technical & commercial, please contact the same. We expect your Valuable order and we assure you for a best support.

Terms & Conditions:-

Taxes & Duties : Inclusive of all taxes.
Payment : 100% advance along with confirmed work order.

Yours truly,

Prakash | Manager
Zenon Power Solutions
Star Arcade, Opp. Chaithanya Hospital,
Talap, Kannur, Kerala - 670002, India.
T +91 497 2700213 | M +91 9562077737.
Visit: www.zenonpower.com





Service quotation' for UPS at - TB Centre, KNR (16th May 2024)

1 message

Zenon Power Solutions <zenonpower@gmail.com>
To: avinashtcyrix@gmail.com
Cc: shebincyrix@gmail.com

Thu, 16 May, 2024 at 12:27 pm

Dear Sirs,

With response, we would like to attach the Service Quotation for Existing " ONLINE UPS" at TN Centre - Kannur.

OUR KEY POINTS:-

Providing best Service & AMC for existing any kind of UPS, INVERTER etc.

Best Exchange/Upgrade value from the time of new purchases.

More than 200 banking / 450 Corporate, Hospital & Institutional customers.

Best quality, Best Price, Best after Sales Supports.

More details; please logon www.zenonpower.com

If any other clarification is both technical & commercial, please contact the same.

We expect your Valuable order and we assure you for the best after sales support.

Yours truly,

Prakash | Manager – Service.

Zenon Power Solutions | Star Arcade | Opp : Chaithanya Hospital,

Talap | Kannur | Kerala - 670002 | India.

T +91 497 2700213 | M +91 9562077737.

Visit: www.zenonpower.com



H. S. S. S.
 S. S. S. S. S. S.
 Lab technician
 D T e ken man