


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	TH NILESWAR
3	Name of Equipment with Make, Model and Serial Number	Name:- Digital BP apparatus Make:- Dr.morphen Model:-BP02XL SN:- BA499200511406D
4	Equipment ID & Barcode	96324 1420254
5	Date of purchase / Year of manufacture /Installation Date	28-03-2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	15-02-2024 (Toll free)
9	Action taken	Checked the machine and found that main board, display board, motor are defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendation.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	2200
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED.
16	Reasons for recommending the equipment as BER	Checked the machine and found that main board, display board and motor are defective. Machine installed on 28/3/2018 and aged up to 5 + years. Spare quote not submitted since spares are not available in market and so recommending for Condemnation.
17	Name & Signature of CYRIX Authority	VISHNU P 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
 checked the machine and found main board, display board, and motor are defective. Equipment installed on 28/03/2018. As per cyrix, spares are not available in market. So recommending the equipment for BER.

Signature of JC BM (NHM)


 J. M. M. Engineering
 Biomedical Engineering (NHM)
 Trogyakeralam (NHM)
 Kozhikode, Kerala

Date 21/3/24



Signature of
 Superintendent Medical Officer (i/c)
 സൂപ്രണ്ടി
 താലൂക്ക് ഹെഡ് ക്വാർട്ടേഴ്സ് ആശുപത്രി
 നിലേശ്വരം, പി. ഒ. നിലേശ്വരം - 671314



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 122791

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>TH Nileswar</u></p> <p>Address <u>Nileswar</u> <u>Kasaragod</u></p> <p>Ph : <u>9496239794</u></p>	<p>Call Registration Date : <u>15/02/24</u></p> <p>Caller ID : <u>96324</u></p> <p>Date of Visit : <u>16/02/24</u></p> <p>Asset No. : <u>1420254</u></p> <p>EQPT Name : <u>Digital BP Apparatus</u></p> <p>Manufacture <u>Dr. moephan</u> Model : <u>BPO2XL</u></p> <p>S. No. <u>BAP-9900511466</u> Dept. <u>Palliative</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not switching on.

Action Taken : Checked the machine and found problem with main bowl, display and monitor. Need to replace these spares for further checking of the complaint.

Completed Date : 16/02/24 Time : 11:00 Am Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Jayesh kushnan</u>	<u>16/02/24</u>	<u>10:30 Am</u>	<u>11:00 Am</u>

Customer Remark Completed Pending

Service Engineer Name : Jayesh kushnan

Signature : [Signature]

Date : 16/02/24

Contact Number : 7593847156

Customer Name : [Signature]

Signature : [Signature]

Contact Number : 1612123
9400869161

Designation : Nursing officer UNP



Handwritten notes and signatures in the bottom right corner.

	BP apparatus	OB	Receipt	Balance
17/10/01	Received from mo		1	1
31/4/2007	Received from mo TU nilesbwer	1	1	2
10/2013	from Untied fund C.A	2	1	3
28/3/2018	from Untied fund Digital BP apparatus		1	
31/3/2019		Broken -	1	1 (Not working condition)



Handwritten signature
 സമുപദേശ
 കമ്മ്യൂണിക്കേഷൻ ഓഫീസർ
 നിലമ്പൂർ, പി.ഒ. നിലമ്പൂർ - 671314