


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC OLLUR
3	Name of Equipment with Make, Model and Serial Number	Refrigerator Make - Whirlpool Model - nill S/N - NA
4	Equipment ID/ Barcode	#98210 - 0832342
5	Date of purchase/ Year of manufacture/Installation Date	30/03/2004
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	26-02-2024
9	Action taken	Checked the equipment refrigerator not cooling and found compressor, gasket, timer, thermostat, freezer door faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	Door gasket -2,700/- Compressor& gas charging - 7,500 Timer - 850 Thermostat - 1000 Freezer door - 500 TOTAL - 12,550/-

13	Asset Value	19,900/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	63%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report and Quotation attached
16	Reasons for recommending the equipment as BER	Checked equipment found Compressor, gasket, freezer door, thermostat and timer are faulty. equipment installed on 30-03-2004 and aged up to 19 years 11 month, repair cost is 63% and both criteria met for RBER. As per tender clause 5.3.14.1 we are recommending equipment for condemnation.
17	Signature of CYRIX Authority	Bijo 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Very old and damaged.


Signature of Junior Consultant (Biomedical) NHM



Date


Signature of Superintendent/Medical Officer (i/c)

Handwritten Malayalam text:
സാമൂഹിക ആരോഗ്യ കേന്ദ്രം
ബി.സി.പി. ഓഫീസ്



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 189929

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>C.H.C</u></p> <p>Address <u>ollur</u> <u>thirissur</u></p> <p>Ph : <u>9447523680</u></p>	<p>Call Registration Date : <u>26/2/24</u></p> <p>Caller ID : <u>98210</u></p> <p>Date of Visit : <u>27/2/24</u></p> <p>Asset No. : <u>0832342</u></p> <p>EQPT Name : <u>pharmacy refrigerator</u></p> <p>Manufacture <u>whisper</u> Model : <u>vill</u></p> <p>S. No. <u>NA</u> Dept. <u>op</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Cooling issue. temperature high

Action Taken : checked the equipment refrigerator not cooling and found compressor gasket, timer, thermostat, freezer door faulty, need to replace these spares for further checking

Completed Date : 27/2/24 Time : 3:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jinto</u>	<u>27/2/24</u>	<u>12:00pm</u>	<u>3:00pm</u>

Customer Remark Completed Pending



<p>Service Engineer Name : <u>Jinto Jeyathirissur</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>27/2/24</u></p> <p>Contact Number : <u>7593847125</u></p>	<p>Customer Name : <u>Bonimol Jikwianel</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>27/2/24</u></p> <p>Contact Number : <u>9447523680</u></p> <p>Designation : <u>Nurse</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
--	---

682038, Poonithura, Kochi - 682 038

Cool House

Deto

Opposite SBI BANK, Kodakara Thrissur

☎ 9895171425

Quotation

To,
Cyrix Healthcare Pvt Ltd
Kochi, Ernaakulam
☎ 9847299500

Quotation# Quote-257

Date: 21-03-2024

Dear Sir/Mam,

Thank you for your valuable inquiry. We are pleased to quote as below:

#	DESCRIPTION	QTY	PRICE	TOTAL
1	Compressor with Gas charging	1 Nos	₹ 7,500.00	₹ 7,500.00
2	Door Gasket	1 Nos	₹ 2,700.00	₹ 2,700.00
3	Thermostat	1 Nos	₹ 1,000.00	₹ 1,000.00
4	Timer	1 Nos	₹ 850.00	₹ 850.00
5	Freezer door	1 nos	₹ 500.00	₹ 500.00
GRAND TOTAL				₹ 12,550.00

We hope you find our offer to be in line with your requirement.

COOL HOUSE For COOL HOUSE
Refrigeration & Air Condition
Opp. SBI Kodakara- 680 682


AUTHORIZED SIGNATURE



JkU



JkU

[Handwritten signature]



OLUR THRISSUR 680 006

680 006
680 006


Refrigerator 165 to 175 liter c¹

Balac

Receipt

03/2004	IN-NO-2277 Scientific			
	Ad for TSR	1		1 M
	Whirlpool			
②	Stabilizers (V guard)	1		1 M
	Issued to Chemistry Dept		1	nil M
	Issued to Chemistry (Stabilizer)		1	nil M
	Issued to lab			

Received
 J. K. S.
 Lab Dept



[Signature]
 மதுராட்சி மன்றம்
 மதுராட்சி மன்றம்
 கட்டிட எண் 680 306