


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PHC EDAKKAD
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE Make : SAMSO Model: NA SN:N/A
4	Equipment ID/ Barcode	1344504 #107047
5	Date of purchase/ Year of manufacture/Installation Date	07-03-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 18-04-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found mainboard, Load cells damaged. Need to replace the these spares for further checking .
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	




13	Asset Value	715/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 07-03-2019. aged up to 5 years . Quotation not submitted since Spare parts are not available on market. So we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that machine not working due to load cell damage, main board issue. aged up to 5 years. so recommended for BER due to service cost economically not feasible. 
13/05/2024.

Signature of JC BM (NHM)

CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

Date: 24/5/24



Signature of
Superintendent/Medical Officer (i/c)



MEDICAL OFFICER
PRIMARY HEALTH CENTRE, EDAKKAL
P.O. ATTADAPPA, KANNUR - 670 002



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 183098

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>phe edakkad</u></p> <p>Address : <u>Kannur</u> <u>Kerala</u></p> <p>Ph : <u>9567469180</u></p>	<p>Call Registration Date : <u>18-4-24</u></p> <p>Caller ID : <u>107047</u></p> <p>Date of Visit : <u>19-4-24</u></p> <p>Asset No. : <u>1344504</u></p> <p>EQPT Name : <u>weighing Machine</u></p> <p>Manufacture : <u>sansa</u> Model : <u>NA</u></p> <p>S. No. : <u>NA</u> Dept. : <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Main boards, lead cells damaged

Action Taken : checked the machine and found that machine re-
had soldering on found main board, lead cells damaged
Need to replace these spares for further checking.

Completed Date : 19-4-24 Time : 11:40 Am Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Shyamkrishnan</u>	<u>19-4-24</u>	<u>11:00 Am</u>	<u>11:40 Am</u>

Customer Remark Completed Pending



Service Engineer Name : Shyamkrishnan

Signature : [Signature]

Date : 19-4-24

Contact Number : 8201188514

Customer Name : balakrishnan A.V.

Signature : [Signature]

Date : 19/4/24

Contact Number : 9567 469 180

Designation : Medical Officer

Hospital Seal : PRIMARY HEALTH CENTRE, EDADAKKAD, P.O. ATTADAPPA, KANNUR - 670 008

8 'WEIGHING MACHINE'
 EDUKKAD - HANT - (2000 CAPACITY 100kg) PL

INVOICE NO DATE	PREVIOUS when reported when used	Real	Issue	Price	Value	Notes
73-2014 11-3-2014	14/11/2014 11-3-2014	2000	2000	2000	1700.50	
	14/11/2014					80131594
						90131575



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
1344504

