


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	PHC KUNNAMANGALAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : WEIGHING MACHINE Make : SAMSO Model : NA SN : NA
4	Equipment ID & Barcode	80798 & 1143172
5	Date of purchase / Year of manufacture /Installation Date	12- 10- 2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	17/11/2023(Toll free)
9	Action taken	Checked and found Main board, display and load sensor defective need to replace these spares for further checking working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	5000
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found main board, display and load sensor defective need to replace these spares for further checking and working condition of the equipment. Equipment installed on 12/10/2019 aged up to 4years 1months. Quotation not submitted since spares are not available in the market So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	AMRUTHA M 


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked and found that main board, display load sensors are damaged. Equipment installed on 12/10/2019. So recommended for condemnation. Spare are not available in the market


 Signature of JC BM (NHM)
 Consultant Bio Medical Engineer
 National Health Mission
 Kozhikode

Date




 Signature of
 Superintendent / Medical Officer (i/c)
 MEDICAL OFFICER
 Primary Health Centre
 Kunnammangalam



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 124018

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC KUNNAMANGALAM</u></p> <p>Address <u>KOZHIKODE</u></p> <p>Ph : <u>8129753282</u></p>	<p>Call Registration Date : <u>17-11-2023</u></p> <p>Caller ID : <u>80798</u></p> <p>Date of Visit : <u>18-11-2023</u></p> <p>Asset No. : <u>11A-3172</u></p> <p>EQPT Name : <u>Weighing Machine</u></p> <p>Manufacture <u>SAMSO</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Public Health-sub Center</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not getting on.

Action Taken : Checked and found main board, display and load sensor defective, need to replace these spares for further checking working condition of these equipment.

Completed Date : 01-12-23 Time : 1 pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Abijith.</u>	<u>01-12-23</u>	<u>12.30 pm</u>	<u>1 pm</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Abijith A</u></p> <p>Signature : </p> <p>Date : <u>01-12-2023</u></p> <p>Contact Number : <u>7593847141</u></p>	<p>Customer Name : <u>Abijith</u></p> <p>Signature : </p> <p>Date : <u>01-12-23</u></p> <p>Contact Number : <u>7593847141</u></p> <p>Designation : <u>MEDICAL OFFICER</u></p> <p>Hospital Seal : </p>
---	---



KARUNYA MEDICINE DEPOT

SWARAJ ROUND EAST, TRISSUR

PH: 0487-2975601 EMAIL : kcpdepottr@gmail.com

State : Kerala

State Code : 32

DL No1: KL-TSR-139473, 1

DL No2: KL-TSR-139474, 1

[ORIGINAL]

INVOICE

GSTIN : 32AADCY4029M1ZK

Name : KZH-KUNNAMANGALAM-PHC

Cust.Code : KUNNY

**Address : MEDICAL OFFICER
PRIMARY HEALTH CENTRE
KUNNAMANGALAM**

Phone No :

GSTIN :

DL No1 :

DL No2 :

Invoice No : 4057

Date : 12/10/2019

Time : 04:14 PM

Prepared By : SINDHU K B

Page 2 of 2

Sl#	MFR	Product Name	KCP S.P	Exp.Dt	MRP	Batch	Unit	Qty	Rate	CGST %	CGST Amt	SGST %	SGST Amt	Amount
22	GE	URINE STRIP (GLUCOSE & ALBUMIN) 100 TEST/KIT - LIFECARE	102.68	May/2021	434.00	URS9K10222	NOS	9.0	102.68	6.00	55.45	6.00	55.45	924.12
23	GE	WEIGHING MACHINE (DIGITAL PAEDIATRIC) - PE	4239.44	Dec/2030	6000.00	111	NOS	1.0	4239.44	9.00	381.55	9.00	381.55	4239.44
24	GE	WEIGHING MACHINE (DIGITAL ADULT - SCALE CAPACITY 150 KGS) - PE	1608.04	Jan/2028	2590.00	113	NOS	1.0	1608.04	9.00	144.72	9.00	144.72	1608.04

CGST: 1420.33	SGST: 1420.33	CESS: 189.35	Total: 21964.98
			Handling Charge: 1325.45
			Handling Charge Gst18% : 238.58
			Round off: -0.01
MRP Amount: 47131.00			Net Amount: 23529.00

Rupees: Twenty Three Thousand Five Hundred and Twenty Nine

Terms & Conditions

Sales returns only within 10 days from the date of billing
 Sales return -10AM to 4PM Only
 No Sales return on Sunday
 * Products to be kept in regrigerator will not be taken back after sales.



AUTHORISED SIGNATORY

STOCK BOOK OF

Name of article

SUB CENTER KIT.

Date	No and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought Forward					
3/12/2019		Received from UH					
1		BP apparatus.	1	Put in use		3/12/19	
2		Dust Bin - black	1	Put in use		3/12/19	
3		Dust Bin - Red	1			"	
4		Dust Bin - yellow	1			"	
5		Examination torch	1			"	
6		Fotoscope	1			"	
7		Gloves sterile	25			"	
8		Glucometer strips	300			"	
9		Glucometer - on call	1			"	
10		Hemometer test strip	1			"	
11		Hemometer	1	Put in use		3/12/19	
12		Club cutter	1			"	
13		Infantometer	1			"	
14		Kit for testing Residual Chlorine in DW.	1			"	
15		Carrel	100			"	
16		measuring tap	1			"	
17		Nebulizer	1			"	
18		Salt testing but for room	5			"	
19		Stadiometer	1			"	
20		Stethoscope	1			"	
21		Thermometer Digital	1			"	
22		Urine strip	9.			"	
23		weighing machine	1			"	
		(Digital paediatric)		put in use		3/12/19	

