


**Quotation-027-Re: REGARDING QUOTATION - 103856 - 0110011 - GH THIRUVANANTHAPURAM - DIALYSIS MACHINE**

Grahams Pharmaceuticals &lt;grahamspharmaceuticals@yahoo.com&gt;

Fri 5/10/2024 11:43 AM

To: Anjitha T L &lt;purchasecyrix2@gmail.com&gt;

Cc: Cyrix Purchase Mgr. &lt;purchasemgr@cyrix.in&gt;; Abijith A &lt;abijithacyrix@gmail.com&gt;; Afsal Yunuze &lt;zm1.klbemp@cyrix.in&gt;; Kasyep P V &lt;kasyepcyrix@gmail.com&gt;; Assistant Manager Purchase &lt;ampurchase@cyrix.in&gt;; D. Gunasekaran &lt;d.gunasekaran@freseniusmedicalcare.com&gt;

 1 attachments (56 KB)

Quotation-027 Cyrix-10-5-24.pdf;

Dear Sir,

Please find the Quotation.

This is with reference to our discussion regarding purchase of **Fresenius Hemodialysis Machine spares.**

Further, we would like to inform you that we are the authorised distributor of **M/s Fresenius Medical Care India Pvt Ltd** and are currently operating from Cochin.

We are pleased to quote our best lowest possible rates for your esteemed Institution & hope that you will find the offer competitive & look forward to have favourable response.

Thanking you and assuring you our best services at all the time.

Regards,

Tesmol

**Grahams Pharmaceuticals**

46/2373/C,

Chakkaraparambu Road,

Kottankavu Junction,

Vennala P.O,

Ernakulam. PIN:- 682 028

Office: 0484- 2807376,

0484 -2808376

*"Do not dwell in the past, do not dream of the future, concentrate the mind on the present moment."*

On Friday, 10 May, 2024 at 11:42:00 am IST, Anjitha T L <purchasecyrix2@gmail.com> wrote:

Dear Sir,

Kindly share a quotation for the following spare of Dialysis Machine  
Make- Fresenius

Model- 4008B

Hospital Name: GH Thiruvananthapuram

- 1) Modification Kit Power Supply - M502301 - 1no
- 2) PCB LP630 Motherboard - M473721 - 1no
- 3) PCB LP634 SMT Output board -M461411 - 1no
- 4) Membrane pump Conc/Bic -M307151 - 2no
- 5) Ultrafiltration pump -6707061 -1no
- 6) Flow pump motor with plug -M375791 -1no
- 7) Stepper motor with gear box 4-poles - M393211 - 1no
- 8) Balancing chamber complete -6718071 -1no

--

**Thanks & Regards**

**Anjitha T L**

**Purchase Assistant**

**PH:- 9947719111**

[purchasecyrix2@gmail.com](mailto:purchasecyrix2@gmail.com)

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

**Cyrix Healthcare Pvt Ltd**

**| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India**

**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**



**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 161126

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>G.H. Thiruvananthapuram</u>	Call Registration Date : <u>29/02/2024</u>
Address <u>Thiruvananthapuram</u>	Caller ID : <u>103856</u>
<u>Kerala</u>	Date of Visit : <u>02/04/2024</u>
Ph : <u>7012995543</u>	Asset No. : <u>0110011</u>
	EQPT Name : <u>Dialysis Machine</u>
	Manufacture <u>Presentis</u> Model : <u>4003 B</u>
	S. No. <u>155ATM70</u> Dept. <u>Dialysis Unit</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Flow Alarm - 20% During Dialysis  
Multiple Complaints

Action Taken : Checked and Found Flow Alarm - 20% during dialysis  
Problem Suspecting with mother board (PCB LP630)  
Need to replace mother board. After replacing the mother  
board only can check the further service.

Completed  Date : 02/04/24 Time : 05:30 P.M Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. Mother Board LP630	01		
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Syam Surendran</u>	<u>02/04/24</u>	<u>12:00 PM</u>	<u>05:30 PM</u>

Customer Remark  Pending   
Not Ready - Office closed seal Not Available.

Service Engineer Name : Syam Surendran  
Signature : Syam Specialist Engr Dialysis  
Date : 02/04/2024  
Contact Number : 8078442537

Customer Name : Santhosh N R  
Signature : [Signature]  
Date : 21/4/24  
Contact Number : 9449452008  
Designation : klungy officer.  
Hospital Seal :




Superintendent  
General Hospital  
Thiruvananthapuram

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANATHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Dialysis Machine Make : Fresenius Model: 4008B SI :1V5ATM70
4	Equipment ID/ Barcode	103856 / 0110011
5	Date of purchase/ Year of manufacture/Installation Date	16/04/2012
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	29/03/2024
9	Action taken	Checked and found Modification kit power supply,PCB LP630 motherboard, PCB LP 634,membrane pump,Ultrafiltration pump,Flow pump motor with plug,steppermotor with gear box 4 poles and balancing chamber defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair.
12	Cost of spares (specify parts and cost)	Modification kit power supply-103524 PCB LP630 Mother Board - 44963 PCB LP634 Output Board - 42469 Membrane Pump - 31515 Ultrafiltration Pump - 28469 Flow pump motor with plug - 11170 Stepper Motor with Gear box - 25574 Balance Chamber 4008s - 41163  Total = 386148/-

13	Asset Value	Rs. 435000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	88.76%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report with OEM quotation attached.
16	Reasons for recommending the equipment as BER	Checked and found Modification kit power supply, PCB LP630 Mother Board, PCB LP634 Output Board, Membrane Pump, Ultrafiltration Pump, Flow pump motor with plug, Stepper Motor with Gear box, Balance Chamber 4008s are defective. Equipment installed on April 2012 and covered up to 12+years.the repair cost is 88.76%. As per the tender clause 5.3.14.1 ,both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Mother boards, Pump etc need to be replaced. Machine is old. Also the rate of repair is 88.76% asset value. ∴ can be BER.

**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
13-05-2024  
Signature of JC BM (NHM)

Date:



  
Signature of  
Superintendent/ Medical Officer (i/c)  
Superintendent  
General Hospital  
Thiruvananthapuram

STOCK REGISTER OF... Dialysis Unit  
 Name of Article... DIALYSIS MACHINE

01.16.12  
 16/4/12  
 16/4/12

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		through forward					
16/4/12		Dialysis machine (Donated)	(5)		Cost 4,37,000 per unit		Stack Pg. Page No. 6. Vol. (Donated Item)
		[Fresenius 4008B]					
16/4/12		Dialysis machine	(4)		Cost 5,91,119 per unit		Stack Pg. P. no. 271, Vol. 8
		[B. Braun]					(9)

16/4/12  
 16/4/12  
 16/4/12

29/01/2022 Received from Stone (2) Haemo (113) dialysis Machine SURDIAL SS plus, Sundial ss plus (2) Nipro medical, India Pvt, Ltd, 19CX 14761 and 19CX 1462. Module 01/5/20 exp 30/4/2030 (2) Nephrology. @ Rate of 87,200/- piece. Total Value 4,87,200/- + 4,87,200 = 9,74,400; Tax collected at source = 730. Total = 97530

Issue No: (MIN) No: K1 00037/1P/00979/Janet  
 dtd 15/10/20

Stack Page No. 11-210 and 11-211  
 Total = 97530

Carried over

# GRAHAMS PHARMACEUTICALS

46/2373C, Chakkaraparambu Road, Kottankavu Junction, Vennala PO, Cochin-28,  
Ph: 0484-2807376, 2106068, 2808376

E-mail: grahamspHarmaceuticals@yahoo.com

KER/MAY24/027

Quotation						
Customer: Cyrix Healthcare Pvt Ltd (G H Trivandrum)					Date: 10-05-2024	
					Equipment: HD Machine	
					Serial No:	
S. No.	Description	Part No	Rate ( Rs)	GST	Qty	Amount
1	MODIFICATION KIT POWER SUPPLY *4008B/S *4008S	M502301	103524	18%	1	103524
2	P.C.B. LP 630 MOTHERBOARD *4008B/H/S *4008S V1	M473721	44963	18%	1	44963
3	P.C.B. LP 634 OUTPUT BOARD *4008B/H/S *400S V10	M570571	42469	18%	1	42469
4	Membrane Pump 4008(84138190)	M307151	31515	12%	1	31515
5	ULTRAFILTRATION PUMP *4008B/H/S *4008S V10 *4	6707061	28469	18%	1	28469
6	FLOW PUMP MOTOR WITH PLUG *4008B/H/S *4008S	M375791	11170	18%	1	11170
7	STEPPER MOTOR WITH GEAR BOX 4-POLES *4008B/H	M393211	25574	18%	1	25574
8	BALANCE CHAMBER 4008B/H/S 4008S V10	6718071	41163	18%	1	41163
9						
10						
11						
12						
13						
14						
15						
Rupees : Three Lakh Eighty Six Thousand One Hundred and Forty Eight Only				TOTAL		328847
				GST		57301.5
				GRAND TOTAL		386148

1. Payment Terms : Immediate After Delivery

2. Quotation Validity: 30 Days

For any further clarification, Kindly contact

Mr. Gunasekaran D(8086026265)

