



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No.: 190680

485

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.ki@cyrilx.in

Health Facility : <u>THQH</u>	Call Registration Date : <u>4.2.2024</u>
Address : <u>Chirayinkeezh</u>	Caller ID : <u>99514</u>
<u>Thiruvananthapuram</u>	Date of Visit : <u>4.2.2024</u>
Ph : <u>9027406659</u>	Asset No. : <u>0121066</u>
	EQPT Name : <u>Multispectrum</u>
	Manufacture <u>BPL</u> Model : <u>Excella Eco</u>
	S. No. <u>DATQAJ4525</u> Dept. <u>OT</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Refrigerator sensor board after servicing, still machine is not functional. Crosschecked with another machine and found that leadlines to sensor board, BM 100 module also defective. Need to replace sensor board and BM100 module for further checking of the machine. Checked for spares with OEM.

Completed Date : 16.4.2024 Time : 12:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Krishna Prakesh VS</u>	<u>16.4.2024</u>	<u>11:15 am</u>	<u>12:00 pm</u>

Customer Remark



Completed Pending

Dr. P.K. Prakesh
SUPERINTENDENT
Talker Headquarters Hospital
Chirayinkeezhu

Service Engineer Name : Krishna Prakesh VS
Signature : [Signature]
Date : 16.4.2024
Contact Number : 9072522086

Customer Name : Prakesh VS
Signature : [Signature]
Date : 16/04/2024
Contact Number : 9037404659
Designation : Biomedical Engineer
Hospital Seal :

485


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	THQH CHIRAYINKEEZH
3	Name of Equipment with Make, Model and Serial Number	Equipment : Multipara monitor Make : BPL Model : Excello eco Serial Number : DATQ4J4525
4	Equipment ID/ Barcode	ID : 99514 / Barcode : 0121066
5	Date of purchase/ Year of manufacture/Installation Date	23/12/2014
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	04/03/2024
9	Action taken	Checked and found BM100 module board and ARM board defective. Enquired the spare from vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	BM100 Module Board - 38940/- ARM Board - 23010/- Total - 61950/-

PERMIT TO REPAIR

13	Asset Value	Rs. 84900/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	72.96%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report and OEM Quotation Attached.
16	Reasons for recommending the equipment as BER	Checked and found BM100 module board and ARM board defective. The equipment was installed on 23/12/2014 and covered upto 9+ years. Repairing cost of the equipment is 72.96%. As per the tender clause 5.3.14.1 both criteria met, So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**


* Attach Photograph



Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

As per the RBER criteria, this can be recommended BER. as the repair cost is high.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)

National Health Mission
Thiruvananthapuram


10-05-2024
Signature of JC BM (NHM)



Signature of
Superintendent/Medical Officer (i/c)
SUPERINTENDENT
Head Quarters Hospital
Chirayinkeezhu

Date: 10-05-2024

248

Multiparameter monitor
with reusable neonatal pads

B/F P. No 249

23/12/14 Rd from KMSCL, Trivandrum

Supplied by BPL Medical Technologies Pvt Ltd, Cochin
Invoice no ERM/FG/50378 dt 23/12/14

NRHM/ child health.

BPL MPM Exello - Eco

Model - Exello Eco

Make - BPL Medical Technologies Ltd.

Eqpt code - 75 - 245

Rs	78000.00
	<u>2852.00</u>
	20552.00
5% vat	4042.85
	<u>15</u>
	31960.00

I No. I No.

23/12/14 Issued to Med. wd

I No: not given

11/3/15 Rd from KMSCL, Triv.

(CHS PS 12-13 / Trauma care unit) Invoice - 9846969584

Supplied by BPL Medical Techno.

Invoice no ERM/FG/50447 dt 23/12/14

BPL MPM Exello - Eco

Model - Exello - Eco warranty syn (10/3/15)

Eqpt code - 75 - 279

	78000.00
	<u>3900.00</u>
	81900.00

I No. I No.

11/3/15 Issued to casualty

I No: not given
P. V. K. P. K.

Trauma care

VA. page no: 11



SUPERINTENDENT

Thaluk Hospital

Chirayinkeezhu



QTN/BISO/040424

04/04/2024

QUOTATION

To,
CYRIX HEALTH CARE PVT LTD
ERNAKULAM

Dear Sir,

Please find below our lowest quotation for the following items:-

SL NO.	ITEM DESCRIPTION	COST PER UNIT	UNIT PRICE
1	BM100 MODULE BOARD -BPL MODEL FOR EXCELLO ECO MULTIPARA MONITOR	33000	33000+18%TAX
2	ARM BOARD FOR EXCELLO ECO MULTIPARA MONITOR	19500	19500+18%TAX


Terms and conditions:

1. Payment 100% in advance along with Purchase order
2. Delivery within 15 days

Thanking you and assuring you our best attention always.

For Biomedix Solutions




Authorized Signatory

Mobile: 9846969564 | Email: biomedixservice@gmail.com

Branches: Kozhikode, Kannur, Kottayam-Kollam

For enquiries contact: biomedixsales@gmail.com, biomedixservice@gmail.com



BIOMEDIX SOLUTIONS

No: 7/437- B, Mannarakayam P.O,
Kathilankalpadi, Kanjirapally, Kottayam - 686506,
Mob:9846969564,



SUPERINTENDENT
T. K. H. Quarters Hospital
Chirayinkeezhu

Re: REGARDING QUOTATION - 99514 - 0121066 - THQH CHIRAYINKEEZH - MULTIPARA MONITORAbijith A <abijithacyrix@gmail.com>

Thu 5/16/2024 3:37 PM

To:Purchase Cyrix <purchasecyrix2@gmail.com>Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Anjitha,

Please share mail communication

On Wed, Apr 3, 2024 at 3:19 PM Abijith A <abijithacyrix@gmail.com> wrote:

Reminder

On Sat, Mar 16, 2024 at 10:05 AM Abijith A <abijithacyrix@gmail.com> wrote:

Dear Anjitha,

Please arrange a quotation.

----- Forwarded message -----

From: **Abijith A** <abijithacyrix@gmail.com>

Date: Thu, Mar 14, 2024 at 1:48 PM

Subject: REGARDING QUOTATION - 99514 - 0121066 - THQH CHIRAYINKEEZH - MULTIPARA MONITOR

To: Biomedixservice <biomedixservice@gmail.com>, BISO Homecare <biomedixqln@gmail.com>Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, <tanimachinnucyrix@gmail.com>

Dear Sir,

kindly share the quotation and stock availability of BM100 module and arm board for Make-Bpl Model-Excello Eco Multipara Monitor.

Please do the needful.

--

Thanks & Regards,

Abijith A**Divisional Incharge****Mob-7593847141**abijithacyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

--

Thanks & Regards,