



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 197775

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>Che</u></p> <p>Address <u>Che Mappanad</u> <u>Kollam</u></p> <p>Ph : <u>94416445717</u></p>	<p>Call Registration Date : <u>22/3/2024</u></p> <p>Caller ID : <u>102698</u></p> <p>Date of Visit : <u>23/3/2024</u></p> <p>Asset No. : <u>0280603</u></p> <p>EQPT Name : <u>Stirizer</u></p> <p>Manufacture <u>Mdicraft</u> Model : <u>12x12</u></p> <p>S. No. <u>NA</u> Dept. <u>Hospital</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
Coil damaged, chamber broken & making

Action Taken :
Checked and found that Stirizer Blasted coil damaged & chamber broken it's not separable, as proof images attained

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Rajesh</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>25/03/2024</u></p> <p>Contact Number : <u>7593844128</u></p>	<p>Customer Name : <u>Suni Joseph</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>25/3/24</u></p> <p>Contact Number : <u>9445021294</u></p> <p>Designation : <u>ASST SURGEON</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
--	--




Tresa Leen 574134
DR. TRESA LEEN
MBBS, FCCP
Reg No: 26620
Asst. Surgeon
PHC Mappanad

REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOLLAM
2	Name of Hospital	CHC MAYYANAD
3	Name of Equipment with Make, Model and Serial Number	Equipment : STERLIZER Make: MEDICRAFT Model: 12X12 SN : NA
4	Equipment ID & Barcode	Ticket id: 102698 / Barcode : 0230603
5	Date of purchase / Year of manufacture / Installation Date	22-05-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	22-03-2024
9	Action taken	Checked and found that the sterilizer blasted, chamber broken, coil damaged
10	Present status of the equipment (Fully damaged / partially damaged) .	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	Rs .14160
14	# Percentage value of the cost of spareswith respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report & images attached.
16	Reasons for recommending the equipment as BER	Checked and found that the sterilizer blasted, chamber broken, coil damaged it's not in repairable condition totally damaged. Its installed on 22-05-2019 outlived 4 year 10 month .so recommending to condemnation
17	Name & Signature of CYRIX Authority	RAJESHLAL 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Consultant (Biomedical) NHM:


checked the machine and found that body broken . Recommended for Beyond economic repair .



05/04/24

Signature of JC BM (NHM)

ABMO
NHM, KOLLAM

Date




10/4/24


Signature of
Superintendent / Medical Officer (1/2)
PHC Mayyand

Dr. TRESA LEELA
MBBS, FCOG
Reg No: 26620
PHC Mayyand



101

DI. TRESA-L
 MSB PCGR
 Reg No: 1000
 Assl. Subd. P.C.
 PHC Tithy, Gand.



Tresha Beer
-
07/4.
DR. TRESHA BEER
M.B. (M.D.)
Raj. N. (M.D.)
A.B. (M.D.)
M.B. (M.D.)
M.B. (M.D.)
M.B. (M.D.)

STOCK REGISTER

Name 42 RO AUTOCLAVE 12*12 12x22

Date	Received from or issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	P.				
22/5/19	INNO 2019/20/M93 1. Autoclave 12*12 R. 10000 Amt 14160	Dtd: 21/5/19			1		1	From MEDICRAE
23/5/19	Issued to dental op:					1		
21/8/19	INNO: 2019/20/MC55 1. STAINLESS STEEL AUTOCLAV 12x22	Dtd: 20/8/19			1		1	From MEDICRAE SANKAR HOSPITAL Junction, Kollam.
2/9/19	Issued to op		24000/-			1		
						1		

Subscrip page no: 71

Tresa Leen
-
17/4.



DR. TRESA LEEN
MBBS, FCOG
Reg No: 26020
MCh, Surgeon
PHC *Malappuram*