



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 191093

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>C.H.C. Ramapuram</u></p> <p>Address <u>Ramapuram</u> <u>Kottayam</u></p> <p>Ph : <u>6238898244</u></p>	<p>Call Registration Date : <u>01-03-2024</u></p> <p>Caller ID : <u>99039</u></p> <p>Date of Visit : <u>02-03-2024</u></p> <p>Asset No. : <u>0521360</u></p> <p>EQPT Name : <u>Sterilizer</u></p> <p>Manufacture <u>Coley</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>Palliative</u></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working well

**Action Taken :** Checked. Top door, internal gasket, safety valve, coil, power cord and pressure gauge are defective. Needs to be replace these spares for further checking and working condition. Enquired spares with OEM.

Completed  Date : 02.03.2024 Time : 11:55am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Nidhish Roy</u>	<u>02-03-2024</u>	<u>09:00am</u>	<u>11:55am</u>

**Customer Remark** Completed  Pending



**Service Engineer Name :** Nidhish Roy

**Signature :** NBR

**Date :** 02-03-2024

**Contact Number :** 9072575296

**Customer Name :** Betty

**Signature :** Betty

**Date :** 02/03/2024

**Contact Number :** 6238898244

**Designation :** Palliative Nurse

**Hospital Seal :**




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	CHC RAMAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Sterilizer Make :GL Model :GL SN:NIL
4	Equipment ID & Barcode	99039 & 0531360
5	Date of purchase / Year of manufacture /Installation Date	26/11/2013
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	01-3-2024 (Toll free)
9	Action taken	Checked and found out the problem with top door , gasket, safety valve ,coil, power code, and pressure gauge
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	8145/-


13	Asset Value	Rs . 8800/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	92.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor quotation attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with top door , gasket, safety valve, coil, power code, and pressure gauge Machine installed in 26/11/2013 and covered up to 11 +year. The repairing cost 92.55%, both criteria met. So recommending the machine for condemnation as per the tender close 5.3.14.1
17	Name & Signature of CYRIX Authority	Nidhish roy 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - Rs 8800  
 Cost of spare Rs. 8145  
 Installation Date - 26/11/2013  
 Life period - 10 year as per the BER guidelines recommended as BER

  
 Signature of JC BM (NHM)

Date

 Seal

  
 Signature of Medical Officer in Charge  
 Superintendent Medical Officer  
 F.H.C. Ramapuram

8 BULB. 40 wats

Date		Receipt	nil	Balance
29-02	Donated by the Rotary club of Ramapuram.	10 Nos.		10 Nos.
26-11-02			10	NIL

AUTOCLAVE DRUM

26/11/13	Donated by Augustine Thomas, Nannvelli, Amnakara			
	Rtd Jk Mehta & sons as per invoice no. 17744 dtd 23/11/13			
	Electric Autoclave Size: 12x12	Unit price	8800/-	
		1	1	

26/11/13	Issued to palliative primary nurse	1		NIL
----------	------------------------------------	---	--	-----

Dressing Drum

26/11/13	Donated by Augustine Thomas, Nannvelli, Amnakara			
	Rtd Jk Mehta & S. Mehta & sons as per invoice no. 17744 dtd 23/11/13			

26/11/13	Dressing Drum 9x9 Issued to palliative primary nurse	1		NIL
----------	---	---	--	-----



Arun GC &lt;arungcyrix@gmail.com&gt;

---

**RBER PALLICKATHOD AND RAMAPURAM PHC**

---

**CLOUD 99** <cloud99biotech@gmail.com>  
To: "arungcyrix@gmail.com" <arungcyrix@gmail.com>

Mon, Apr 8, 2024 at 3:43 PM

Dear Sir  
please find the attachments for RBER OF PALLICKATHOD AND RAMAPURAM

Thanks and regards

Sunil Thomas  
Cloud99 Biotech  
Medical Engineering Works  
Room No: 22, CSI Golden Jubilee Complex  
Kottayam 686 001  
Kerala, India  
Ph: 9744473043, 9744473243  
email: [cloud99biotech@gmail.com](mailto:cloud99biotech@gmail.com)

---

**2 attachments**

**0141- CHC RAMAPURAM, AUTOCLAVE CONDEMNATION.docx**  
34K



**0142-PHC PALLIKKATHOD, AUTOCLAVE CONDEMNATION.docx**  
35K



# CLOUD99 BIOTECH

Office : 97444 73043  
Service : 97444 73243

*Sales and service of all kinds of medical equipments*  
Room No: 22, First Floor, CSI Golden Jubilee Complex,  
Behind Shastri Road Waiting Shed, Kottayam-686001  
Email: cloud99biotech@gmail.com

## QUOTATION

Ref No: C99B/0139/2023-24

Date: 14/03/2024

To,

CHC RAMAPURAM  
PALA, KOTTAYAM

Dear Sir

We are pleased to submit a quotation for CHC Ramapuram for Cooker Type Autoclave

SL No	ITEM NAME	PRICE PER UNIT	GST	PRICE INCLUDING GST
1	GASKET	550	12%	649
2	COIL 2000 W	1100	18%	1298
3	POWER CODE WITH COIL CONNECTOR	450	18%	531
4	PRESSURE GAUGE	650	18%	767
5	SAFTY VALVE	550	18%	649
6	STEAM RELEASE VALVE	350	18%	413
7	TOP DOOR	3250	18%	3835
	<b>TOTAL</b>			<b>8145</b>

### Terms & Conditions

1. Taxes: GST @ 18 % inclusive
2. Payment Terms: 100% advance payment
3. Quotation validity: 15 days from date of quotation

We hope you will find our rates are competitive and wait for your valued order.

Thanking you,

Yours Faithfully

FOR CLOUD99 BIOTECH

AUTHORIZED SIGNATORY



We deal with sales and service of all kinds of hospital equipments, neonatal equipments, laboratory equipments, dental equipments, suction, hospital furniture, radiology equipments, physiotherapy equipments, all type of autoclaves, ICU equipments, etc.

