



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 166294

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>phc</u>	Call Registration Date : <u>4/3/2024</u>
Address : <u>pallickathodu</u>	Caller ID : <u>99392</u>
<u>Kottayam</u>	Date of Visit : <u>7/3/2024</u>
Ph : <u>9744967284</u>	Asset No. : <u>0543330</u>
	EQPT Name : <u>Autoclave</u>
	Manufacture : <u>NA</u> Model : <u>NA</u>
	S. No. : <u>NA</u> Dept. : <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working well

Action Taken : checked top door, internal gasket, safety valve coil, power cord and pressure gauge are defective. Needs to be replace these spares for further cheary and working condition. Engused spares with OEM

Completed Date : 7/3/2024 Time : 1:15pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AMAL BABU</u>	<u>7/3/2024</u>	<u>1:15pm</u>	<u>2:05pm</u>

Customer Remark : Equipment is not able to use. Completed Pending

Nishamol Nishamol Besheed
Nursing Officer CrI

Service Engineer Name : <u>AMAL BABU</u>	Customer Name :
Signature :	Signature :
Date : <u>7/3/2024</u>	Date :
Contact Number : <u>9567525983</u>	Contact Number :
	Designation : <u>Medical Officer In-Charge</u>
	Hospital Seal : <u>PHC Pallickathodu</u>
	<u>Aruvikuzh</u>


14/2/24
57

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC PALLICKATHODU
3	Name of Equipment with Make, Model and Serial Number	Equipment : Sterlizer Make :NA Model :NA SN:NIL
4	Equipment ID & Barcode	99392 & 0543330
5	Date of purchase / Year of manufacture /Installation Date	31/01/2012
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	04-3-2024 (Toll free)
9	Action taken	Checked and found out the problem with top door , gasket, safety valve ,coil, power code, and pressure gauge
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Gasket - 649/- Coil 2000w - 1298/- Power code - 531/- Pressure gauge - 767/- Safety valve - 649/- Steam release valve - 413/- TOTAL- 4307/-


13	Asset Value	Rs . 4950/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	87%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor quotation attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with top door , gasket, safety valve ,coil, power code, and pressure gauge machine installed in 31/01/2012 and covered up to 12 +year. The repairing cost 87%, both criteria met. So recommending the machine for condemnation as per the tender close 5.3.14.1
17	Name & Signature of CYRIX Authority	AMAL BABU 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset Value = 4950. / -
 cost of Spare : 4307 / - (87.1.)
 Installation Date - 31/1/2012
 Life period - 12 year
 As per the RBER guideline recommend as
 RBER


 Signature of JC BM (NHM)



Date 7/3/24

Seal


 Signature of Medical Officer In-Charge
 Superintendent / Medical Officer (i/c)



Arun GC <arungcyrix@gmail.com>

RBER PALLICKATHOD AND RAMAPURAM PHC

CLOUD 99 <cloud99biotech@gmail.com>
To: "arungcyrix@gmail.com" <arungcyrix@gmail.com>

Mon, Apr 8, 2024 at 3:43 PM

Dear Sir
please find the attachments for RBER OF PALLICKATHOD AND RAMAPURAM

Thanks and regards

Sunil Thomas
Cloud99 Biotech
Medical Engineering Works
Room No: 22, CSI Golden Jubilee Complex
Kottayam 686 001
Kerala, India
Ph: 9744473043, 9744473243
email: cloud99biotech@gmail.com

2 attachments

0141- CHC RAMAPURAM, AUTOCLAVE CONDEMNATION.docx
34K



0142-PHC PALLIKKATHOD, AUTOCLAVE CONDEMNATION.docx
35K



CLOUD99

Office : 97444 73043

Service : 97444 73243

Sales and service of all kinds of medical equipments
Room No: 22, First Floor, CSI Golden Jubilee Complex,
Behind Shastri Road Waiting Shed, Kottayam-686001
Email: cloud99biotech@gmail.com

QUOTATION

Ref No: C99B/0142/2023-24

Date: 20/03/2024

To,

PHC PALLIKKATHOD
PAMPADY, KOTTAYAM

Dear Sir

We are pleased to submit a quotation for PHC Pallikkathod for Cooker Type Autoclave.

SL No	ITEM NAME	PRICE PER UNIT	GST	PRICE INCLUDING GST
1	GASKET	550	12%	649
2	COIL 2000 W	1100	18%	1298
3	POWER CODE WITH COIL CONNECTOR	450	18%	531
4	PRESSURE GAUGE	650	18%	767
5	SAFTY VALVE	550	18%	649
6	STEAM RELEASE VALVE	350	18%	413
	TOTAL			4307

Terms & Conditions

1. Taxes: GST @ 18 % inclusive
2. Payment Terms: 100% advance payment
3. Quotation validity: 15 days from date of quotation

We hope you will find our rates are competitive and wait for your valued order.

Thanking you,

Yours Faithfully



FOR CLOUD99 BIOTECH



RBER **CYRIX**
INDUSTRIAL EQUIPMENT
"Commitment to Reliability"

Ticket No. 19392
 RBER approved on: 7/31/2024
 Condemn on:
 Remark:



Handwritten signature
 2/11/24

In-Charge
 Kathodu
 686 503



RBER **CYRIX**
REPAIR & MAINTENANCE
"Caring is Different"

Ticket No.: 99892
 RBER approved on: 7/3/2024
 Condemn on:
 Remark:



2/4/24

In-Charge
 Rathodu
 686 503