


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

470

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANATHAPURAM
2	Name of Hospital	DISTRICT HOSPITAL, NEDUMANGADU
3	Name of Equipment with Make, Model and Serial Number	Equipment : NEBULISER Make : PHILIPS Model : RESPIRONICS SN : 1022722944
4	Equipment ID & Barcode	100888 & 0112397
5	Date of purchase / Year of manufacture /Installation Date	13-06-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	12-03-2024 (Toll free)
9	Action taken	Checked the machine and found compressor motor with piston defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 1353
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found compressor motor with piston defective. Machine installed on 13.06.2018 and covered up to 5+years. Quotation not submitted since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AKHIL MANOHAR 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Compressor motor and piston damaged. Hence it is BER.


MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


10-4-2024

Signature of JC BM (NHM)



Date _____ Seal _____


Superintending Officer (i/c)
മെഡിക്കൽ ഓഫീസ്



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 157244

470

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/84 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99300 Website : www.cyrilx.com | E-mail : bomp.kl@cyrilx.in

Health Facility : <u>Disease Hospital</u>	Call Registration Date : <u>12.03.2024</u>
Address : <u>Nedumangala</u>	Caller ID : <u>100888</u>
<u>Trivandrum</u>	Date of Visit : <u>15.03.2024</u>
Ph : <u>6282146538</u>	Asset No. : <u>0112397</u>
	EQPT Name : <u>Nebulizer</u>
	Manufacture <u>Philips</u> Model : <u>Respironics</u>
	S. No. <u>1022722944</u> Dept. <u>C- Ward</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not getting on

Action Taken : Checked and found compressor motor with piston defective. Need to replace these spares for further checking and working condition of the equipment.

Completed Date : 15/03/2024 Time : 11:45 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	<u>-</u>	<u>-</u>	<u>-</u>
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Aksh</u>	<u>15/03/24</u>	<u>11:00 AM</u>	<u>11:45 AM</u>


Customer Remark Completed Pending

Service Engineer Name : <u>Aksh M...</u>	Customer Name : <u>...</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>15/03/2024</u>	Date : <u>15/03/24</u>
Contact Number : <u>8921568751</u>	Contact Number : <u>6282146538</u>
	Designation : <u>...</u>
	Hospital Seal : <u>[Seal]</u>

NEBULIZER

113

Date	From whom received / to whom issued.	Cost	Receipt	Issued	Balance	Signatures	Remarks
13/6/2018	Tk. in from new stock instrument register page no. 168	1027/-	01		09	<i>[Signature]</i>	
28/7/2018	Tk. out to observation ward. instrument register volume I page no. 184.			01	03		Rd 1000 500 H/W 012
7/3/2019	Rtd from the store (F.W Register Page No 153)		1		4	<i>[Signature]</i>	
13/11/2019	Received from: Cancer Care Unit Instrument Book Reg No: T. page no 47		2.		6.	<i>[Signature]</i>	
2/1/2020	Tk out L.R. one nebulizer			1	5	<i>[Signature]</i>	
2/3/22	Rtd from store		2		7	<i>[Signature]</i>	
4/4/22	Received from store		2		9	<i>[Signature]</i>	
23.0.24	Rtd from store		01		10		
1.3.24	"		02		12	<i>[Signature]</i>	
10.03	from Cancer Care unit		01		13	<i>[Signature]</i>	


 അഡ്വൈസ്
 ജില്ലാ ആശുപത്രി
 മെട്രോമെന്റാട്



സുപ്രബ
ജില്ലാ ആശുപത്രി
നെടുമങ്ങാട്

1095063

InnoSpire Essence (EU Plug)

230V~50Hz 1A 30 min I / 30 min O

SN 1022722944

LOT 160512



Respironics Respiratory Drug Delivery (UK) Ltd
Chichester Business Park, City Fields Way, Tangmere
Chichester, PO20 2FT UK

Made in China

CE 0123




IP21



100

1108975 R00

സുപ്രണ്ട്
ജില്ലാ ആശുപത്രി
നെടുമങ്ങാട്



സുപ്രണ്ട്
ജില്ലാ ആശുപത്രി
നെടുമങ്ങാട്

[Handwritten signature]