



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 219303

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility ..... <u>CHC CHAVARA</u>	Call Registration Date : ..... <u>11/05/2024</u>
Address ..... <u>KOLLAM</u>	Caller ID : ..... <u>110619</u>
..... <u>KERALA</u>	Date of Visit : ..... <u>14/05/2024</u>
Ph : ..... <u>9744536733</u>	Asset No. : ..... <u>0230162</u>
	EQPT Name : ..... <u>Nebuliser</u>
	Manufacture ..... <u>Infu Neb</u> Model : <u>MCN-5600A</u>
	S. No. ..... <u>112018</u> Dept. ..... <u>IP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : ..... Nebuliser not getting ON.

Action Taken : ..... Checked the nebuliser and found that the motor defective and piston damaged. Need to replace these spares for further checking working condition of these equipment.

Completed  Date : 14/05/2024 Time : 3:00pm Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Willfest Sunil Willson</u>	<u>14/05/2024</u>	<u>2:45pm</u>	<u>3:00pm</u>

Customer Remark .....  Completed  Pending

Service Engineer Name : <u>Willfest Sunil Willson</u>	Customer Name : <u>Thimma Chavara</u>
Signature : <u>Willfest</u>	Signature : <u>Thimma</u>
Date : <u>14/05/2024</u>	Date : <u>14/5/24</u>
Contact Number : <u>9778526652</u>	Contact Number : <u>9744536733</u>
	Designation : <u>S.N.O.</u>
	Hospital Seal : <u>S.N.O.</u>




Medical Officer in Charge  
Community Health Centre

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	Kollam
2	Name of Hospital	CHC CHAVARA
3	Name of Equipment with Make, Model and Serial Number	NEBULIZER Make : INFINEB Model : MCN-S600A SI no: 112018
4	Equipment ID & Barcode	Barcode : 0230162 / Ticket id : 110619
5	Date of purchase / Year of manufacture /Installation Date	20-04-2012
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	11-05-2024
9	Action taken	Checked the machine found that the Nebulizer motor and piston are damaged. Need to replace these spares for further checking &working condition of these equipment
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not Recommended for Repair
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	Rs. 1353/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that the nebulizer motor and piston are defective, enquired for spares in the market but spare are not available. Equipment is installed on 20-04-2012 and aged upto 12 year. So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	WILFRET SUNIL 

\*Not mandatory #Based on the period of life and value as per the BER guidelines


\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Very old machine and fully damaged. So Recommended for Beyond Economic Repair

  
 Signature of ABMO 03/06/24  
 ABMO (NHM)  
 NHM, KOLLAM

Date 3/06/2024

  
 Seal

  
 Signature of Superintendent / Medical Officer (i/c)

Medical Officer  
 Community Health Centre  
 Chavara

WIRE & CASE



*[Handwritten signature]*  
Medical Officer in  
Family Health Cent  
para

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0230162



ISO 9001:2008 Certified Co. PH: 9744555073  
Date: 17.12.24  
Date: 18.12.23  
0230162  
NEBULIZER  
CHCH37  
HEALTHCARE PVT LTD  
CYRIX

Compressor Nebulizer  
Model : MCN-S600A  
Input : 230V / 50 Hz  
Power : 80 VA  
Serial No. : 112018  
Manufactured By :  
Innovative Medical Devices  
38A, Nahapur Village, Sector-7,  
Rohtak, Delhi-110085  
Website : www.inthealthcare.com



*[Signature]*  
Medical Officer in Charge  
Community Health Centre  
Chavara

NEBULIZER

(BIC from page no - 84 Old stock register)

45

Date	Particulars	Rd Issued	Bal
07/07/90	Rd from store	1	1
12/06/2008	Rd from store	2	3
12/05/2008	Rd from store	1	4
12/01/2009	Rd from store	1	5
20/04/12	Rd from store	1	6
20/04/12	Rd from store	1	7
20/05/2020	Rd from store	1	8
23/02/2023	As per order no 406/2023/mo care chavch & dte 23/02/2023 consumed Snc Care		
			Bal 3
			=

No.

Sub stock - H/N - Shankar  
SND.

*Shankar*  
 District Engineer in Charge,  
 Maharashtra State  
 Government  
 Mumbai