



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 217817

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>TICOH MALLAPPALLY</u>	Call Registration Date : <u>30-05-2024</u>
Address <u>MALLAPPALLY</u>	Caller ID : <u>11 42 41</u>
<u>PATHANAMTHITTA</u>	Date of Visit : <u>31-05-2024</u>
Ph : <u>9496266263</u>	Asset No. : <u>0321540</u>
	EQPT Name : <u>nebuliser</u>
	Manufacture <u>Philips</u> Model : <u>InnoSpire Elegance</u>
	S. No. <u>1023608634</u> Dept. <u>WARD</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the machine Found the problem with compressor motor with piston . Compressor motor and piston defective

Completed Date : 31/05/24 Time : 11:30 AM Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1. /	/	/	/
2. /	/	/	/
3. /	/	/	/

Cyrix Engineer	Date	Start Time	End Time
<u>Abhilash Kumar</u>	<u>31/05/24</u>	<u>11:00 AM</u>	<u>11:30 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Abhilash Kumar
Signature : [Signature]
Date : 31-05-2024
Contact Number : 9400181609

Customer Name : Preetba Nathan
Signature : [Signature]
Date : 31/5/2024
Contact Number : 9498266000
Designation : Senior Superintendent
Hospital Seal : TALUK HQ. HOSPITAL MALLAPPALLY




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PATHANAMTHITTA
2	Name of Hospital	THQH MALLAPALLY
3	Name of Equipment with Make, Model and Serial Number	NAME- NEBULISER MAKE- Philips MODEL- Innospire Elegance SL.NO- 1023608634
4	Equipment ID & Barcode	114241 & 0321540
5	Date of purchase / Year of manufacture /Installation Date	29-03-2019
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	30/05/2024
9	Action taken	CHECKED AND FOUND THAT COMPRESSOR MOTOR WITH PISTON DEFECTIVE.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	Rs. 1353 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	CHECK AND FOUND THAT COMPRESSOR MOTOR WITH PISTON DEFECTIVE. EQUIPMENT INSTALLED ON 29/03/2019 AND AGED UPTO 4+ YEARS. QUOTATION NOT SUBMITTED SINCE SPARES ARE NOT AVAILABLE IN THE MARKET.SO RECOMMENDING THE EQUIPMENT FOR RBER.
17	Name & Signature of CYRIX Authority	Abhilash Kumar 


*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Physically verified on 10/6/2024. Machine is fully damaged. Recommended to BER as per rule. Stock entry details not provided from hospital.


 Signature of JC BM (NHM) 12.6.24


 Seal


 SUPERINTENDENT
 TALUK H.O. HOSPITAL
 Signature of
 Superintendent / Medical Officer (i/c)
 MALLARPALLY

Date

ഭരണ ഭാഷ മാതൃഭാഷ

താലൂക്ക് ആസ്ഥാന ആശുപത്രി മല്ലപ്പള്ളി

മല്ലപ്പള്ളി വെസ്റ്റ് പി. ഒ, മല്ലപ്പള്ളി, പിൻ -689 585

ഫോൺ:0469-2683084 ഇമെയിൽ ഐഡി:thospitalmallappally@gmail.com

കത്ത് നം.എ2-1343/2024

താലൂക്ക് ആശുപത്രി, മല്ലപ്പള്ളി

പ്രേഷകൻ

സൂപ്രണ്ട്

താലൂക്ക് ആശുപത്രി

മല്ലപ്പള്ളി

സ്വീകർത്താവ്

മാനേജിംഗ് ഡയറക്ടർ

കെ.എം.സി.എൽ. ലിമിറ്റഡ്

തിരുവനന്തപുരം

സർ

വിഷയം:ഉപയോഗരഹിതമായ ഉപകരണങ്ങൾ ഗർഹണം ചെയ്യുന്നതിനുള്ള അനുമതി
-സംബന്ധിച്ച്

സൂചന:സീനിയർ നഴ്സിംഗ് ഓഫീസറുടെ കത്ത് തീയതി 13.05.2024

മല്ലപ്പള്ളി താലൂക്ക് ആശുപത്രിയിലെ പ്രവർത്തന രഹിതമായ താഴെ പറയുന്ന ഉപകരണങ്ങളുടെ Stock Details ഈ സ്ഥാപനത്തിൽ ലഭ്യമല്ലാത്തതിനാൽ ഗർഹണം ചെയ്യുന്നതിനുള്ള അനുമതി നൽകമെന്ന് അഭ്യർത്ഥിക്കുന്നു.

ക്രമ നം.	ഉപകരണം	ബാർകോഡ്
1	Nebuliser	0321540
2	Nebuliser	0321671

മല്ലപ്പള്ളി
20.05.2024



സൂപ്രണ്ട്

താലൂക്ക് ആശുപത്രി

മല്ലപ്പള്ളി

**SUPERINTENDENT
TALUK H.Q. HOSPITAL
MALLAPPALLY**

