


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|---|--|
| 1 | Name of District | Ernakulam |
| 2 | Name of Hospital | 24X7 PHC MANJALLOOR |
| 3 | Name of Equipment with Make, Model and Serial Number | Equipment: BP APPARATUS Make : LIFECARE Model : BSX555 Sr.no : NA |
| 4 | Equipment ID & Barcode | 104729 & 0744813 |
| 5 | Date of purchase / Year of manufacture /Installation Date | 27/02/2019 |
| 6 | Warranty details (Yes/No) | No warranty |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | No AMC/CAMC |
| 8 | Date of breakdown(Date of registration of complaint through email/ Toll free) | 04-04-24 (Toll Free) |
| 9 | Action taken | Checked the machine and found that motor and main board are faulty. |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | Fully damaged |
| 11 | Recommendations for repair (required service details) | Not recommending for repair. |
| 12 | Cost of spares (specify parts and cost) | NA |



| | | |
|----|--|---|
| 13 | Asset Value | Rs.2200 |
| 14 | # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | NA |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Cyrix service report attached |
| 16 | Reasons for recommending the equipment as BER | Checked and found motor and main board are defective. Equipment installed on 27/02/2019 and the machine outlived 5+years. Quotation not submitted since the spares are not available in the market. So recommending the equipment for condemnation. |
| 17 | Name & Signature of CYRIX Authority | JIBIN THOMAS GEORGE  |

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
*call on 14/4/24.
 BER submitted on 14/5/24*

Dewathy
 Signature of JC BM (NHM) 
 Bio-Medical Engineer
 National Association of Biomedical Engineers (NABE) 


Superintendent / Medical Officer (i/c) *JTG*
 തിരുവനന്തപുരം ജില്ലാ ആരോഗ്യ ഓഫീസ്
 കൂടുംബാരോഗ്യ കേന്ദ്രം
 മഞ്ഞമുളൂർ പിൻ : 686 670

Signature of seal  Date



Electronic Blood Pressure Monitor
Model No.:BSX555
Voltage:6V(4*AA Batteries)
Pressure Range/Accuracy:(0-290) \pm 3mmHg
Pulse Range/Accuracy:(40-180)Beats/min \pm 5%
SN:H202103000120 Lot No.:555CAL19

J
18/6/1





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 196466

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|---|---|
| <p>Health Facility <u>24X7 PHC MANJALLOOR</u></p> <p>Address <u>Ernathulam,</u> <u>Ernathulam (Dt.)</u></p> <p>Ph :</p> | <p>Call Registration Date : <u>04-04-2024</u></p> <p>Caller ID : <u>104729</u></p> <p>Date of Visit : <u>05-04-2024</u></p> <p>Asset No. : <u>0744813</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>Lifecare</u> Model : <u>BSX555</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p> |
|---|---|

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : BP Apparatus not working

Action Taken : checked the machine, and found that motor and mainboard are faulty.

Completed Date : 05-04-24 Time : 10:30AM Spare Required

Spare Replaced Requested

| No. | Description | Qty. | Part Number | PR Number |
|-----|-------------------|--------------|-------------|-----------|
| 1. | <u>Motor</u> | <u>1 No.</u> | | |
| 2. | <u>Main Board</u> | <u>1 No.</u> | | |
| 3. | | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|----------------------------|-----------------|---------------|----------------|
| <u>Jibin Thomas George</u> | <u>05-04-24</u> | <u>9:00AM</u> | <u>10:30AM</u> |

Customer Remark Completed Pending

Service Engineer Name : Jibin Thomas
Signature : Jibin
Date : 05-04-2024
Contact Number : 9047958051

Customer Name : Bridee John
Signature : Bridee John
Date : 18/6/24
Contact Number : 7025091176
Designation : Nursing Officer
Hospital Seal :

