



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 183368

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>THQH Haripad</u>	Call Registration Date : <u>27/01/2024</u>
Address <u>Dist Alappuzha</u>	Caller ID : <u>92951</u>
Ph : <u>9446857772, 9388955075</u>	Date of Visit : <u>29/01/2024</u>
	Asset No. : <u>0420910</u>
	EQPT Name : <u>Bilirubinometer</u>
	Manufacture <u>Drager</u> Model : <u>Jm-105</u>
	S. No. <u>B3201063</u> Dept. <u>Labor Room</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : machine not getting on

Action Taken : checked the machine and found that sensor set, power supply PCB & Battery are defective. Enquired the OEM for quotation

Completed Date : 20/02/2024 Time : 2.00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Savath Chandran</u>	<u>20/02/2024</u>	<u>1.00pm</u>	<u>2.00pm</u>
	<u>u</u>	<u>u</u>	<u>u</u>



Customer Remark Bindu M.L Senior & In-charge officer THQH, Haripad Completed Pending

Service Engineer Name : Savath Chandran
Signature : [Signature]
Date : 20/02/2024
Contact Number : 8075052698

Customer Name : Bindu M.L
Signature : [Signature]
Date : 20/2/2024
Contact Number : 9446857772
Designation : Senior & In-charge officer
Hospital Seal :


Superintendent
Taluk Head Quarters Hospital
Haripad- 690 594

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH HARIPPAD
3	Name of Equipment with Make, Model and Serial Number	BILIRUBINOMETER MAKE:- DRAGER MODEL :- JM-105 SN :- B3201063
4	Equipment ID & Barcode	92951 & 0420910
5	Date of purchase / Year of manufacture /Installation Date	29/07/2015
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	27/01/2024
9	Action taken	Checked the machine and found that Sensor Set, Power Supply PCB & Battery are defective. Enquired the OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	Not Recommended For Repair
12	Cost of spares (specify parts and cost)	Sensor Set :- 163759.22/- Power Supply PCB :-54960.86/- Battery :- 25280/ Total :- 244000.08/-

13	Asset Value	RS :- 260000/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	93%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report And OEM Quotation
16	Reasons for recommending the equipment as BER	Checked the machine and found that Sensor Set, Power Supply PCB & Battery are defective. Enquired the OEM for quotation. Equipment installed on 29/07/2015 aged up to 8 years and 6 months. Percentage of service cost is high (93%) and economically not viable. Both criteria met as per the tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	SHARATH CHANDRAN 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

JYOTHISH THOMAS
Consultant (Biomedical Engineering)
NHM ALAPPUZHA

 21/02/24

Signature of JC BM (NHM)

Date *18/02/24*



 18/02/24

Signature of
Superintendent / Medical Officer (i/c)
Superintendent

Taluk Head Quarters Hospital
Haripad- 690 514

QUOTATION



To, **Customer Code** : C05618
Name : CYRIX HEALTHCARE PRIVATE LIMITED
Address : 1ST FLOOR, 30/641B, KOCHI DHANUSHKODI ROAD,
City : Ernakulam, KL, 682038
Phone No : 9 809 237 001
Document Date : 03-Feb-24
Customer PO No. :
External Doc No. :
GST No.: 32AAFCC2499H2ZM

Quotation No : VSI/SQ/SR/02-24/0030
Customer PO Date :
Service Person: Vishnu Muraleedhar
Service Person Email :
 vishnu.muraleedhar@draeger.com
Service Person Phone No. :
 9446610210
Company Email ID:
 customercareindia@draeger.com

Item Code	Description	Qty	Rate	HSN	CGST Amount	CGST %	IGST Amount	IGST %	SGST Amount	SGST %	Amt
MU00791	Sensor set 1712-0651-05	1	138779	90279090	0.00	0	24980.22	18	0.00	0	138779
MU24868	Power supply PCB, JM-105	1	46577	850440	0.00	0	8383.86	18	0.00	0	46577
MU24842	Battery, JM-105	1	19750	85075000	0.00	0	5530.00	28	0.00	0	19750
Sub Total											205106.00
Freight											0.00
IGST											38894.08
Grand Total											244,000.08

Draeger India Private Limited

Registered Office :
 10th Floor, Commerz II, International Business Park, Oberoi
 Garden City, Goregaon-East
 Mumbai, Maharashtra - 400063, India
 Tel. : +91 22 61555600
 FAX : +91 22 61555670 CIN No. U51507MH2004FTC14399
 Website : GST No. : 27AACCD3263D1Z0

HSBC BANK LTD : :
 52/60, Mahatma Gandhi
 Road,, Mumbai - 400 001-
 IFSC - HSBC0400002
 Acc. No - 006051767001
 MICR - 400039002
 SWIFT-HSBCINBB

For M/s. Draeger India Private Limited

Jithin Raj
 Digitally signed
 by Jithin Raj
 Date:
 2024.02.03
 23:12:42 +05'30'

Authorised Signatory

Terms and Conditions :

- Order must be place in the name of "Draeger India Pvt Ltd
 - Payment : 100% advance in favour Draeger India Pvt Ltd
 - Delivery : Subject to availability or 4 to 6 Weeks from the date of receiving the P.O.
 - Validity : 30 Days.
 - Service Charge Extra if equipments not in Contract
 - NOTE: Dräger shall not be responsible for any delay or other failure in performing its contractual obligations, if the performance of such obligations is prevented or materially affected as a direct or indirect result of the corona virus pandemic. Such causes include, without limitation, unpunctual or incorrect delivery by suppliers, impairment of Dräger's own production due to employee absences or precautionary measures. In such cases the delivery time shall be extended automatically.
- Dräger is entitled to withdraw from this agreement fully or partly by written notice, should the contractual performance become impossible.
 Any and all other rights pursuant to the standard terms and conditions for sales and service of remain unaffected

This is a system generated 'Sales Quotation', hence needs no signature.



(Handwritten signature)

Superintendent
Taluk Head Quarters Hospital
Haripad- 690 514



STOCK BOOK OF.....

Name of Article Printer Jayadev meter Jm - 105

Date	No. and Date of voucher or Invoice	From whom received or to whom issued	Receipt	Issued	Balance After each transaction	Initials of receiver	Remarks
		Brought Forward					
23/9/12		BF from NRHM stock register PG no 129 NO KMCL/EP/549/653 Dragey medical (India) Pvt LTD 357-230B/ Surya habita building Automobile Road, Kochi 682025 Invoice no. LW/PSIC/15-10/0087 Warranty 3yrs from 29/7/15 Name of Service Engineer Jayaram mobile no 9539873030 DORPT - 29/7/15 Rs - 2,31,505.00 Jithin - 9593007399. Libu - 9567862619					
21/2/2020							

Superintendent
Taluk Head Quarters Hospital
Haripad - 596 514



Carried over