


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	CHC PINARAYI
3	Name of Equipment with Make, Model and Serial Number	REFRIGERATOR Make: WHIRLPOOL Model: GENIOUS SN: NA
4	Equipment ID/ Barcode	1331244 #105521
5	Date of purchase/ Year of manufacture/Installation Date	19-04-2015
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 09-04-2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, found that Compressor, Thermostat and filter also defective. Enquired for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Compressor-7,300/- Gas charging-2,400/- Filter-480/- Service- 2,000/- Thermostat-1,500/- Total=14,160/-


13	Asset Value	17961/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	78.8374%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not switching ON, found that Compressor, Thermostat and filter also defective. Enquired for quotation. Equipment installed on 19-04-2015. Aged up to 9 years. Repairing cost is 78.8374% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that machine not switching on, compressor, thermostat & filter defective. Aged up to 9 years. So recommended for BER as per BER protocol & quotation.




03/05/2024 Signature of JOIBM (NHM)

NATIONAL HEALTH MISSION

KANNUR - 670002

Date: 8/05/2024




Signature of
Superintendent Medical Officer (i/c)
COMMUNITY HEALTH CENTRE,
PINARAYI-570741



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 189643

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>CHE Pinarayi</u></p> <p>Address : <u>Kannur</u> <u>Kerala</u></p> <p>Ph : <u>98 46 64 6001</u></p>	<p>Call Registration Date : <u>9-4-2024</u></p> <p>Caller ID : <u>105521</u></p> <p>Date of Visit : <u>11-4-2024</u></p> <p>Asset No. : <u>1331244</u></p> <p>EQPT Name : <u>Refrigerator</u></p> <p>Manufacture : <u>Whirlpool</u> Model : <u>Cuenians</u></p> <p>S. No. : <u>NA</u> Dept. : <u>pharmacy</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Compressor, thermostat and filter are defective.

Action Taken : checked the equipment. Found equipment it had switching on. Found that compressor, thermostat and filter are defective. Enquired for quotation.

Completed Date : 11-4-24 Time : 12:30pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Shyamkrishnan</u>	<u>11-4-24</u>	<u>10:30 AM</u>	<u>12:30 PM</u>

Customer Remark Completed Pending



MEDICAL OFFICER
COMMUNITY HEALTH CENTRE
PINARAYI 670741
19/4/24
9846646001
Pharmacy

Service Engineer Name : Shyamkrishnan

Signature : [Signature]

Date : 11-4-24

Contact Number : 8281188514

Customer Name : [Signature]

Signature : [Signature]

Date : 19/4/24

Contact Number : 9846646001

Designation : Pharmacy

Hospital Seal :



Required one time service 1331244#105521 @chc pinarayi

2 messages

avinash t <avinashtcyrix@gmail.com>
To: ecotonicsolutions@gmail.com

Fri, 21 Jun, 2024 at 8:39 pm

Dear sir,
Required your one time service for CHC pinarayi refrigerator to rectify problem. Please check and provide estimate

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

ECOTONIC AUTOMATION SOLUTIONS <ecotonicsolutions@gmail.com>
To: avinash t <avinashtcyrix@gmail.com>

Fri, 21 Jun, 2024 at 9:39 pm

Hi Avinash
Your quote for above mentioned Machine is attached
PFA

Regards
Ecotonic Automation Solutions
+91 75103 48448
[Quoted text hidden]

STOCK BOOK OF REFRIGERATOR

Name of article... ~~Refrigerator~~ 165-Lt. Whar P. Sol. C. Stabiliser

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of receiver	Remarks
		Brought forward	--				

2-1-99 Transferred from Lab- on loan basis. 1-Nr. 1-Nr. Rs. 13300-00

See page No. 52 of Stock Register of Lab.

~~19/5/2016~~ ~~Received from Lab. Card~~ ~~Physically made~~ ~~Registered in Stock~~ ~~of Pharmacy~~ 1-Nr. 1-Nr.

7/5/16 bought from Local purchase register vol. IV (Ca) page 58 1 1 Rs: 15,000

2 transfered to Inventory register page no 41, vol. II


MEDICAL OFFICER
COMMUNITY HEALTH CENT.
PINARAY 1-670741

7/5/16 kept ~~in~~ in Drug 1 1 ml

4/11/21 Del in ~~store~~ ^{room} 01 01
(kept in store) on line 01
For Card box

(Reg. no. 74 of NARM Stock Register - Inu ward)

24/11/21 + unit Del in Inu (the item was same that was received from ward on 4/11/21 as the item was used from Covid hospital Accordingly physically on 4/11/21 and was issued to Inu unit. ~~Physically~~ ~~kept~~ ~~in~~ ~~store~~ ~~on~~ ~~room~~ ~~Delivered~~ ~~to~~ ~~pharmacy~~ ~~for~~ ~~immediate~~ ~~use~~ ~~and~~ ~~after~~ ~~10~~ ~~months~~ ~~that~~ ~~Covid~~ ~~hospital~~ ~~gave~~ ~~us~~ ~~issue~~ ~~note~~ ~~through~~ ~~DDMS~~ ~~on~~ ~~24/11/21~~ ~~and~~ ~~Recd~~ ~~that~~ ~~issue~~ ~~note~~ ~~through~~ ~~DDMS~~ ~~and~~ ~~the~~ ~~both~~ ~~transaction~~ ~~are~~ ~~same~~ ~~item~~ ~~(Rehyreudo~~ ~~made~~ ~~by~~ ~~Parasone)~~





Sellu

Shened

